

Data Path : Z:\HPCHEM1\BNA F\DATA\BF061016\
 Data File : BF087883.D
 Acq On : 10 Jun 2016 19:20
 Operator : UM/SJ
 Sample : H3395-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 STA-928-PLATFORM(0-4)

Manual Integrations
 APPROVED

SOHIL
 6/13/2016 7:11:13 PM

Quant Time: Jun 11 02:12:14 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 07 18:51:55 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.81	152	121596	20.00	ng	-0.03
21) Naphthalene-d8	8.10	136	483452	20.00	ng	-0.02
38) Acenaphthene-d10	9.85	164	205790	20.00	ng	-0.03
63) Phenanthrene-d10	11.32	188	276291	20.00	ng	-0.03
75) Chrysene-d12	13.97	240	234112	20.00	ng	-0.02
86) Perylene-d12	15.37	264	220358	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	891305	125.31	ng	-0.01
7) Phenol-d6	6.46	99	1154028	133.88	ng	-0.02
23) Nitrobenzene-d5	7.38	82	693574	83.77	ng	-0.03
41) 2,4,6-Tribromophenol	10.64	330	193012	88.83	ng	-0.03
44) 2-Fluorobiphenyl	9.18	172	938774	70.73	ng	-0.03
78) Terphenyl-d14	12.91	244	661815	64.33	ng	-0.03
Target Compounds						Qvalue
48) Acenaphthylene	9.71	152	317683	15.00	ng	98
49) Dimethylphthalate	9.56	163	92854	6.23	ng	99
70) Phenanthrene	11.35	178	364967	25.17	ng	100
71) Anthracene	11.40	178	229224	15.67	ng	97
72) Carbazole	11.55	167	35719	2.61	ng	99
74) Fluoranthene	12.55	202	1434304	93.74	ng	94
77) Pyrene	12.76	202	1257756	72.40	ng	99
80) Benzo(a)anthracene	13.95	228	901247	67.55	ng	# 92
82) Chrysene	13.99	228	758511m	60.43	ng	
85) Indeno(1,2,3-cd)pyrene	16.73	276	367292	31.50	ng	# 100
87) Benzo(b)fluoranthene	14.98	252	1222979m	79.63	ng	
88) Benzo(k)fluoranthene	14.99	252	268623m	23.19	ng	
89) Benzo(a)pyrene	15.31	252	646821	52.05	ng	98
90) Dibenzo(a,h)anthracene	16.74	278	99600	8.63	ng	96
91) Benzo(a,h,i)perylene	17.14	276	300288	25.29	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF061016\
Data File : BF087883.D
Acq On : 10 Jun 2016 19:20
Operator : UM/SJ
Sample : H3395-01
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
STA-928-PLATFORM(0-4)

Manual Integrations
APPROVED

SOHIL
6/13/2016 7:11:13 PM

Quant Time: Jun 11 02:12:14 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060716.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 07 18:51:55 2016
Response via : Initial Calibration

