

Data Path : Z:\HPCHEM1\BNA F\DATA\BF061017\
 Data File : BF095871.D
 Acq On : 11 Jun 2017 5:17
 Operator : SJ/MA
 Sample : I3591-01
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 E2-Q11-ESW

Manual Integrations
 APPROVED

mohammad
 6/12/2017 1:28:50 PM

Quant Time: Jun 12 07:14:00 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 05 16:15:31 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.37	152	59664	20.00	ng	-0.03
21) Naphthalene-d8	9.40	136	236766	20.00	ng	-0.04
38) Acenaphthene-d10	12.22	164	94366	20.00	ng	-0.04
63) Phenanthrene-d10	14.62	188	149176	20.00	ng	-0.03
75) Chrysene-d12	18.33	240	132369	20.00	ng	-0.02
86) Perylene-d12	20.00	264	93439	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.26	112	351748	93.94	ng	-0.02
7) Phenol-d6	6.93	99	450202	100.43	ng	-0.03
23) Nitrobenzene-d5	8.28	82	271544	71.23	ng	-0.04
41) 2,4,6-Tribromophenol	13.52	330	61241	73.35	ng	-0.04
44) 2-Fluorobiphenyl	11.18	172	532184	78.48	ng	-0.04
78) Terphenyl-d14	16.99	244	357860	67.09	ng	-0.02
Target Compounds						
49) Dimethylphthalate	11.87	163	65436	9.13	ng	# 99
82) Chrysene	18.37	228	17155	2.23	ng	# 93
87) Benzo(b)fluoranthene	19.61	252	13253m	2.27	ng	
89) Benzo(a)pyrene	19.95	252	12242	2.28	ng	98
91) Benzo(g,h,i)perylene	21.53	276	10337	2.22	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF061017\
Data File : BF095871.D
Acq On : 11 Jun 2017 5:17
Operator : SJ/MA
Sample : I3591-01
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
E2-Q11-ESW

Manual Integrations
APPROVED

mohammad
6/12/2017 1:28:50 PM

Quant Time: Jun 12 07:14:00 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 05 16:15:31 2017
Response via : Initial Calibration

