



Data Path : Z:\HPCHEM1\BNA F\Data\BF061017\
Data File : BF095867.D
Acq On : 11 Jun 2017 3:09
Operator : SJ/MA
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 12 07:56:54 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 05 16:15:31 2017
Response via : Initial Calibration



TIC: BF095867.D

(69) Pentachlorophenol (C)



14.398min (+0.001) 0.00ng

response 7078

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	73.51
264.00	64.60	62.78
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF061017\
 Data File : BF095867.D
 Acq On : 11 Jun 2017 3:09
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 12 07:56:54 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 05 16:15:31 2017
 Response via : Initial Calibration



TIC: BF095867.D

(76) Benzidine

16.668min (-0.018) 0.00ng

response 147856

Ion	Exp%	Act%
184.00	100	100
185.00	19.40	15.41#
183.00	12.20	11.26
0.00	0.00	0.00