

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061021\
 Data File : BF124225.D
 Acq On : 10 Jun 2021 9:48
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 10 11:31:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 10 01:53:56 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	87	0.00
2	1,4-Dioxane	40.000	35.793	10.5	80	0.00
3	Pyridine	40.000	35.567	11.1	80	0.00
4	n-Nitrosodimethylamine	40.000	34.289	14.3	77	0.00
5 S	2-Fluorophenol	80.000	75.266	5.9	84	0.00
6	Aniline	40.000	35.528	11.2	82	0.00
7 S	Phenol-d6	80.000	74.273	7.2	84	0.00
8	2-Chlorophenol	40.000	36.223	9.4	80	0.00
9	Benzaldehyde	40.000	42.161	-5.4	99	0.00
10 C	Phenol	40.000	35.601	11.0	81	0.00
11	bis(2-Chloroethyl)ether	40.000	36.551	8.6	81	0.00
12	1,3-Dichlorobenzene	40.000	37.642	5.9	85	0.00
13 C	1,4-Dichlorobenzene	40.000	37.301	6.7	84	0.00
14	1,2-Dichlorobenzene	40.000	37.732	5.7	85	0.00
15	Benzyl Alcohol	40.000	37.743	5.6	87	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	33.287	16.8	77	0.00
17	2-Methylphenol	40.000	37.806	5.5	86	0.00
18	Hexachloroethane	40.000	37.926	5.2	85	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.993	7.5	88	0.00
20	3+4-Methylphenols	40.000	36.651	8.4	83	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	86	0.00
22	Acetophenone	40.000	40.147	-0.4	89	0.00
23 S	Nitrobenzene-d5	80.000	77.710	2.9	88	0.00
24	Nitrobenzene	40.000	38.062	4.8	83	0.00
25	Isophorone	40.000	36.709	8.2	83	0.00
26 C	2-Nitrophenol	40.000	39.271	1.8	87	0.00
27	2,4-Dimethylphenol	40.000	36.999	7.5	82	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.272	6.8	87	0.00
29 C	2,4-Dichlorophenol	40.000	39.506	1.2	89	0.00
30	1,2,4-Trichlorobenzene	40.000	38.816	3.0	88	0.00
31	Naphthalene	40.000	36.960	7.6	83	0.00
32	Benzoic acid	40.000	36.152	9.6	86	0.00
33	4-Chloroaniline	40.000	40.472	-1.2	91	0.00
34 C	Hexachlorobutadiene	40.000	38.396	4.0	86	0.00
35	Caprolactam	40.000	38.938	2.7	89	0.00
36 C	4-Chloro-3-methylphenol	40.000	36.940	7.7	87	0.00
37	2-Methylnaphthalene	40.000	37.965	5.1	84	0.00
38	1-Methylnaphthalene	40.000	38.643	3.4	86	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	83	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.796	-2.0	92	0.00
41 P	Hexachlorocyclopentadiene	40.000	34.760	13.1	78	0.00
42 S	2,4,6-Tribromophenol	80.000	80.344	-0.4	89	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.782	3.0	84	0.00
44	2,4,5-Trichlorophenol	40.000	38.366	4.1	85	0.00
45 S	2-Fluorobiphenyl	80.000	79.512	0.6	88	0.00
46	1,1'-Biphenyl	40.000	39.832	0.4	90	0.00
47	2-Chloronaphthalene	40.000	38.094	4.8	87	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061021\
 Data File : BF124225.D
 Acq On : 10 Jun 2021 9:48
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 10 11:31:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 10 01:53:56 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.150	-0.4	89	0.00
49	Acenaphthylene	40.000	38.867	2.8	86	0.00
50	Dimethylphthalate	40.000	38.006	5.0	88	0.00
51	2,6-Dinitrotoluene	40.000	37.778	5.6	87	0.00
52 C	Acenaphthene	40.000	38.665	3.3	87	0.00
53	3-Nitroaniline	40.000	37.209	7.0	84	0.00
54 P	2,4-Dinitrophenol	40.000	41.718	-4.3	97	0.00
55	Dibenzofuran	40.000	38.845	2.9	88	0.00
56 P	4-Nitrophenol	40.000	47.596	-19.0	101	0.00
57	2,4-Dinitrotoluene	40.000	38.905	2.7	83	0.00
58	Fluorene	40.000	39.421	1.4	86	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.905	2.7	89	0.00
60	Diethylphthalate	40.000	39.246	1.9	87	0.00
61	4-Chlorophenyl-phenylether	40.000	38.293	4.3	87	0.00
62	4-Nitroaniline	40.000	39.393	1.5	86	0.00
63	Azobenzene	40.000	37.074	7.3	85	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	84	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.746	0.6	90	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.099	4.8	87	0.00
67	4-Bromophenyl-phenylether	40.000	38.493	3.8	91	0.00
68	Hexachlorobenzene	40.000	36.836	7.9	85	0.00
69	Atrazine	40.000	43.945	-9.9	91	0.00
70 C	Pentachlorophenol	40.000	36.743	8.1	85	0.00
71	Phenanthrene	40.000	38.254	4.4	87	0.00
72	Anthracene	40.000	37.197	7.0	87	0.00
73	Carbazole	40.000	37.464	6.3	86	0.00
74	Di-n-butylphthalate	40.000	38.680	3.3	89	0.00
75 C	Fluoranthene	40.000	36.505	8.7	83	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	78	0.00
77	Benzydine	40.000	41.332	-3.3	76	0.00
78	Pyrene	40.000	40.025	-0.1	83	0.00
79 S	Terphenyl-d14	80.000	80.313	-0.4	82	0.00
80	Butylbenzylphthalate	40.000	39.238	1.9	81	0.00
81	Benzo(a)anthracene	40.000	37.371	6.6	79	0.00
82	3,3'-Dichlorobenzidine	40.000	37.895	5.3	80	0.00
83	Chrysene	40.000	37.696	5.8	79	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	37.505	6.2	77	0.00
85 c	Di-n-octyl phthalate	40.000	37.893	5.3	80	0.00
86 I	Perylene-d12	20.000	20.000	0.0	78	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.360	1.6	85	0.01
88	Benzo(b)fluoranthene	40.000	37.473	6.3	81	0.00
89	Benzo(k)fluoranthene	40.000	38.548	3.6	81	0.00
90 C	Benzo(a)pyrene	40.000	38.969	2.6	83	0.00
91	Dibenzo(a,h)anthracene	40.000	38.274	4.3	83	0.01
92	Benzo(g,h,i)perylene	40.000	37.842	5.4	80	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061021\
Data File : BF124225.D
Acq On : 10 Jun 2021 9:48
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 10 11:31:29 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 10 01:53:56 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0