

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061021\
 Data File : BF124236.D
 Acq On : 10 Jun 2021 16:58
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 10 18:02:46 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 10 01:53:56 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.581	0.510	12.2	84	0.00
3	Pyridine	1.489	1.371	7.9	88	0.00
4	n-Nitrosodimethylamine	0.885	0.780	11.9	84	0.00
5 S	2-Fluorophenol	1.329	1.308	1.6	94	0.00
6	Aniline	2.040	1.856	9.0	89	0.00
7 S	Phenol-d6	1.786	1.707	4.4	93	0.00
8	2-Chlorophenol	1.478	1.420	3.9	91	0.00
9	Benzaldehyde	0.794	0.869	-9.4	110	0.00
10 C	Phenol	1.930	1.789	7.3	90	0.00
11	bis(2-Chloroethyl)ether	1.407	1.331	5.4	90	0.00
12	1,3-Dichlorobenzene	1.728	1.655	4.2	92	0.00
13 C	1,4-Dichlorobenzene	1.734	1.647	5.0	91	0.00
14	1,2-Dichlorobenzene	1.658	1.558	6.0	90	0.00
15	Benzyl Alcohol	1.610	1.492	7.3	91	0.00
16	2,2'-oxybis(1-Chloropropane	2.195	1.855	15.5	83	0.00
17	2-Methylphenol	1.202	1.158	3.7	93	0.00
18	Hexachloroethane	0.683	0.677	0.9	95	0.00
19 P	n-Nitroso-di-n-propylamine	1.261	1.185	6.0	95	0.00
20	3+4-Methylphenols	1.592	1.522	4.4	92	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
22	Acetophenone	0.570	0.559	1.9	96	0.00
23 S	Nitrobenzene-d5	0.504	0.483	4.2	96	0.00
24	Nitrobenzene	0.506	0.460	9.1	88	0.00
25	Isophorone	0.909	0.844	7.2	93	0.00
26 C	2-Nitrophenol	0.226	0.218	3.5	95	0.00
27	2,4-Dimethylphenol	0.319	0.300	6.0	93	0.00
28	bis(2-Chloroethoxy)methane	0.450	0.429	4.7	98	0.00
29 C	2,4-Dichlorophenol	0.369	0.347	6.0	94	0.00
30	1,2,4-Trichlorobenzene	0.413	0.392	5.1	96	0.00
31	Naphthalene	1.196	1.100	8.0	92	0.00
32	Benzoic acid	0.196	0.196	0.0	105	0.00
33	4-Chloroaniline	0.445	0.430	3.4	96	0.00
34 C	Hexachlorobutadiene	0.292	0.264	9.6	90	0.00
35	Caprolactam	0.101	0.093	7.9	93	0.00
36 C	4-Chloro-3-methylphenol	0.403	0.379	6.0	98	0.00
37	2-Methylnaphthalene	0.839	0.776	7.5	91	0.00
38	1-Methylnaphthalene	0.784	0.748	4.6	94	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
40	1,2,4,5-Tetrachlorobenzene	0.709	0.703	0.8	102	0.00
41 P	Hexachlorocyclopentadiene	0.314	0.274	12.7	84	0.00
42 S	2,4,6-Tribromophenol	0.283	0.263	7.1	94	0.00
43 C	2,4,6-Trichlorophenol	0.462	0.443	4.1	94	0.00
44	2,4,5-Trichlorophenol	0.496	0.459	7.5	93	0.00
45 S	2-Fluorobiphenyl	1.589	1.525	4.0	97	0.00
46	1,1'-Biphenyl	1.688	1.632	3.3	99	0.00
47	2-Chloronaphthalene	1.374	1.307	4.9	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061021\
 Data File : BF124236.D
 Acq On : 10 Jun 2021 16:58
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 10 18:02:46 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 10 01:53:56 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.496	0.461	7.1	94	0.00
49	Acenaphthylene	1.998	1.839	8.0	93	0.00
50	Dimethylphthalate	1.655	1.530	7.6	97	0.00
51	2,6-Dinitrotoluene	0.379	0.357	5.8	99	0.00
52 C	Acenaphthene	1.305	1.188	9.0	93	0.00
53	3-Nitroaniline	0.355	0.317	10.7	92	0.00
54 P	2,4-Dinitrophenol	0.160	0.197	-23.1	112	0.00
55	Dibenzofuran	2.042	1.863	8.8	93	0.00
56 P	4-Nitrophenol	0.254	0.276	-8.7	105	0.00
57	2,4-Dinitrotoluene	0.500	0.476	4.8	93	0.00
58	Fluorene	1.567	1.438	8.2	91	0.00
59	2,3,4,6-Tetrachlorophenol	0.402	0.370	8.0	95	0.00
60	Diethylphthalate	1.670	1.576	5.6	96	0.00
61	4-Chlorophenyl-phenylether	0.801	0.756	5.6	98	0.00
62	4-Nitroaniline	0.357	0.312	12.6	87	0.00
63	Azobenzene	1.741	1.574	9.6	94	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
65	4,6-Dinitro-2-methylphenol	0.161	0.160	0.6	96	0.00
66 c	n-Nitrosodiphenylamine	0.760	0.725	4.6	92	0.00
67	4-Bromophenyl-phenylether	0.275	0.265	3.6	97	0.00
68	Hexachlorobenzene	0.311	0.299	3.9	94	0.00
69	Atrazine	0.215	0.231	-7.4	94	0.00
70 C	Pentachlorophenol	0.146	0.141	3.4	91	0.00
71	Phenanthrene	1.249	1.198	4.1	92	0.00
72	Anthracene	1.258	1.202	4.5	95	0.00
73	Carbazole	1.096	1.002	8.6	89	0.00
74	Di-n-butylphthalate	1.409	1.345	4.5	93	0.00
75 C	Fluoranthene	1.311	1.169	10.8	86	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	82	0.00
77	Benzydine	0.429	0.433	-0.9	72	0.00
78	Pyrene	1.921	1.911	0.5	86	0.00
79 S	Terphenyl-d14	1.457	1.521	-4.4	90	0.00
80	Butylbenzylphthalate	0.779	0.804	-3.2	90	0.00
81	Benzo(a)anthracene	1.457	1.422	2.4	86	0.00
82	3,3'-Dichlorobenzidine	0.427	0.408	4.4	84	0.00
83	Chrysene	1.406	1.343	4.5	84	0.00
84	Bis(2-ethylhexyl)phthalate	0.927	0.955	-3.0	89	0.00
85 c	Di-n-octyl phthalate	1.237	1.319	-6.6	94	0.00
86 I	Perylene-d12	1.000	1.000	0.0	91	0.00
87	Indeno(1,2,3-cd)pyrene	1.448	1.490	-2.9	92	0.01
88	Benzo(b)fluoranthene	1.554	1.500	3.5	98	0.00
89	Benzo(k)fluoranthene	1.471	1.295	12.0	87	0.00
90 C	Benzo(a)pyrene	1.349	1.259	6.7	92	0.00
91	Dibenzo(a,h)anthracene	1.252	1.258	-0.5	92	0.00
92	Benzo(g,h,i)perylene	1.235	1.243	-0.6	90	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061021\
Data File : BF124236.D
Acq On : 10 Jun 2021 16:58
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 10 18:02:46 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 10 01:53:56 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0