

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061021\
 Data File : BF124236.D
 Acq On : 10 Jun 2021 16:58
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 10 18:02:46 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 10 01:53:56 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	93	0.00
2	1,4-Dioxane	40.000	35.083	12.3	84	0.00
3	Pyridine	40.000	36.819	8.0	88	0.00
4	n-Nitrosodimethylamine	40.000	35.271	11.8	84	0.00
5 S	2-Fluorophenol	80.000	78.742	1.6	94	0.00
6	Aniline	40.000	36.396	9.0	89	0.00
7 S	Phenol-d6	80.000	76.465	4.4	93	0.00
8	2-Chlorophenol	40.000	38.429	3.9	91	0.00
9	Benzaldehyde	40.000	43.768	-9.4	110	0.00
10 C	Phenol	40.000	37.084	7.3	90	0.00
11	bis(2-Chloroethyl)ether	40.000	37.853	5.4	90	0.00
12	1,3-Dichlorobenzene	40.000	38.301	4.2	92	0.00
13 C	1,4-Dichlorobenzene	40.000	38.002	5.0	91	0.00
14	1,2-Dichlorobenzene	40.000	37.581	6.0	90	0.00
15	Benzyl Alcohol	40.000	37.069	7.3	91	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	33.807	15.5	83	0.00
17	2-Methylphenol	40.000	38.547	3.6	93	0.00
18	Hexachloroethane	40.000	39.688	0.8	95	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.593	6.0	95	0.00
20	3+4-Methylphenols	40.000	38.234	4.4	92	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	95	0.00
22	Acetophenone	40.000	39.221	1.9	96	0.00
23 S	Nitrobenzene-d5	80.000	76.569	4.3	96	0.00
24	Nitrobenzene	40.000	36.372	9.1	88	0.00
25	Isophorone	40.000	37.172	7.1	93	0.00
26 C	2-Nitrophenol	40.000	38.639	3.4	95	0.00
27	2,4-Dimethylphenol	40.000	37.676	5.8	93	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.087	4.8	98	0.00
29 C	2,4-Dichlorophenol	40.000	37.647	5.9	94	0.00
30	1,2,4-Trichlorobenzene	40.000	37.951	5.1	96	0.00
31	Naphthalene	40.000	36.802	8.0	92	0.00
32	Benzoic acid	40.000	40.144	-0.4	105	0.00
33	4-Chloroaniline	40.000	38.652	3.4	96	0.00
34 C	Hexachlorobutadiene	40.000	36.129	9.7	90	0.00
35	Caprolactam	40.000	36.578	8.6	93	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.585	6.0	98	0.00
37	2-Methylnaphthalene	40.000	37.040	7.4	91	0.00
38	1-Methylnaphthalene	40.000	38.165	4.6	94	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	94	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.690	0.8	102	0.00
41 P	Hexachlorocyclopentadiene	40.000	33.436	16.4	84	0.00
42 S	2,4,6-Tribromophenol	80.000	74.204	7.2	94	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.365	4.1	94	0.00
44	2,4,5-Trichlorophenol	40.000	37.037	7.4	93	0.00
45 S	2-Fluorobiphenyl	80.000	76.753	4.1	97	0.00
46	1,1'-Biphenyl	40.000	38.666	3.3	99	0.00
47	2-Chloronaphthalene	40.000	38.048	4.9	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061021\
 Data File : BF124236.D
 Acq On : 10 Jun 2021 16:58
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 10 18:02:46 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 10 01:53:56 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.215	7.0	94	0.00
49	Acenaphthylene	40.000	36.823	7.9	93	0.00
50	Dimethylphthalate	40.000	36.995	7.5	97	0.00
51	2,6-Dinitrotoluene	40.000	37.609	6.0	99	0.00
52 C	Acenaphthene	40.000	36.431	8.9	93	0.00
53	3-Nitroaniline	40.000	35.699	10.8	92	0.00
54 P	2,4-Dinitrophenol	40.000	42.291	-5.7	112	0.00
55	Dibenzofuran	40.000	36.487	8.8	93	0.00
56 P	4-Nitrophenol	40.000	43.452	-8.6	105	0.00
57	2,4-Dinitrotoluene	40.000	38.083	4.8	93	0.00
58	Fluorene	40.000	36.698	8.3	91	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	36.812	8.0	95	0.00
60	Diethylphthalate	40.000	37.751	5.6	96	0.00
61	4-Chlorophenyl-phenylether	40.000	37.717	5.7	98	0.00
62	4-Nitroaniline	40.000	34.982	12.5	87	0.00
63	Azobenzene	40.000	36.176	9.6	94	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	89	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.926	0.2	96	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.163	4.6	92	0.00
67	4-Bromophenyl-phenylether	40.000	38.601	3.5	97	0.00
68	Hexachlorobenzene	40.000	38.504	3.7	94	0.00
69	Atrazine	40.000	43.016	-7.5	94	0.00
70 C	Pentachlorophenol	40.000	37.180	7.1	91	0.00
71	Phenanthrene	40.000	38.343	4.1	92	0.00
72	Anthracene	40.000	38.203	4.5	95	0.00
73	Carbazole	40.000	36.545	8.6	89	0.00
74	Di-n-butylphthalate	40.000	38.173	4.6	93	0.00
75 C	Fluoranthene	40.000	35.654	10.9	86	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	82	0.00
77	Benzydine	40.000	38.091	4.8	72	0.00
78	Pyrene	40.000	39.796	0.5	86	0.00
79 S	Terphenyl-d14	80.000	83.552	-4.4	90	0.00
80	Butylbenzylphthalate	40.000	41.271	-3.2	90	0.00
81	Benzo(a)anthracene	40.000	39.038	2.4	86	0.00
82	3,3'-Dichlorobenzidine	40.000	38.159	4.6	84	0.00
83	Chrysene	40.000	38.213	4.5	84	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.222	-3.1	89	0.00
85 c	Di-n-octyl phthalate	40.000	42.647	-6.6	94	0.00
86 I	Perylene-d12	20.000	20.000	0.0	91	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.905	7.7	92	0.01
88	Benzo(b)fluoranthene	40.000	38.627	3.4	98	0.00
89	Benzo(k)fluoranthene	40.000	35.215	12.0	87	0.00
90 C	Benzo(a)pyrene	40.000	37.333	6.7	92	0.00
91	Dibenzo(a,h)anthracene	40.000	36.506	8.7	92	0.00
92	Benzo(g,h,i)perylene	40.000	36.453	8.9	90	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061021\
Data File : BF124236.D
Acq On : 10 Jun 2021 16:58
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 10 18:02:46 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 10 01:53:56 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0