

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061022\
 Data File : BF128755.D
 Acq On : 10 Jun 2022 12:18
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 11 01:26:20 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 08 14:28:39 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	85	0.00
2	1,4-Dioxane	40.000	39.207	2.0	79	-0.03
3	Pyridine	40.000	39.282	1.8	80	-0.03
4	n-Nitrosodimethylamine	40.000	35.197	12.0	71	-0.02
5 S	2-Fluorophenol	80.000	78.761	1.5	82	-0.01
6	Aniline	40.000	41.080	-2.7	88	0.00
7 S	Phenol-d6	80.000	82.361	-3.0	87	-0.01
8	2-Chlorophenol	40.000	41.403	-3.5	88	0.00
9	Benzaldehyde	40.000	44.427	-11.1	116	0.00
10 C	Phenol	40.000	40.679	-1.7	85	-0.01
11	bis(2-Chloroethyl)ether	40.000	40.717	-1.8	86	0.00
12	1,3-Dichlorobenzene	40.000	40.545	-1.4	86	0.00
13 C	1,4-Dichlorobenzene	40.000	40.837	-2.1	87	0.00
14	1,2-Dichlorobenzene	40.000	41.315	-3.3	89	-0.01
15	Benzyl Alcohol	40.000	40.628	-1.6	87	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.541	6.1	82	0.00
17	2-Methylphenol	40.000	41.644	-4.1	90	-0.01
18	Hexachloroethane	40.000	41.348	-3.4	86	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.065	-2.7	87	0.00
20	3+4-Methylphenols	40.000	41.608	-4.0	88	-0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	91	0.00
22	Acetophenone	40.000	39.732	0.7	87	0.00
23 S	Nitrobenzene-d5	80.000	79.648	0.4	88	0.00
24	Nitrobenzene	40.000	39.826	0.4	88	-0.01
25	Isophorone	40.000	38.796	3.0	86	0.00
26 C	2-Nitrophenol	40.000	42.180	-5.4	89	0.00
27	2,4-Dimethylphenol	40.000	40.934	-2.3	87	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.365	1.6	86	-0.01
29 C	2,4-Dichlorophenol	40.000	39.691	0.8	85	-0.01
30	1,2,4-Trichlorobenzene	40.000	39.951	0.1	87	0.00
31	Naphthalene	40.000	40.308	-0.8	92	0.00
32	Benzoic acid	40.000	26.124	34.7#	55	0.00
33	4-Chloroaniline	40.000	40.308	-0.8	93	-0.01
34 C	Hexachlorobutadiene	40.000	39.966	0.1	92	0.00
35	Caprolactam	40.000	41.247	-3.1	94	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.635	0.9	91	0.00
37	2-Methylnaphthalene	40.000	40.158	-0.4	92	0.00
38	1-Methylnaphthalene	40.000	40.214	-0.5	94	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	92	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.742	-4.4	93	0.00
41 P	Hexachlorocyclopentadiene	40.000	32.679	18.3	70	-0.01
42 S	2,4,6-Tribromophenol	80.000	81.773	-2.2	88	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.464	3.8	86	0.00
44	2,4,5-Trichlorophenol	40.000	39.262	1.8	87	0.00
45 S	2-Fluorobiphenyl	80.000	81.935	-2.4	92	0.00
46	1,1'-Biphenyl	40.000	40.523	-1.3	92	0.00
47	2-Chloronaphthalene	40.000	40.613	-1.5	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061022\
 Data File : BF128755.D
 Acq On : 10 Jun 2022 12:18
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 11 01:26:20 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 08 14:28:39 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.710	-4.3	93	0.00
49	Acenaphthylene	40.000	40.038	-0.1	90	0.00
50	Dimethylphthalate	40.000	40.677	-1.7	91	0.00
51	2,6-Dinitrotoluene	40.000	42.241	-5.6	92	0.00
52 C	Acenaphthene	40.000	40.486	-1.2	90	-0.01
53	3-Nitroaniline	40.000	41.874	-4.7	91	0.00
54 P	2,4-Dinitrophenol	40.000	35.483	11.3	76	0.00
55	Dibenzofuran	40.000	40.538	-1.3	92	0.00
56 P	4-Nitrophenol	40.000	35.406	11.5	81	0.00
57	2,4-Dinitrotoluene	40.000	43.739	-9.3	94	0.00
58	Fluorene	40.000	40.858	-2.1	90	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	36.874	7.8	81	0.00
60	Diethylphthalate	40.000	40.170	-0.4	90	0.00
61	4-Chlorophenyl-phenylether	40.000	40.686	-1.7	90	-0.01
62	4-Nitroaniline	40.000	41.439	-3.6	91	0.00
63	Azobenzene	40.000	40.140	-0.4	90	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	90	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.362	-5.9	91	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.408	-3.5	91	0.00
67	4-Bromophenyl-phenylether	40.000	42.189	-5.5	93	0.00
68	Hexachlorobenzene	40.000	41.019	-2.5	89	0.00
69	Atrazine	40.000	39.581	1.0	86	0.00
70 C	Pentachlorophenol	40.000	41.369	-3.4	88	0.00
71	Phenanthrene	40.000	40.412	-1.0	90	0.00
72	Anthracene	40.000	40.508	-1.3	90	0.00
73	Carbazole	40.000	40.258	-0.6	90	0.00
74	Di-n-butylphthalate	40.000	40.154	-0.4	87	0.00
75 C	Fluoranthene	40.000	40.444	-1.1	87	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	77	0.00
77	Benzydine	40.000	31.274	21.8	65	0.00
78	Pyrene	40.000	43.676	-9.2	87	0.00
79 S	Terphenyl-d14	80.000	86.769	-8.5	82	0.00
80	Butylbenzylphthalate	40.000	42.893	-7.2	80	0.00
81	Benzo(a)anthracene	40.000	41.160	-2.9	78	0.00
82	3,3'-Dichlorobenzidine	40.000	42.709	-6.8	82	0.00
83	Chrysene	40.000	40.661	-1.7	78	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.450	-6.1	80	0.00
85 c	Di-n-octyl phthalate	40.000	40.749	-1.9	76	0.00
86 I	Perylene-d12	20.000	20.000	0.0	77	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.068	-5.2	89	0.00
88	Benzo(b)fluoranthene	40.000	41.782	-4.5	78	0.00
89	Benzo(k)fluoranthene	40.000	39.638	0.9	74	0.00
90 C	Benzo(a)pyrene	40.000	40.594	-1.5	77	0.00
91	Dibenzo(a,h)anthracene	40.000	42.013	-5.0	87	0.00
92	Benzo(g,h,i)perylene	40.000	41.399	-3.5	88	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061022\
Data File : BF128755.D
Acq On : 10 Jun 2022 12:18
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 11 01:26:20 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 08 14:28:39 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0