


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061022\
 Data File : BF128754.D
 Acq On : 10 Jun 2022 11:20
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jun 08 14:28:39 2022






AutoFind: Scans 1581, 1582, 1583; Background Corrected with Scan 1574

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.8	50015	PASS
68	69	0.00	2	1.8	895	PASS
69	198	0.00	100	35.1	48927	PASS
70	69	0.00	2	0.6	293	PASS
127	198	10	80	45.5	63489	PASS
197	198	0.00	2	0.1	146	PASS
198	198	100	100	100.0	139541	PASS
199	198	5	9	6.9	9660	PASS
275	198	10	60	28.5	39714	PASS
365	198	1	100	4.4	6175	PASS
441	198	0.01	100	17.3	24206	PASS
442	442	50	100	100.0	151125	PASS
443	442	15	24	18.9	28589	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061022\
Data File : BF128754.D
Acq On : 10 Jun 2022 11:20
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Jun 11 06:55:15 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 08 14:28:39 2022
Response via : Initial Calibration



TIC: BF128754.D\data.ms

(70) Pentachlorophenol (C)

11.145min (-0.006) 15946.90 ng




response 76701

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.70	63.45
264.00	61.80	62.40
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061022\
 Data File : BF128754.D
 Acq On : 10 Jun 2022 11:20
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Jun 11 06:55:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 08 14:28:39 2022
 Response via : Initial Calibration



TIC: BF128754.D\data.ms

(77) Benzidine

12.680min (-0.000) 0.00 ng

response 727054

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.90	14.50
183.00	11.60	11.94
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
6/10/2022	BNA_F	<u>BF128754.D</u>
Compound Name	Response	Retention Time
DDT	355987	13.498
DDD	6950	13.198
DDE	1678	12.857
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
8628	364615	2.37