

Data Path : Z:\HPCHEM1\BNA F\DATA\BF061116\
 Data File : BF087924.D
 Acq On : 12 Jun 2016 2:55
 Operator : UM/SJ
 Sample : PB91188BL
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB91188BL

Quant Time: Jun 12 06:03:33 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 07 18:51:55 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.80	152	102914	20.00	ng	-0.05
21) Naphthalene-d8	8.09	136	409183	20.00	ng	-0.03
38) Acenaphthene-d10	9.84	164	200354	20.00	ng	-0.05
63) Phenanthrene-d10	11.32	188	402963	20.00	ng	-0.03
75) Chrysene-d12	13.95	240	330122	20.00	ng	-0.03
86) Perylene-d12	15.37	264	251507	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.39	112	638212	106.02	ng	-0.03
7) Phenol-d6	6.43	99	736825	100.99	ng	-0.05
23) Nitrobenzene-d5	7.37	82	579779	82.74	ng	-0.05
41) 2,4,6-Tribromophenol	10.63	330	215238	101.74	ng	-0.05
44) 2-Fluorobiphenyl	9.16	172	1004380	78.32	ng	-0.05
78) Terphenyl-d14	12.91	244	942461	64.96	ng	-0.03

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF061116\
Data File : BF087924.D
Acq On : 12 Jun 2016 2:55
Operator : UM/SJ
Sample : PB91188BL
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB91188BL

Quant Time: Jun 12 06:03:33 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060716.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 07 18:51:55 2016
Response via : Initial Calibration

