


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF061116\
 Data File : BF087922.D
 Acq On : 12 Jun 2016 1:56
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF060716.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jun 13 14:10:49 2016



AutoFind: Scans 874, 875, 876; Background Corrected with Scan 871

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.7	12594	PASS
68	69	0.00	2	1.4	179	PASS
69	198	0.00	100	38.5	12530	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	54.3	17689	PASS
197	198	0.00	2	0.4	127	PASS
198	198	100	100	100.0	32547	PASS
199	198	5	9	6.4	2095	PASS
275	198	10	60	24.7	8034	PASS
365	198	1	100	2.8	917	PASS
441	442	0.01	24	15.5	4335	PASS
442	198	50	100	85.9	27962	PASS
443	442	15	24	19.3	5396	PASS