

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061120\
 Data File : BF120596.D
 Acq On : 11 Jun 2020 10:01
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 11 12:36:05 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 11 11:28:22 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	0.00
2	1,4-Dioxane	0.553	0.518	6.3	101	0.00
3	Pyridine	1.554	1.452	6.6	99	0.00
4	n-Nitrosodimethylamine	0.645	0.610	5.4	100	0.00
5 S	2-Fluorophenol	1.167	1.178	-0.9	111	0.00
6	Aniline	1.859	1.925	-3.6	106	0.00
7 S	Phenol-d6	1.453	1.517	-4.4	108	0.00
8	2-Chlorophenol	1.318	1.279	3.0	103	0.00
9	Benzaldehyde	0.919	0.942	-2.5	108	0.00
10 C	Phenol	1.550	1.663	-7.3	110	0.00
11	bis(2-Chloroethyl)ether	1.291	1.256	2.7	104	0.00
12	1,3-Dichlorobenzene	1.401	1.392	0.6	107	0.00
13 C	1,4-Dichlorobenzene	1.383	1.400	-1.2	114	0.00
14	1,2-Dichlorobenzene	1.347	1.348	-0.1	110	0.00
15	Benzyl Alcohol	1.030	1.076	-4.5	111	0.00
16	2,2'-oxybis(1-Chloropropane	1.859	1.999	-7.5	118	0.00
17	2-Methylphenol	1.124	1.178	-4.8	123	0.00
18	Hexachloroethane	0.510	0.523	-2.5	114	0.00
19 P	n-Nitroso-di-n-propylamine	0.843	0.861	-2.1	114	0.00
20	3+4-Methylphenols	1.405	1.336	4.9	110	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	116	0.00
22	Acetophenone	0.447	0.439	1.8	112	0.00
23 S	Nitrobenzene-d5	0.358	0.349	2.5	110	0.00
24	Nitrobenzene	0.374	0.374	0.0	118	0.00
25	Isophorone	0.697	0.679	2.6	110	0.00
26 C	2-Nitrophenol	0.198	0.193	2.5	106	0.00
27	2,4-Dimethylphenol	0.292	0.284	2.7	107	0.00
28	bis(2-Chloroethoxy)methane	0.404	0.410	-1.5	104	0.00
29 C	2,4-Dichlorophenol	0.296	0.292	1.4	101	0.00
30	1,2,4-Trichlorobenzene	0.329	0.324	1.5	102	0.00
31	Naphthalene	0.982	0.994	-1.2	108	0.00
32	Benzoic acid	0.224	0.222	0.9	101	0.00
33	4-Chloroaniline	0.451	0.450	0.2	110	0.00
34 C	Hexachlorobutadiene	0.192	0.199	-3.6	118	0.00
35	Caprolactam	0.095	0.094	1.1	117	0.00
36 C	4-Chloro-3-methylphenol	0.296	0.304	-2.7	118	0.00
37	2-Methylnaphthalene	0.641	0.669	-4.4	118	0.00
38	1-Methylnaphthalene	0.603	0.616	-2.2	114	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	127	0.00
40	1,2,4,5-Tetrachlorobenzene	0.597	0.593	0.7	115	0.00
41 P	Hexachlorocyclopentadiene	0.335	0.335	0.0	117	0.00
42 S	2,4,6-Tribromophenol	0.233	0.234	-0.4	115	0.00
43 C	2,4,6-Trichlorophenol	0.424	0.415	2.1	114	0.00
44	2,4,5-Trichlorophenol	0.481	0.470	2.3	113	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061120\
 Data File : BF120596.D
 Acq On : 11 Jun 2020 10:01
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 11 12:36:05 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 11 11:28:22 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.307	1.190	9.0	113	0.00
46	1,1'-Biphenyl	1.493	1.498	-0.3	116	0.00
47	2-Chloronaphthalene	1.221	1.222	-0.1	116	0.00
48	2-Nitroaniline	0.384	0.397	-3.4	121	0.00
49	Acenaphthylene	1.885	1.924	-2.1	117	0.00
50	Dimethylphthalate	1.440	1.439	0.1	117	0.00
51	2,6-Dinitrotoluene	0.326	0.330	-1.2	117	0.00
52 C	Acenaphthene	1.124	1.130	-0.5	116	0.00
53	3-Nitroaniline	0.392	0.388	1.0	116	0.00
54 P	2,4-Dinitrophenol	0.175	0.169	3.4	116	0.00
55	Dibenzofuran	1.724	1.763	-2.3	118	0.00
56 P	4-Nitrophenol	0.295	0.285	3.4	114	0.00
57	2,4-Dinitrotoluene	0.433	0.443	-2.3	119	0.00
58	Fluorene	1.329	1.307	1.7	117	0.00
59	2,3,4,6-Tetrachlorophenol	0.378	0.379	-0.3	117	0.00
60	Diethylphthalate	1.419	1.469	-3.5	119	0.00
61	4-Chlorophenyl-phenylether	0.686	0.670	2.3	115	0.00
62	4-Nitroaniline	0.389	0.386	0.8	115	0.00
63	Azobenzene	1.288	1.377	-6.9	123	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	124	0.00
65	4,6-Dinitro-2-methylphenol	0.120	0.121	-0.8	115	0.00
66 c	n-Nitrosodiphenylamine	0.614	0.629	-2.4	118	0.00
67	4-Bromophenyl-phenylether	0.227	0.231	-1.8	115	0.00
68	Hexachlorobenzene	0.244	0.247	-1.2	116	0.00
69	Atrazine	0.190	0.191	-0.5	116	0.00
70 C	Pentachlorophenol	0.136	0.139	-2.2	114	0.00
71	Phenanthrene	0.994	0.999	-0.5	114	0.00
72	Anthracene	1.002	1.022	-2.0	115	0.00
73	Carbazole	0.992	0.991	0.1	114	0.00
74	Di-n-butylphthalate	1.115	1.178	-5.7	118	0.00
75 C	Fluoranthene	1.137	1.118	1.7	105	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
77	Benzidine	0.676	0.748	-10.7	103	0.00
78	Pyrene	1.417	1.522	-7.4	103	0.00
79 S	Terphenyl-d14	0.905	0.929	-2.7	103	0.00
80	Butylbenzylphthalate	0.629	0.639	-1.6	98	0.00
81	Benzo(a)anthracene	1.202	1.201	0.1	100	0.00
82	3,3'-Dichlorobenzidine	0.459	0.462	-0.7	99	0.00
83	Chrysene	1.233	1.231	0.2	98	0.00
84	Bis(2-ethylhexyl)phthalate	0.768	0.819	-6.6	106	0.00
85 c	Di-n-octyl phthalate	1.465	1.501	-2.5	98	0.00
86 I	Perylene-d12	1.000	1.000	0.0	94	0.00
87	Indeno(1,2,3-cd)pyrene	1.217	1.185	2.6	87	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061120\
Data File : BF120596.D
Acq On : 11 Jun 2020 10:01
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 11 12:36:05 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 11 11:28:22 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.191	1.177	1.2	96	0.00
89	Benzo(k)fluoranthene	1.054	1.109	-5.2	89	0.00
90 C	Benzo(a)pyrene	1.066	1.076	-0.9	88	0.00
91	Dibenzo(a,h)anthracene	0.976	0.970	0.6	88	0.00
92	Benzo(q,h,i)perylene	0.983	0.947	3.7	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0