


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061120\
 Data File : BF120595.D
 Acq On : 11 Jun 2020 8:32
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF061020.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jun 11 11:28:22 2020



AutoFind: Scans 1612, 1613, 1614; Background Corrected with Scan 1608

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.3	40040	PASS
68	69	0.00	2	1.6	614	PASS
69	198	0.00	100	34.0	37482	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	47.5	52386	PASS
197	198	0.00	2	0.6	612	PASS
198	198	100	100	100.0	110314	PASS
199	198	5	9	6.9	7565	PASS
275	198	10	60	27.0	29816	PASS
365	198	1	100	3.0	3354	PASS
441	198	0.01	100	13.8	15180	PASS
442	198	50	100	88.4	97493	PASS
443	442	15	24	18.3	17888	PASS