

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061121\
 Data File : BF124276.D
 Acq On : 11 Jun 2021 23:23
 Operator : JU/CG
 Sample : M2625-14
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 18-B12(190-194)

Quant Time: Jun 14 11:25:53 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 10 01:53:56 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.840	152	28672	20.000	ng	-0.01
21) Naphthalene-d8	8.122	136	109585	20.000	ng	-0.01
39) Acenaphthene-d10	9.875	164	68053	20.000	ng	-0.02
64) Phenanthrene-d10	11.363	188	133636	20.000	ng	-0.01
76) Chrysene-d12	14.010	240	80531	20.000	ng	-0.01
86) Perylene-d12	15.486	264	84599	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.469	112	188805	99.127	ng	-0.01
7) Phenol-d6	6.481	99	238072	92.990	ng	-0.02
23) Nitrobenzene-d5	7.404	82	181303	65.618	ng	-0.01
42) 2,4,6-Tribromophenol	10.669	330	89897	93.240	ng	-0.01
45) 2-Fluorobiphenyl	9.198	172	310666	57.445	ng	-0.01
79) Terphenyl-d14	12.951	244	324577	55.336	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061121\
Data File : BF124276.D
Acq On : 11 Jun 2021 23:23
Operator : JU/CG
Sample : M2625-14
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
18-B12(190-194)

Quant Time: Jun 14 11:25:53 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 10 01:53:56 2021
Response via : Initial Calibration

