


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF061217\
 Data File : BF095872.D
 Acq On : 12 Jun 2017 15:05
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF060517.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jun 05 16:15:31 2017



AutoFind: Scans 2310, 2311, 2312; Background Corrected with Scan 2303

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.5	11910	PASS
68	69	0.00	2	1.6	180	PASS
69	198	0.00	100	33.4	11211	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.0	16424	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	33520	PASS
199	198	5	9	6.5	2189	PASS
275	198	10	60	23.0	7693	PASS
365	198	1	100	2.5	822	PASS
441	442	0.01	24	15.6	4979	PASS
442	198	50	100	95.5	32005	PASS
443	442	15	24	17.8	5699	PASS