

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF061220\
 Data File : BF120616.D
 Acq On : 12 Jun 2020 17:50
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 12 18:26:23 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 11 11:28:22 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.553	0.512	7.4	88	-0.04
3	Pyridine	1.554	1.438	7.5	87	-0.04
4	n-Nitrosodimethylamine	0.645	0.626	2.9	91	-0.04
5 S	2-Fluorophenol	1.167	1.154	1.1	97	-0.01
6	Aniline	1.859	1.950	-4.9	95	0.00
7 S	Phenol-d6	1.453	1.508	-3.8	95	0.00
8	2-Chlorophenol	1.318	1.298	1.5	93	0.00
9	Benzaldehyde	0.919	0.920	-0.1	93	0.00
10 C	Phenol	1.550	1.656	-6.8	97	0.00
11	bis(2-Chloroethyl)ether	1.291	1.299	-0.6	95	0.00
12	1,3-Dichlorobenzene	1.401	1.423	-1.6	97	0.00
13 C	1,4-Dichlorobenzene	1.383	1.435	-3.8	103	0.00
14	1,2-Dichlorobenzene	1.347	1.354	-0.5	97	0.00
15	Benzyl Alcohol	1.030	1.073	-4.2	98	0.00
16	2,2'-oxybis(1-Chloropropane	1.859	1.983	-6.7	104	0.00
17	2-Methylphenol	1.124	1.139	-1.3	105	0.00
18	Hexachloroethane	0.510	0.539	-5.7	104	0.00
19 P	n-Nitroso-di-n-propylamine	0.843	0.868	-3.0	102	0.00
20	3+4-Methylphenols	1.405	1.374	2.2	100	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
22	Acetophenone	0.447	0.445	0.4	101	0.00
23 S	Nitrobenzene-d5	0.358	0.360	-0.6	100	0.00
24	Nitrobenzene	0.374	0.383	-2.4	108	0.00
25	Isophorone	0.697	0.672	3.6	96	0.00
26 C	2-Nitrophenol	0.198	0.188	5.1	91	0.00
27	2,4-Dimethylphenol	0.292	0.279	4.5	93	0.00
28	bis(2-Chloroethoxy)methane	0.404	0.414	-2.5	94	0.00
29 C	2,4-Dichlorophenol	0.296	0.290	2.0	89	0.00
30	1,2,4-Trichlorobenzene	0.329	0.329	0.0	92	0.00
31	Naphthalene	0.982	1.011	-3.0	97	0.00
32	Benzoic acid	0.224	0.216	3.6	87	-0.01
33	4-Chloroaniline	0.451	0.448	0.7	97	0.00
34 C	Hexachlorobutadiene	0.192	0.203	-5.7	106	0.00
35	Caprolactam	0.095	0.092	3.2	102	-0.01
36 C	4-Chloro-3-methylphenol	0.296	0.298	-0.7	103	0.00
37	2-Methylnaphthalene	0.641	0.673	-5.0	105	0.00
38	1-Methylnaphthalene	0.603	0.643	-6.6	106	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
40	1,2,4,5-Tetrachlorobenzene	0.597	0.627	-5.0	105	0.00
41 P	Hexachlorocyclopentadiene	0.335	0.345	-3.0	104	0.00
42 S	2,4,6-Tribromophenol	0.233	0.239	-2.6	101	0.00
43 C	2,4,6-Trichlorophenol	0.424	0.421	0.7	100	0.00
44	2,4,5-Trichlorophenol	0.481	0.482	-0.2	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF061220\
 Data File : BF120616.D
 Acq On : 12 Jun 2020 17:50
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 12 18:26:23 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 11 11:28:22 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.307	1.269	2.9	104	0.00
46	1,1'-Biphenyl	1.493	1.562	-4.6	104	0.00
47	2-Chloronaphthalene	1.221	1.265	-3.6	103	0.00
48	2-Nitroaniline	0.384	0.407	-6.0	107	0.00
49	Acenaphthylene	1.885	1.949	-3.4	102	0.00
50	Dimethylphthalate	1.440	1.449	-0.6	102	0.00
51	2,6-Dinitrotoluene	0.326	0.322	1.2	98	0.00
52 C	Acenaphthene	1.124	1.174	-4.4	103	0.00
53	3-Nitroaniline	0.392	0.398	-1.5	103	0.00
54 P	2,4-Dinitrophenol	0.175	0.158	9.7	93	0.00
55	Dibenzofuran	1.724	1.785	-3.5	103	0.00
56 P	4-Nitrophenol	0.295	0.279	5.4	96	0.00
57	2,4-Dinitrotoluene	0.433	0.429	0.9	99	0.00
58	Fluorene	1.329	1.373	-3.3	106	0.00
59	2,3,4,6-Tetrachlorophenol	0.378	0.376	0.5	100	0.00
60	Diethylphthalate	1.419	1.431	-0.8	100	0.00
61	4-Chlorophenyl-phenylether	0.686	0.710	-3.5	105	0.00
62	4-Nitroaniline	0.389	0.394	-1.3	101	0.00
63	Azobenzene	1.288	1.395	-8.3	107	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
65	4,6-Dinitro-2-methylphenol	0.120	0.116	3.3	94	0.00
66 c	n-Nitrosodiphenylamine	0.614	0.635	-3.4	102	0.00
67	4-Bromophenyl-phenylether	0.227	0.232	-2.2	98	0.00
68	Hexachlorobenzene	0.244	0.252	-3.3	101	0.00
69	Atrazine	0.190	0.184	3.2	95	0.00
70 C	Pentachlorophenol	0.136	0.137	-0.7	96	0.00
71	Phenanthrene	0.994	1.024	-3.0	100	0.00
72	Anthracene	1.002	1.036	-3.4	100	0.00
73	Carbazole	0.992	1.004	-1.2	98	0.00
74	Di-n-butylphthalate	1.115	1.175	-5.4	101	0.00
75 C	Fluoranthene	1.137	1.183	-4.0	95	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
77	Benzidine	0.676	0.571	15.5	81	0.00
78	Pyrene	1.417	1.400	1.2	98	0.00
79 S	Terphenyl-d14	0.905	0.866	4.3	98	0.00
80	Butylbenzylphthalate	0.629	0.613	2.5	97	0.00
81	Benzo(a)anthracene	1.202	1.214	-1.0	104	0.00
82	3,3'-Dichlorobenzidine	0.459	0.465	-1.3	103	0.00
83	Chrysene	1.233	1.245	-1.0	102	0.00
84	Bis(2-ethylhexyl)phthalate	0.768	0.815	-6.1	109	0.00
85 c	Di-n-octyl phthalate	1.465	1.479	-1.0	99	0.00
86 I	Perylene-d12	1.000	1.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	1.217	1.340	-10.1	104	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF061220\
Data File : BF120616.D
Acq On : 12 Jun 2020 17:50
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 12 18:26:23 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 11 11:28:22 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.191	1.193	-0.2	103	0.00
89	Benzo(k)fluoranthene	1.054	1.122	-6.5	95	0.00
90 C	Benzo(a)pyrene	1.066	1.107	-3.8	95	0.00
91	Dibenzo(a,h)anthracene	0.976	1.090	-11.7	105	-0.01
92	Benzo(q,h,i)perylene	0.983	1.115	-13.4	107	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0