

Data Path : Z:\HPCHEM1\BNA F\DATA\BF061416\
 Data File : BF088013.D
 Acq On : 14 Jun 2016 22:25
 Operator : UM/SJ
 Sample : H3489-03 5X
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 S7-B5(0-6)C

Manual Integrations
 APPROVED

sohil
 6/15/2016 6:45:22 PM

Quant Time: Jun 15 04:58:34 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 15 04:08:24 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.78	152	102062	20.00	ng	-0.01
21) Naphthalene-d8	8.07	136	404576	20.00	ng	0.00
38) Acenaphthene-d10	9.83	164	177436	20.00	ng	0.00
63) Phenanthrene-d10	11.30	188	266876	20.00	ng	0.00
75) Chrysene-d12	13.94	240	258985	20.00	ng	0.00
86) Perylene-d12	15.35	264	216434	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.36	112	142036	23.79	ng	-0.01
7) Phenol-d6	6.41	99	224649	31.05	ng	-0.01
23) Nitrobenzene-d5	7.35	82	140582	20.29	ng	-0.01
41) 2,4,6-Tribromophenol	10.61	330	8607	4.59	ng	-0.01
44) 2-Fluorobiphenyl	9.14	172	273710	20.00	ng	-0.01
78) Terphenyl-d14	12.89	244	208587	18.33	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	9.54	163	56165	4.37	ng	99
74) Fluoranthene	12.51	202	60738	4.11	ng	98
77) Pyrene	12.74	202	56427	2.94	ng	98
80) Benzo(a)anthracene	13.93	228	35222m	2.39	ng	
82) Chrysene	13.97	228	29099m	2.10	ng	
87) Benzo(b)fluoranthene	14.95	252	39410m	2.61	ng	
89) Benzo(a)pyrene	15.29	252	25597	2.10	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF061416\
Data File : BF088013.D
Acq On : 14 Jun 2016 22:25
Operator : UM/SJ
Sample : H3489-03 5X
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
S7-B5(0-6)C

Manual Integrations
APPROVED

sohil
6/15/2016 6:45:22 PM

Quant Time: Jun 15 04:58:34 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060716.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 15 04:08:24 2016
Response via : Initial Calibration

