

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061421\
 Data File : BF124318.D
 Acq On : 14 Jun 2021 21:39
 Operator : JU/CG
 Sample : M2674-06DL 10X
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TMR-DS-06-060921DL

Quant Time: Jun 15 02:28:01 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 14 12:10:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.828	152	31061	20.00	ng	0.00
21) Naphthalene-d8	8.116	136	116882	20.00	ng	# 0.00
39) Acenaphthene-d10	9.875	164	59827	20.00	ng	0.00
64) Phenanthrene-d10	11.369	188	88546	20.00	ng	0.00
76) Chrysene-d12	14.015	240	75477	20.00	ng	0.00
86) Perylene-d12	15.504	264	52754	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.481	112	17567	8.51	ng	0.02
7) Phenol-d6	6.498	99	22799	8.22	ng	0.02
23) Nitrobenzene-d5	7.392	82	25505	8.65	ng	-0.01
42) 2,4,6-Tribromophenol	10.669	330	7020	8.28	ng	0.00
45) 2-Fluorobiphenyl	9.192	172	29990	6.31	ng	0.00
79) Terphenyl-d14	12.951	244	28587	5.20	ng	0.00
Target Compounds						Qvalue
10) Phenol	6.516	94	56344	18.80	ng	89
17) 2-Methylphenol	7.275	107	134396	72.01	ng	83
20) 3+4-Methylphenols	7.275	107	134396	54.34	ng	# 70
24) Nitrobenzene	7.375	77	7413	2.51	ng	# 1
27) 2,4-Dimethylphenol	7.804	122	96184	51.65	ng	88
30) 1,2,4-Trichlorobenzene	8.157	180	9986	4.14	ng	# 22
31) Naphthalene	8.139	128	73322	10.49	ng	96
32) Benzoic acid	7.928	122	75702	66.25	ng	# 26
34) Hexachlorobutadiene	8.469	225	15480	9.08	ng	# 23
37) 2-Methylnaphthalene	8.833	142	86020	17.55	ng	97
38) 1-Methylnaphthalene	8.933	142	52252	11.41	ng	93
46) 1,1'-Biphenyl	9.292	154	16227	3.21	ng	91
48) 2-Nitroaniline	9.469	65	9848	6.64	ng	# 34
53) 3-Nitroaniline	9.827	138	7259	6.83	ng	# 10
54) 2,4-Dinitrophenol	9.957	184	3574	9.40	ng	# 1
55) Dibenzofuran	10.080	168	28071	4.60	ng	# 92
56) 4-Nitrophenol	9.851	139	4837	6.37	ng	95
58) Fluorene	10.427	166	36549	7.80	ng	# 78
70) Pentachlorophenol	11.233	266	121	4.82	ng	# 18
71) Phenanthrene	11.392	178	48556	8.78	ng	# 94
72) Anthracene	11.392	178	48556	8.72	ng	94
75) Fluoranthene	12.586	202	16630	2.86	ng	81
77) Benzidine	12.704	184	822	4.49	ng	# 1
78) Pyrene	12.816	202	22283	3.07	ng	# 82
87) Indeno(1,2,3-cd)pyrene	16.986	276	3174	4.99	ng	# 76
91) Dibenzo(a,h)anthracene	17.009	278	252	4.24	ng	# 1
92) Benzo(g,h,i)perylene	17.445	276	1727	4.24	ng	# 60


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061421\
Data File : BF124318.D
Acq On : 14 Jun 2021 21:39
Operator : JU/CG
Sample : M2674-06DL 10X
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TMR-DS-06-060921DL

Quant Time: Jun 15 02:28:01 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 14 12:10:02 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.828 min Scan# 806
Delta R.T. -0.006 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39




Instrument :
BNA_F
ClientSampleId :
TMR-DS-06-060921DL


Tgt Ion:152 Resp: 31061
Ion Ratio Lower Upper
152 100
150 172.6 127.0 190.6
115 73.1 54.3 81.5



#5
2-Fluorophenol
Concen: 8.51 ng
RT: 5.481 min Scan# 577
Delta R.T. 0.023 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39

Tgt Ion:112 Resp: 17567
Ion Ratio Lower Upper
112 100
64 69.1 61.4 92.0
63 38.3 33.0 49.4









#7
Phenol-d6
Concen: 8.22 ng
RT: 6.498 min Scan# 750
Delta R.T. 0.018 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39




Instrument :
BNA_F
ClientSampleId :
TMR-DS-06-060921DL


Tgt Ion:	99	Resp:	22799
Ion Ratio	Lower	Upper	
99	100		
42	21.5	19.8	29.8
71	43.7	31.1	46.7



#10
Phenol
Concen: 18.80 ng
RT: 6.516 min Scan# 753
Delta R.T. 0.023 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39

Tgt Ion:	94	Resp:	56344
Ion Ratio	Lower	Upper	
94	100		
65	38.2	18.1	58.1
66	62.7	29.0	69.0









#17
2-Methylphenol
Concen: 72.01 ng
RT: 7.275 min Scan# 882
Delta R.T. 0.176 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39




Instrument :
BNA_F
ClientSampleId :
TMR-DS-06-060921DL


Tgt Ion:	107	Resp:	134396
Ion	Ratio	Lower	Upper
107	100		
108	90.2	89.8	134.8
77	41.1	40.9	61.3
79	42.0	40.2	60.4



#20
3+4-Methylphenols
Concen: 54.34 ng
RT: 7.275 min Scan# 882
Delta R.T. 0.023 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39


Tgt Ion:	107	Resp:	134396
Ion	Ratio	Lower	Upper
107	100		
108	90.2	71.5	111.5
77	41.1	82.0	122.0#
79	42.0	14.1	54.1








#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.116 min Scan# 1025
Delta R.T. -0.000 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39

Instrument :
BNA_F
ClientSampleId :
TMR-DS-06-060921DL







Tgt Ion:136 Resp: 116882
Ion Ratio Lower Upper
136 100
137 10.6 8.7 13.1
54 9.4 9.8 14.6#
68 5.3 4.1 6.1



#23
Nitrobenzene-d5
Concen: 8.65 ng
RT: 7.392 min Scan# 902
Delta R.T. -0.012 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39

Tgt Ion: 82 Resp: 25505
Ion Ratio Lower Upper
82 100
128 33.5 29.4 44.2
54 57.6 54.1 81.1









#24
Nitrobenzene
Concen: 2.51 ng
RT: 7.375 min Scan# 899
Delta R.T. -0.047 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39




Instrument :
BNA_F
ClientSampleId :
TMR-DS-06-060921DL


Tgt Ion: 77 Resp: 7413
Ion Ratio Lower Upper
77 100
123 0.0 34.5 51.7#
65 196.9 12.2 18.2#



#27
2,4-Dimethylphenol
Concen: 51.65 ng
RT: 7.804 min Scan# 972
Delta R.T. 0.023 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39

Tgt Ion: 122 Resp: 96184
Ion Ratio Lower Upper
122 100
107 120.2 106.3 159.5
121 47.0 45.4 68.0









#30
1,2,4-Trichlorobenzene
Concen: 4.14 ng
RT: 8.157 min Scan# 1032
Delta R.T. 0.100 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39




Instrument :
BNA_F
ClientSampleId :
TMR-DS-06-060921DL


Tgt Ion:180 Resp: 9986
Ion Ratio Lower Upper
180 100
182 0.0 77.2 115.8#
145 22.1 26.0 39.0#



#31
Naphthalene
Concen: 10.49 ng
RT: 8.139 min Scan# 1029
Delta R.T. -0.000 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39

Tgt Ion:128 Resp: 73322
Ion Ratio Lower Upper
128 100
129 10.5 8.9 13.3
127 15.6 10.8 16.2









#32
Benzoic acid
Concen: 66.25 ng
RT: 7.928 min Scan# 993
Delta R.T. 0.012 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39




Instrument :
BNA_F
ClientSampleId :
TMR-DS-06-060921DL


Tgt Ion:122 Resp: 75702
Ion Ratio Lower Upper
122 100
105 18.6 107.8 147.8#
77 53.4 81.6 121.6#



#34
Hexachlorobutadiene
Concen: 9.08 ng
RT: 8.469 min Scan# 1085
Delta R.T. 0.212 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39

Tgt Ion:225 Resp: 15480
Ion Ratio Lower Upper
225 100
223 0.0 50.4 75.6#
227 7.2 51.4 77.0#









#37
2-Methylnaphthalene
Concen: 17.55 ng
RT: 8.833 min Scan# 1147
Delta R.T. -0.000 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39




Instrument :
BNA_F
ClientSampleId :
TMR-DS-06-060921DL


Tgt Ion:142 Resp: 86020
Ion Ratio Lower Upper
142 100
141 92.4 71.8 107.6



#38
1-Methylnaphthalene
Concen: 11.41 ng
RT: 8.933 min Scan# 1164
Delta R.T. -0.000 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39

Tgt Ion:142 Resp: 52252
Ion Ratio Lower Upper
142 100
141 83.9 72.8 109.2









#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.875 min Scan# 1324
Delta R.T. -0.000 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39




Instrument :
BNA_F
ClientSampleId :
TMR-DS-06-060921DL


Tgt Ion:164 Resp: 59827
Ion Ratio Lower Upper
164 100
162 113.7 80.7 121.1
160 52.0 37.3 55.9



#42
2,4,6-Tribromophenol
Concen: 8.28 ng
RT: 10.669 min Scan# 1459
Delta R.T. -0.000 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39

Tgt Ion:330 Resp: 7020
Ion Ratio Lower Upper
330 100
332 82.8 77.1 115.7
141 78.1 27.9 41.9#





#45

2-Fluorobiphenyl

Concen: 6.31 ng

RT: 9.192 min Scan# 1208

Delta R.T. -0.006 min

Lab File: BF124318.D

Acq: 14 Jun 2021 21:39





Instrument :

BNA_F

ClientSampleId :

TMR-DS-06-060921DL

Tgt Ion:172	Resp:	29990
Ion Ratio	Lower	Upper
172	100	
171	36.1	28.1 42.1
170	25.7	19.2 28.8



#46

1,1'-Biphenyl

Concen: 3.21 ng




RT: 9.292 min Scan# 1225


Delta R.T. -0.006 min

Lab File: BF124318.D

Acq: 14 Jun 2021 21:39

Tgt Ion:154	Resp:	16227
Ion Ratio	Lower	Upper
154	100	
153	34.7	20.7 60.7
76	14.4	0.0 31.7









#48
2-Nitroaniline
Concen: 6.64 ng
RT: 9.469 min Scan# 1255
Delta R.T. 0.047 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39




Instrument :
BNA_F
ClientSampleId :
TMR-DS-06-060921DL


Tgt Ion: 65 Resp: 9848
Ion Ratio Lower Upper
65 100
92 22.1 47.1 70.7#
138 16.5 70.2 105.2#



#53
3-Nitroaniline
Concen: 6.83 ng
RT: 9.827 min Scan# 1316
Delta R.T. -0.012 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39

Tgt Ion: 138 Resp: 7259
Ion Ratio Lower Upper
138 100
108 36.7 10.2 15.4#
92 20.9 103.3 154.9#









#54
2,4-Dinitrophenol
Concen: 9.40 ng
RT: 9.957 min Scan# 1338
Delta R.T. 0.006 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39




Instrument :
BNA_F
ClientSampleId :
TMR-DS-06-060921DL


Tgt Ion:184 Resp: 3574
Ion Ratio Lower Upper
184 100
63 489.4 73.5 110.3#
154 124.7 52.5 78.7#



#55
Dibenzofuran
Concen: 4.60 ng
RT: 10.080 min Scan# 1359
Delta R.T. -0.000 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39

Tgt Ion:168 Resp: 28071
Ion Ratio Lower Upper
168 100
139 45.3 33.2 49.8
169 19.6 10.8 16.2#









#56
4-Nitrophenol
Concen: 6.37 ng
RT: 9.851 min Scan# 1320
Delta R.T. -0.165 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39




Instrument :
BNA_F
ClientSampleId :
TMR-DS-06-060921DL


Tgt Ion:139 Resp: 4837
Ion Ratio Lower Upper
139 100
109 80.7 63.2 103.2
65 138.2 110.0 150.0



#58
Fluorene
Concen: 7.80 ng
RT: 10.427 min Scan# 1418
Delta R.T. -0.000 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39

Tgt Ion:166 Resp: 36549
Ion Ratio Lower Upper
166 100
165 111.7 73.5 110.3#
167 25.4 10.7 16.1#









#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.369 min Scan# 1578
Delta R.T. -0.000 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39




Instrument :
BNA_F
ClientSampleId :
TMR-DS-06-060921DL


Tgt Ion:188 Resp: 88546
Ion Ratio Lower Upper
188 100
94 9.0 6.9 10.3
80 10.0 8.2 12.2



#70
Pentachlorophenol
Concen: 4.82 ng
RT: 11.233 min Scan# 1555
Delta R.T. 0.053 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39

Tgt Ion:266 Resp: 121
Ion Ratio Lower Upper
266 100
268 0.0 51.6 77.4#
264 0.0 52.6 78.8#









#71
Phenanthrene
Concen: 8.78 ng
RT: 11.392 min Scan# 1582
Delta R.T. -0.000 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39




Instrument :
BNA_F
ClientSampleId :
TMR-DS-06-060921DL


Tgt Ion:178 Resp: 48556
Ion Ratio Lower Upper
178 100
176 17.1 15.6 23.4
179 18.8 12.5 18.7#



#72
Anthracene
Concen: 8.72 ng
RT: 11.392 min Scan# 1582
Delta R.T. -0.053 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39

Tgt Ion:178 Resp: 48556
Ion Ratio Lower Upper
178 100
176 17.1 15.4 23.0
179 18.8 12.6 18.8









#75
Fluoranthene
Concen: 2.86 ng
RT: 12.586 min Scan# 1785
Delta R.T. -0.000 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39




Instrument :
BNA_F
ClientSampleId :
TMR-DS-06-060921DL


Tgt Ion:	202	Resp:	16630
Ion Ratio	Lower	Upper	
202	100		
101	16.3	0.0	30.9
203	27.3	0.0	37.5



#76
Chrysene-d12
Concen: 20.00 ng
RT: 14.015 min Scan# 2028
Delta R.T. -0.000 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39

Tgt Ion:	240	Resp:	75477
Ion Ratio	Lower	Upper	
240	100		
120	10.7	7.2	10.8
236	26.5	19.0	28.4









#77
Benzidine
Concen: 4.49 ng
RT: 12.704 min Scan# 1805
Delta R.T. -0.000 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39




Instrument :
BNA_F
ClientSampleId :
TMR-DS-06-060921DL


Tgt Ion:184 Resp: 822
Ion Ratio Lower Upper
184 100
185 82.8 12.1 18.1#
183 81.4 7.8 11.8#



#78
Pyrene
Concen: 3.07 ng
RT: 12.816 min Scan# 1824
Delta R.T. -0.000 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39

Tgt Ion:202 Resp: 22283
Ion Ratio Lower Upper
202 100
200 29.2 16.1 24.1#
203 26.1 14.8 22.2#









#79
Terphenyl-d14
Concen: 5.20 ng
RT: 12.951 min Scan# 1847
Delta R.T. -0.006 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39




Instrument :
BNA_F
ClientSampleId :
TMR-DS-06-060921DL


Tgt Ion:244 Resp: 28587
Ion Ratio Lower Upper
244 100
212 10.8 5.7 8.5#
122 8.2 9.7 14.5#



#86
Perylene-d12
Concen: 20.00 ng
RT: 15.504 min Scan# 2281
Delta R.T. 0.012 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39

Tgt Ion:264 Resp: 52754
Ion Ratio Lower Upper
264 100
260 23.7 18.7 28.1
265 21.4 17.8 26.8









#87
Indeno(1,2,3-cd)pyrene
Concen: 4.99 ng
RT: 16.986 min Scan# 2533
Delta R.T. -0.000 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39




Instrument :
BNA_F
ClientSampleId :
TMR-DS-06-060921DL


Tgt Ion:276 Resp: 3174
Ion Ratio Lower Upper
276 100
138 10.1 16.2 24.2#
277 38.1 20.1 30.1#



#91
Dibenzo(a,h)anthracene
Concen: 4.24 ng
RT: 17.009 min Scan# 2537
Delta R.T. 0.012 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39

Tgt Ion:278 Resp: 252
Ion Ratio Lower Upper
278 100
139 96.4 10.2 15.2#
279 108.0 19.4 29.0#





#92
Benzo(g,h,i)perylene
Concen: 4.24 ng
RT: 17.445 min Scan# 2611
Delta R.T. 0.006 min
Lab File: BF124318.D
Acq: 14 Jun 2021 21:39

Instrument :
BNA_F
ClientSampleId :
TMR-DS-06-060921DL

Tgt Ion:276 Resp: 1727
Ion Ratio Lower Upper
276 100
277 42.2 18.2 27.2#
138 0.0 13.8 20.6#

