

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061422\
 Data File : BF128798.D
 Acq On : 14 Jun 2022 13:34
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 15 01:00:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 14 14:53:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	79	0.00
2	1,4-Dioxane	40.000	35.300	11.8	66	0.00
3	Pyridine	40.000	37.730	5.7	71	0.00
4	n-Nitrosodimethylamine	40.000	33.246	16.9	62	-0.01
5 S	2-Fluorophenol	80.000	77.622	3.0	75	0.00
6	Aniline	40.000	40.591	-1.5	80	0.00
7 S	Phenol-d6	80.000	79.721	0.3	78	0.00
8	2-Chlorophenol	40.000	39.747	0.6	78	0.00
9	Benzaldehyde	40.000	40.273	-0.7	97	0.00
10 C	Phenol	40.000	39.778	0.6	77	0.00
11	bis(2-Chloroethyl)ether	40.000	39.043	2.4	76	0.00
12	1,3-Dichlorobenzene	40.000	39.397	1.5	77	0.00
13 C	1,4-Dichlorobenzene	40.000	39.335	1.7	78	0.00
14	1,2-Dichlorobenzene	40.000	39.966	0.1	80	0.00
15	Benzyl Alcohol	40.000	37.258	6.9	74	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.558	11.1	72	0.00
17	2-Methylphenol	40.000	39.953	0.1	80	0.00
18	Hexachloroethane	40.000	40.189	-0.5	78	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.540	1.2	77	0.00
20	3+4-Methylphenols	40.000	40.747	-1.9	80	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	77	0.00
22	Acetophenone	40.000	42.269	-5.7	78	0.00
23 S	Nitrobenzene-d5	80.000	83.723	-4.7	78	0.00
24	Nitrobenzene	40.000	41.850	-4.6	78	0.00
25	Isophorone	40.000	40.888	-2.2	76	0.00
26 C	2-Nitrophenol	40.000	45.005	-12.5	81	0.00
27	2,4-Dimethylphenol	40.000	42.286	-5.7	76	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.246	-0.6	74	0.00
29 C	2,4-Dichlorophenol	40.000	41.804	-4.5	76	0.00
30	1,2,4-Trichlorobenzene	40.000	40.970	-2.4	75	0.00
31	Naphthalene	40.000	40.632	-1.6	78	0.00
32	Benzoic acid	40.000	24.727	38.2#	44	0.00
33	4-Chloroaniline	40.000	41.383	-3.5	80	0.00
34 C	Hexachlorobutadiene	40.000	40.953	-2.4	79	0.00
35	Caprolactam	40.000	41.554	-3.9	80	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.432	-1.1	78	0.00
37	2-Methylnaphthalene	40.000	40.885	-2.2	79	0.00
38	1-Methylnaphthalene	40.000	40.645	-1.6	80	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	79	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.554	-3.9	79	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.055	7.4	68	0.00
42 S	2,4,6-Tribromophenol	80.000	80.099	-0.1	74	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.283	6.8	72	0.00
44	2,4,5-Trichlorophenol	40.000	37.655	5.9	72	0.00
45 S	2-Fluorobiphenyl	80.000	81.389	-1.7	79	0.00
46	1,1'-Biphenyl	40.000	40.777	-1.9	79	0.00
47	2-Chloronaphthalene	40.000	40.866	-2.2	80	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061422\
 Data File : BF128798.D
 Acq On : 14 Jun 2022 13:34
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 15 01:00:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 14 14:53:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.444	-3.6	79	0.00
49	Acenaphthylene	40.000	40.666	-1.7	79	0.00
50	Dimethylphthalate	40.000	40.793	-2.0	79	0.00
51	2,6-Dinitrotoluene	40.000	42.124	-5.3	79	0.00
52 C	Acenaphthene	40.000	37.022	7.4	70	0.00
53	3-Nitroaniline	40.000	43.788	-9.5	82	0.00
54 P	2,4-Dinitrophenol	40.000	40.749	-1.9	75	0.00
55	Dibenzofuran	40.000	40.728	-1.8	79	0.00
56 P	4-Nitrophenol	40.000	34.166	14.6	67	0.00
57	2,4-Dinitrotoluene	40.000	43.782	-9.5	81	0.00
58	Fluorene	40.000	41.044	-2.6	78	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	34.780	13.0	65	0.00
60	Diethylphthalate	40.000	40.563	-1.4	78	0.00
61	4-Chlorophenyl-phenylether	40.000	41.193	-3.0	78	0.00
62	4-Nitroaniline	40.000	43.906	-9.8	83	0.00
63	Azobenzene	40.000	40.060	-0.2	77	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	80	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.303	-5.8	81	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.338	-0.8	78	0.00
67	4-Bromophenyl-phenylether	40.000	41.256	-3.1	80	0.00
68	Hexachlorobenzene	40.000	41.369	-3.4	80	0.00
69	Atrazine	40.000	39.843	0.4	76	0.00
70 C	Pentachlorophenol	40.000	43.540	-8.8	83	0.00
71	Phenanthrene	40.000	40.116	-0.3	79	0.00
72	Anthracene	40.000	40.923	-2.3	81	0.00
73	Carbazole	40.000	42.177	-5.4	83	0.00
74	Di-n-butylphthalate	40.000	40.397	-1.0	78	0.00
75 C	Fluoranthene	40.000	43.145	-7.9	83	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	87	0.00
77	Benzydine	40.000	37.958	5.1	89	0.00
78	Pyrene	40.000	36.542	8.6	82	0.00
79 S	Terphenyl-d14	80.000	74.633	6.7	80	0.00
80	Butylbenzylphthalate	40.000	37.344	6.6	79	0.00
81	Benzo(a)anthracene	40.000	40.142	-0.4	86	0.00
82	3,3'-Dichlorobenzidine	40.000	42.194	-5.5	92	0.00
83	Chrysene	40.000	40.177	-0.4	87	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.495	3.8	82	0.00
85 c	Di-n-octyl phthalate	40.000	40.301	-0.8	85	0.00
86 I	Perylene-d12	20.000	20.000	0.0	91	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.020	7.4	93	0.00
88	Benzo(b)fluoranthene	40.000	38.829	2.9	86	0.00
89	Benzo(k)fluoranthene	40.000	41.763	-4.4	92	0.00
90 C	Benzo(a)pyrene	40.000	40.559	-1.4	91	0.00
91	Dibenzo(a,h)anthracene	40.000	37.634	5.9	92	0.00
92	Benzo(g,h,i)perylene	40.000	36.287	9.3	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061422\
Data File : BF128798.D
Acq On : 14 Jun 2022 13:34
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 15 01:00:49 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 14 14:53:53 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0