

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF061515\  
 Data File : BF079982.D  
 Acq On : 15 Jun 2015 16:18  
 Operator : TP/IZ  
 Sample : G2639-02  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 SP15-LF2MW1

Quant Time: Jun 16 01:14:03 2015  
 Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF061115.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Jun 15 15:44:31 2015  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 7.35  | 152  | 40133    | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.64  | 136  | 159747   | 20.00  | ng    | 0.00     |
| 38) Acenaphthene-d10        | 10.41 | 164  | 82219    | 20.00  | ng    | 0.00     |
| 63) Phenanthrene-d10        | 11.93 | 188  | 163560   | 20.00  | ng    | 0.01     |
| 75) Chrysene-d12            | 15.03 | 240  | 177829   | 20.00  | ng    | 0.13     |
| 86) Perylene-d12            | 16.99 | 264  | 176556   | 20.00  | ng    | 0.18     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.96  | 112  | 137408   | 62.78  | ng    | 0.00     |
| 7) Phenol-d6                | 6.95  | 99   | 115399   | 38.70  | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.91  | 82   | 270537   | 111.74 | ng    | 0.00     |
| 41) 2,4,6-Tribromophenol    | 11.21 | 330  | 123016   | 148.13 | ng    | 0.00     |
| 44) 2-Fluorobiphenyl        | 9.72  | 172  | 562097   | 95.73  | ng    | 0.00     |
| 78) Terphenyl-d14           | 13.69 | 244  | 739947   | 95.93  | ng    | 0.07     |

| Target Compounds | Qvalue |
|------------------|--------|
|------------------|--------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF061515\  
Data File : BF079982.D  
Acq On : 15 Jun 2015 16:18  
Operator : TP/IZ  
Sample : G2639-02  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
SP15-LF2MW1

Quant Time: Jun 16 01:14:03 2015  
Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF061115.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Jun 15 15:44:31 2015  
Response via : Initial Calibration

