


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF061516\
 Data File : BF088039.D
 Acq On : 15 Jun 2016 16:13
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF060716.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jun 13 14:10:49 2016



AutoFind: Scans 871, 872, 873; Background Corrected with Scan 868

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.3	10377	PASS
68	69	0.00	2	1.4	149	PASS
69	198	0.00	100	34.3	10712	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	51.5	16066	PASS
197	198	0.00	2	0.6	180	PASS
198	198	100	100	100.0	31198	PASS
199	198	5	9	6.5	2028	PASS
275	198	10	60	23.9	7465	PASS
365	198	1	100	2.5	790	PASS
441	442	0.01	24	15.5	3634	PASS
442	198	50	100	75.3	23500	PASS
443	442	15	24	18.8	4415	PASS