


Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF061518\  
 Data File : BF106486.D  
 Acq On : 15 Jun 2018 12:43  
 Operator : JU/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF061118.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Fri Jun 15 11:11:47 2018



AutoFind: Scans 1597, 1598, 1599; Background Corrected with Scan 1592

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 35.1         | 42074      | PASS                |
| 68             | 69              | 0.00            | 2               | 1.7          | 659        | PASS                |
| 69             | 198             | 0.00            | 100             | 32.3         | 38798      | PASS                |
| 70             | 69              | 0.00            | 2               | 0.3          | 120        | PASS                |
| 127            | 198             | 10              | 80              | 42.5         | 51008      | PASS                |
| 197            | 198             | 0.00            | 2               | 0.4          | 452        | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 120008     | PASS                |
| 199            | 198             | 5               | 9               | 6.5          | 7832       | PASS                |
| 275            | 198             | 10              | 60              | 26.1         | 31285      | PASS                |
| 365            | 198             | 1               | 100             | 3.0          | 3573       | PASS                |
| 441            | 443             | 0.01            | 100             | 81.6         | 17120      | PASS                |
| 442            | 198             | 50              | 100             | 88.2         | 105869     | PASS                |
| 443            | 442             | 15              | 24              | 19.8         | 20989      | PASS                |