

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061520\
 Data File : BF120621.D
 Acq On : 15 Jun 2020 15:26
 Operator : JU/CG
 Sample : PB129539BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB129539BL

Quant Time: Jun 15 16:05:15 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 11 11:28:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.80	152	166029	20.00	ng	0.00
21) Naphthalene-d8	8.08	136	643966	20.00	ng	0.00
39) Acenaphthene-d10	9.83	164	336955	20.00	ng	-0.01
64) Phenanthrene-d10	11.32	188	661124	20.00	ng	0.00
76) Chrysene-d12	13.95	240	577268	20.00	ng	-0.01
86) Perylene-d12	15.39	264	616413	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.41	112	1116398	115.28	ng	0.00
7) Phenol-d6	6.43	99	1414026	117.23	ng	-0.01
23) Nitrobenzene-d5	7.36	82	896911	77.85	ng	0.00
42) 2,4,6-Tribromophenol	10.62	330	441442	112.68	ng	0.00
45) 2-Fluorobiphenyl	9.16	172	1660538	75.40	ng	0.00
79) Terphenyl-d14	12.90	244	2082590	79.69	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061520\
Data File : BF120621.D
Acq On : 15 Jun 2020 15:26
Operator : JU/CG
Sample : PB129539BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB129539BL

Quant Time: Jun 15 16:05:15 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 11 11:28:22 2020
Response via : Initial Calibration

