


Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF061520\  
 Data File : BF120617.D  
 Acq On : 15 Jun 2020 11:39  
 Operator : JU/CG  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF061020.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Thu Jun 11 11:28:22 2020



AutoFind: Scans 1612, 1613, 1614; Background Corrected with Scan 1607

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 36.2         | 22548      | PASS                |
| 68             | 69              | 0.00            | 2               | 1.7          | 336        | PASS                |
| 69             | 198             | 0.00            | 100             | 32.7         | 20344      | PASS                |
| 70             | 69              | 0.00            | 2               | 0.0          | 0          | PASS                |
| 127            | 198             | 10              | 80              | 46.9         | 29176      | PASS                |
| 197            | 198             | 0.00            | 2               | 0.3          | 196        | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 62253      | PASS                |
| 199            | 198             | 5               | 9               | 6.8          | 4206       | PASS                |
| 275            | 198             | 10              | 60              | 28.0         | 17420      | PASS                |
| 365            | 198             | 1               | 100             | 3.5          | 2182       | PASS                |
| 441            | 198             | 0.01            | 100             | 15.7         | 9769       | PASS                |
| 442            | 198             | 50              | 100             | 95.3         | 59346      | PASS                |
| 443            | 442             | 15              | 24              | 18.1         | 10767      | PASS                |