

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061522\
 Data File : BF128822.D
 Acq On : 15 Jun 2022 13:06
 Operator : CG\JU
 Sample : N3334-07DL 5X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-14A-1A-8DL

Quant Time: Jun 15 13:59:40 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 14 14:53:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.781	152	83968	20.000	ng	0.00
21) Naphthalene-d8	8.063	136	316661	20.000	ng	0.00
39) Acenaphthene-d10	9.828	164	162865	20.000	ng	0.00
64) Phenanthrene-d10	11.316	188	249023	20.000	ng	0.00
76) Chrysene-d12	13.969	240	180296	20.000	ng	# 0.00
86) Perylene-d12	15.445	264	169677	20.000	ng	0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.416	112	47941	9.094	ng	0.00
7) Phenol-d6	6.434	99	64545	10.030	ng	0.00
23) Nitrobenzene-d5	7.340	82	51078	7.510	ng	-0.01
42) 2,4,6-Tribromophenol	10.616	330	14061	7.307	ng	0.00
45) 2-Fluorobiphenyl	9.139	172	91570	7.876	ng	0.00
79) Terphenyl-d14	12.904	244	75963	7.321	ng	0.00
Target Compounds						
						Qvalue
31) Naphthalene	8.087	128	69358	4.268	ng	# 94
37) 2-Methylnaphthalene	8.781	142	209428	20.261	ng	100
38) 1-Methylnaphthalene	8.881	142	152900	15.310	ng	100
58) Fluorene	10.375	166	24173	2.333	ng	# 86
71) Phenanthrene	11.339	178	61547	4.928	ng	# 93
78) Pyrene	12.763	202	28318	2.111	ng	# 83
83) Chrysene	13.998	228	29635	2.766	ng	# 57

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061522\
Data File : BF128822.D
Acq On : 15 Jun 2022 13:06
Operator : CG\JU
Sample : N3334-07DL 5X
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-14A-1A-8DL

Quant Time: Jun 15 13:59:40 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 14 14:53:53 2022
Response via : Initial Calibration

