


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF061615\
 Data File : BF079988.D
 Acq On : 16 Jun 2015 13:45
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF061115.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jun 15 15:44:31 2015



AutoFind: Scans 800, 801, 802; Background Corrected with Scan 797

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.8	13176	PASS
68	69	0.00	2	1.2	143	PASS
69	198	0.00	100	35.6	12100	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	46.2	15708	PASS
197	198	0.00	2	0.4	126	PASS
198	198	100	100	100.0	33986	PASS
199	198	5	9	7.0	2380	PASS
275	198	10	60	23.9	8116	PASS
365	198	1	100	2.4	807	PASS
441	442	0.01	24	15.7	3614	PASS
442	198	50	100	67.7	23013	PASS
443	442	15	24	20.0	4598	PASS