

Data Path : Z:\HPCHEM1\BNA F\Data\BF061617\
 Data File : BF095951.D
 Acq On : 16 Jun 2017 12:39
 Operator : SJ/MA
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 16 23:21:19 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 16 17:12:03 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	72	0.00
2	1,4-Dioxane	0.597	0.590	1.2	68	0.00
3	Pyridine	1.403	1.480	-5.5	73	0.00
4	n-Nitrosodimethylamine	0.534	0.555	-3.9	73	0.00
5 S	2-Fluorophenol	1.255	1.388	-10.6	72	0.00
6	Aniline	1.966	2.036	-3.6	70	0.00
7 S	Phenol-d6	1.503	1.617	-7.6	71	0.00
8	2-Chlorophenol	1.335	1.411	-5.7	70	0.00
9	Benzaldehyde	1.014	0.944	6.9	63	0.00
10 C	Phenol	1.559	1.740	-11.6	72	0.00
11	bis(2-Chloroethyl)ether	1.235	1.381	-11.8	74	0.00
12	1,3-Dichlorobenzene	1.511	1.576	-4.3	74	0.00
13 C</						

Data Path : Z:\HPCHEM1\BNA F\Data\BF061617\
 Data File : BF095951.D
 Acq On : 16 Jun 2017 12:39
 Operator : SJ/MA
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 16 23:21:19 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 16 17:12:03 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.648	1.615	2.0	76	0.00
46	2-Chloronaphthalene	1.288	1.261	2.1	76	0.00
47	2-Nitroaniline	0.377	0.386	-2.4	74	0.00
48	Acenaphthylene	2.202	2.116	3.9	73	0.00
49	Dimethylphthalate	1.519	1.497	1.4	76	0.00
50	2,6-Dinitrotoluene	0.331	0.326	1.5	73	0.00
51 C	Acenaphthene	1.272	1.246	2.0	80	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF061617\
Data File : BF095951.D
Acq On : 16 Jun 2017 12:39
Operator : SJ/MA
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 16 23:21:19 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 16 17:12:03 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.197	1.124	6.1	91	0.00
89 C	Benzo(a)pyrene	1.151	1.152	-0.1	95	0.00
90	Dibenzo(a,h)anthracene	0.976	0.879	9.9	90	0.00
91	Benzo(a,h,i)perylene	0.995	0.898	9.7	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0