

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF061720\
 Data File : BF120632.D
 Acq On : 17 Jun 2020 14:46
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 17 15:36:46 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 11 11:28:22 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.553	0.506	8.5	90	-0.03
3	Pyridine	1.554	1.446	6.9	91	-0.04
4	n-Nitrosodimethylamine	0.645	0.623	3.4	94	-0.04
5 S	2-Fluorophenol	1.167	1.135	2.7	99	-0.01
6	Aniline	1.859	1.999	-7.5	101	0.00
7 S	Phenol-d6	1.453	1.567	-7.8	103	0.00
8	2-Chlorophenol	1.318	1.319	-0.1	98	0.00
9	Benzaldehyde	0.919	0.857	6.7	91	0.00
10 C	Phenol	1.550	1.711	-10.4	104	0.00
11	bis(2-Chloroethyl)ether	1.291	1.316	-1.9	100	0.00
12	1,3-Dichlorobenzene	1.401	1.422	-1.5	101	0.00
13 C	1,4-Dichlorobenzene	1.383	1.421	-2.7	106	0.00
14	1,2-Dichlorobenzene	1.347	1.359	-0.9	102	0.00
15	Benzyl Alcohol	1.030	1.156	-12.2	110	0.00
16	2,2'-oxybis(1-Chloropropane	1.859	2.021	-8.7	110	0.00
17	2-Methylphenol	1.124	1.217	-8.3	116	0.00
18	Hexachloroethane	0.510	0.515	-1.0	103	0.00
19 P	n-Nitroso-di-n-propylamine	0.843	0.972	-15.3	119	0.00
20	3+4-Methylphenols	1.405	1.448	-3.1	110	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
22	Acetophenone	0.447	0.440	1.6	110	0.00
23 S	Nitrobenzene-d5	0.358	0.344	3.9	106	0.00
24	Nitrobenzene	0.374	0.365	2.4	113	0.00
25	Isophorone	0.697	0.730	-4.7	116	0.00
26 C	2-Nitrophenol	0.198	0.192	3.0	102	0.00
27	2,4-Dimethylphenol	0.292	0.298	-2.1	110	0.00
28	bis(2-Chloroethoxy)methane	0.404	0.440	-8.9	110	0.00
29 C	2,4-Dichlorophenol	0.296	0.311	-5.1	105	0.00
30	1,2,4-Trichlorobenzene	0.329	0.323	1.8	100	0.00
31	Naphthalene	0.982	0.992	-1.0	106	0.00
32	Benzoic acid	0.224	0.247	-10.3	110	0.00
33	4-Chloroaniline	0.451	0.473	-4.9	113	0.00
34 C	Hexachlorobutadiene	0.192	0.190	1.0	110	0.00
35	Caprolactam	0.095	0.104	-9.5	127	0.00
36 C	4-Chloro-3-methylphenol	0.296	0.329	-11.1	125	0.00
37	2-Methylnaphthalene	0.641	0.705	-10.0	122	0.00
38	1-Methylnaphthalene	0.603	0.663	-10.0	120	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	138	0.00
40	1,2,4,5-Tetrachlorobenzene	0.597	0.578	3.2	122	0.00
41 P	Hexachlorocyclopentadiene	0.335	0.303	9.6	115	0.00
42 S	2,4,6-Tribromophenol	0.233	0.231	0.9	124	0.00
43 C	2,4,6-Trichlorophenol	0.424	0.421	0.7	126	0.00
44	2,4,5-Trichlorophenol	0.481	0.478	0.6	125	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF061720\
 Data File : BF120632.D
 Acq On : 17 Jun 2020 14:46
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 17 15:36:46 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 11 11:28:22 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.307	1.184	9.4	123	0.00
46	1,1'-Biphenyl	1.493	1.493	0.0	125	0.00
47	2-Chloronaphthalene	1.221	1.204	1.4	124	0.00
48	2-Nitroaniline	0.384	0.380	1.0	126	0.00
49	Acenaphthylene	1.885	1.879	0.3	124	0.00
50	Dimethylphthalate	1.440	1.430	0.7	127	0.00
51	2,6-Dinitrotoluene	0.326	0.323	0.9	125	0.00
52 C	Acenaphthene	1.124	1.132	-0.7	126	0.00
53	3-Nitroaniline	0.392	0.383	2.3	125	0.00
54 P	2,4-Dinitrophenol	0.175	0.162	7.4	121	0.00
55	Dibenzofuran	1.724	1.714	0.6	125	0.00
56 P	4-Nitrophenol	0.295	0.278	5.8	121	0.00
57	2,4-Dinitrotoluene	0.433	0.428	1.2	125	0.00
58	Fluorene	1.329	1.263	5.0	123	0.00
59	2,3,4,6-Tetrachlorophenol	0.378	0.382	-1.1	128	0.00
60	Diethylphthalate	1.419	1.407	0.8	124	0.00
61	4-Chlorophenyl-phenylether	0.686	0.657	4.2	123	0.00
62	4-Nitroaniline	0.389	0.374	3.9	121	0.00
63	Azobenzene	1.288	1.306	-1.4	127	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	135	0.00
65	4,6-Dinitro-2-methylphenol	0.120	0.115	4.2	119	0.00
66 c	n-Nitrosodiphenylamine	0.614	0.610	0.7	124	0.00
67	4-Bromophenyl-phenylether	0.227	0.228	-0.4	123	0.00
68	Hexachlorobenzene	0.244	0.244	0.0	124	0.00
69	Atrazine	0.190	0.152	20.0	100	0.00
70 C	Pentachlorophenol	0.136	0.141	-3.7	125	0.00
71	Phenanthrene	0.994	0.975	1.9	121	0.00
72	Anthracene	1.002	0.983	1.9	121	0.00
73	Carbazole	0.992	0.958	3.4	120	0.00
74	Di-n-butylphthalate	1.115	1.118	-0.3	122	0.00
75 C	Fluoranthene	1.137	1.074	5.5	110	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	126	0.00
77	Benzidine	0.676	0.640	5.3	103	0.00
78	Pyrene	1.417	1.414	0.2	113	0.00
79 S	Terphenyl-d14	0.905	0.867	4.2	113	0.00
80	Butylbenzylphthalate	0.629	0.618	1.7	112	0.00
81	Benzo(a)anthracene	1.202	1.184	1.5	116	0.00
82	3,3'-Dichlorobenzidine	0.459	0.461	-0.4	116	0.00
83	Chrysene	1.233	1.228	0.4	115	0.00
84	Bis(2-ethylhexyl)phthalate	0.768	0.785	-2.2	119	0.00
85 c	Di-n-octyl phthalate	1.465	1.465	0.0	112	0.00
86 I	Perylene-d12	1.000	1.000	0.0	120	0.00
87	Indeno(1,2,3-cd)pyrene	1.217	1.324	-8.8	123	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF061720\
Data File : BF120632.D
Acq On : 17 Jun 2020 14:46
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 17 15:36:46 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 11 11:28:22 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.191	1.140	4.3	118	0.00
89	Benzo(k)fluoranthene	1.054	1.055	-0.1	107	0.00
90 C	Benzo(a)pyrene	1.066	1.055	1.0	109	0.00
91	Dibenzo(a,h)anthracene	0.976	1.064	-9.0	123	-0.01
92	Benzo(q,h,i)perylene	0.983	1.085	-10.4	125	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0