

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061724\
 Data File : BF138199.D
 Acq On : 17 Jun 2024 12:30
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 17 13:02:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061224.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 13 01:52:47 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	112	0.00
2	1,4-Dioxane	40.000	40.056	-0.1	109	0.00
3	Pyridine	40.000	40.293	-0.7	112	0.00
4	n-Nitrosodimethylamine	40.000	38.740	3.1	104	0.00
5 S	2-Fluorophenol	80.000	78.635	1.7	107	0.00
6	Aniline	40.000	40.474	-1.2	109	0.00
7 S	Phenol-d6	80.000	77.778	2.8	106	0.00
8	2-Chlorophenol	40.000	39.407	1.5	106	0.00
9	Benzaldehyde	40.000	40.203	-0.5	117	0.00
10 C	Phenol	40.000	40.332	-0.8	109	0.00
11	bis(2-Chloroethyl)ether	40.000	39.679	0.8	109	0.00
12	1,3-Dichlorobenzene	40.000	37.765	5.6	103	0.00
13 C	1,4-Dichlorobenzene	40.000	38.003	5.0	104	0.00
14	1,2-Dichlorobenzene	40.000	37.907	5.2	103	0.00
15	Benzyl Alcohol	40.000	40.621	-1.6	108	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.289	1.8	107	0.00
17	2-Methylphenol	40.000	40.030	-0.1	108	0.00
18	Hexachloroethane	40.000	39.946	0.1	108	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.547	1.1	109	0.00
20	3+4-Methylphenols	40.000	39.691	0.8	107	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	116	0.00
22	Acetophenone	40.000	37.798	5.5	108	0.00
23 S	Nitrobenzene-d5	80.000	77.952	2.6	108	0.00
24	Nitrobenzene	40.000	39.319	1.7	110	0.00
25	Isophorone	40.000	38.486	3.8	110	0.00
26 C	2-Nitrophenol	40.000	42.033	-5.1	110	0.00
27	2,4-Dimethylphenol	40.000	38.014	5.0	107	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.956	2.6	111	0.00
29 C	2,4-Dichlorophenol	40.000	37.909	5.2	106	0.00
30	1,2,4-Trichlorobenzene	40.000	36.961	7.6	104	0.00
31	Naphthalene	40.000	37.141	7.1	105	0.00
32	Benzoic acid	40.000	38.195	4.5	110	0.00
33	4-Chloroaniline	40.000	38.327	4.2	110	0.00
34 C	Hexachlorobutadiene	40.000	37.098	7.3	105	0.00
35	Caprolactam	40.000	40.554	-1.4	113	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.384	1.5	112	0.00
37	2-Methylnaphthalene	40.000	37.060	7.3	105	0.00
38	1-Methylnaphthalene	40.000	37.098	7.3	106	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	118	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	35.824	10.4	104	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.241	4.4	104	0.00
42 S	2,4,6-Tribromophenol	80.000	74.925	6.3	105	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.826	5.4	105	0.00
44	2,4,5-Trichlorophenol	40.000	38.362	4.1	109	0.00
45 S	2-Fluorobiphenyl	80.000	69.044	13.7	103	0.00
46	1,1'-Biphenyl	40.000	36.165	9.6	106	0.00
47	2-Chloronaphthalene	40.000	37.084	7.3	108	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061724\
 Data File : BF138199.D
 Acq On : 17 Jun 2024 12:30
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 17 13:02:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061224.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 13 01:52:47 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.357	-5.9	114	0.00
49	Acenaphthylene	40.000	36.631	8.4	106	0.00
50	Dimethylphthalate	40.000	38.075	4.8	111	0.00
51	2,6-Dinitrotoluene	40.000	41.389	-3.5	113	0.00
52 C	Acenaphthene	40.000	36.761	8.1	106	0.00
53	3-Nitroaniline	40.000	40.940	-2.3	113	0.00
54 P	2,4-Dinitrophenol	40.000	37.769	5.6	109	0.00
55	Dibenzofuran	40.000	36.531	8.7	105	0.00
56 P	4-Nitrophenol	40.000	41.326	-3.3	112	0.00
57	2,4-Dinitrotoluene	40.000	42.749	-6.9	115	0.00
58	Fluorene	40.000	36.044	9.9	106	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.696	5.8	106	0.00
60	Diethylphthalate	40.000	39.029	2.4	113	0.00
61	4-Chlorophenyl-phenylether	40.000	36.592	8.5	107	0.00
62	4-Nitroaniline	40.000	41.270	-3.2	114	0.00
63	Azobenzene	40.000	38.698	3.3	113	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	121	0.00
65	4,6-Dinitro-2-methylphenol	40.000	35.859	10.4	109	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.396	9.0	107	0.00
67	4-Bromophenyl-phenylether	40.000	36.347	9.1	106	0.00
68	Hexachlorobenzene	40.000	35.672	10.8	105	0.00
69	Atrazine	40.000	37.222	6.9	110	0.00
70 C	Pentachlorophenol	40.000	39.219	2.0	108	0.00
71	Phenanthrene	40.000	36.127	9.7	107	0.00
72	Anthracene	40.000	36.524	8.7	108	0.00
73	Carbazole	40.000	37.038	7.4	109	0.00
74	Di-n-butylphthalate	40.000	39.798	0.5	116	0.00
75 C	Fluoranthene	40.000	37.060	7.3	110	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	127	0.00
77	Benzydine	40.000	70.415	-76.0#	127	0.00
78	Pyrene	40.000	37.241	6.9	111	0.00
79 S	Terphenyl-d14	80.000	72.518	9.4	110	0.00
80	Butylbenzylphthalate	40.000	44.280	-10.7	126	0.00
81	Benzo(a)anthracene	40.000	37.221	6.9	115	0.00
82	3,3'-Dichlorobenzidine	40.000	39.373	1.6	121	0.00
83	Chrysene	40.000	37.613	6.0	117	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	45.573	-13.9	134	0.00
85 c	Di-n-octyl phthalate	40.000	41.047	-2.6	129	0.00
86 I	Perylene-d12	20.000	20.000	0.0	119	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.689	5.8	98	0.00
88	Benzo(b)fluoranthene	40.000	37.814	5.5	113	0.00
89	Benzo(k)fluoranthene	40.000	37.935	5.2	113	0.00
90 C	Benzo(a)pyrene	40.000	37.893	5.3	109	0.00
91	Dibenzo(a,h)anthracene	40.000	38.207	4.5	99	0.00
92	Benzo(g,h,i)perylene	40.000	37.641	5.9	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF061724\
Data File : BF138199.D
Acq On : 17 Jun 2024 12:30
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 17 13:02:23 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF061224.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 13 01:52:47 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0