

Data Path : Z:\HPCHEM1\BNA F\DATA\BF061816\
 Data File : BF088128.D
 Acq On : 18 Jun 2016 12:21
 Operator : UM/SJ
 Sample : H3501-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SS1-0-6

Manual Integrations
 APPROVED

sohil
 6/20/2016 6:56:51 PM

Quant Time: Jun 20 13:16:43 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 17 16:09:13 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.68	152	108827	20.00	ng	-0.01
21) Naphthalene-d8	7.96	136	393371	20.00	ng	-0.02
38) Acenaphthene-d10	9.71	164	170181	20.00	ng	-0.02
63) Phenanthrene-d10	11.20	188	270517	20.00	ng	-0.01
75) Chrysene-d12	13.84	240	171381	20.00	ng	-0.01
86) Perylene-d12	15.22	264	158168	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.29	112	710576	111.62	ng	0.01
7) Phenol-d6	6.35	99	986800	127.91	ng	0.00
23) Nitrobenzene-d5	7.26	82	555467	82.46	ng	-0.02
41) 2,4,6-Tribromophenol	10.51	330	179997	100.17	ng	-0.01
44) 2-Fluorobiphenyl	9.05	172	872021	80.18	ng	-0.02
78) Terphenyl-d14	12.79	244	552074	73.30	ng	-0.01
Target Compounds						Qvalue
32) Benzoic acid	7.74	122	9160	2.58	ng	94
49) Dimethylphthalate	9.45	163	129092	10.47	ng	98
70) Phenanthrene	11.22	178	37716	2.66	ng	97
74) Fluoranthene	12.41	202	89541	5.98	ng	93
77) Pyrene	12.64	202	85159m	6.70	ng	
79) Butylbenzylphthalate	13.27	149	196758	34.99	ng	98
80) Benzo(a)anthracene	13.83	228	35465	3.63	ng	# 87
82) Chrysene	13.86	228	41055m	4.47	ng	
83) Bis(2-ethylhexyl)phthalate	13.83	149	94160	13.76	ng	# 99
85) Indeno(1,2,3-cd)pyrene	16.54	276	27852	3.26	ng	# 100
87) Benzo(b)fluoranthene	14.85	252	64971m	5.89	ng	
88) Benzo(k)fluoranthene	14.86	252	22579m	2.72	ng	
91) Benzo(a,h,i)perylene	16.93	276	31831	3.74	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF061816\
Data File : BF088128.D
Acq On : 18 Jun 2016 12:21
Operator : UM/SJ
Sample : H3501-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SS1-0-6

Manual Integrations
APPROVED

sohil
6/20/2016 6:56:51 PM

Quant Time: Jun 20 13:16:43 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060716.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 17 16:09:13 2016
Response via : Initial Calibration

