


Data Path : Z:\HPCHEM1\BNA_F\Data\BF061915\
 Data File : BF080069.D
 Acq On : 19 Jun 2015 22:48
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF061715.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jun 17 16:41:53 2015



AutoFind: Scans 790, 791, 792; Background Corrected with Scan 787

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.9	9155	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	37.5	8180	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	48.6	10598	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	21829	PASS
199	198	5	9	6.9	1501	PASS
275	198	10	60	22.0	4804	PASS
365	198	1	100	2.1	469	PASS
441	442	0.01	24	16.5	2677	PASS
442	198	50	100	74.3	16224	PASS
443	442	15	24	19.2	3113	PASS