

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061920\
 Data File : BF120662.D
 Acq On : 19 Jun 2020 14:30
 Operator : JU/CG
 Sample : L3055-02
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CENTER

Manual Integrations
 APPROVED

mohammad
 6/22/2020 3:24:50 PM

Quant Time: Jun 19 15:39:45 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 11 11:28:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.80	152	110287	20.00	ng	0.00
21) Naphthalene-d8	8.08	136	403283	20.00	ng	0.00
39) Acenaphthene-d10	9.83	164	213759	20.00	ng	0.00
64) Phenanthrene-d10	11.32	188	422866	20.00	ng	0.00
76) Chrysene-d12	13.96	240	465599	20.00	ng	0.00
86) Perylene-d12	15.40	264	492703	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.40	112	141121	21.94	ng	-0.01
7) Phenol-d6	6.43	99	368086	45.94	ng	-0.01
23) Nitrobenzene-d5	7.36	82	425520	58.98	ng	-0.01
42) 2,4,6-Tribromophenol	10.62	330	32059	12.90	ng	0.00
45) 2-Fluorobiphenyl	9.16	172	874155	62.57	ng	0.00
79) Terphenyl-d14	12.91	244	1322888	62.76	ng	0.00
Target Compounds						
38) 1-Methylnaphthalene	8.89	142	26376m	2.168	ng	Qvalue
50) Dimethylphthalate	9.55	163	77173	5.013	ng	# 98
71) Phenanthrene	11.34	178	71842	3.418	ng	97
75) Fluoranthene	12.53	202	114789	4.776	ng	97
78) Pyrene	12.76	202	103534	3.139	ng	98
88) Benzo(b)fluoranthene	14.99	252	60668m	2.067	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061920\
Data File : BF120662.D
Acq On : 19 Jun 2020 14:30
Operator : JU/CG
Sample : L3055-02
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CENTER

Manual Integrations
APPROVED

mohammad
6/22/2020 3:24:50 PM

Quant Time: Jun 19 15:39:45 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 11 11:28:22 2020
Response via : Initial Calibration

