

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061920\
 Data File : BF120670.D
 Acq On : 19 Jun 2020 18:19
 Operator : JU/CG
 Sample : L3020-17
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-27(16.5-18.5)

Quant Time: Jun 20 02:31:15 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 11 11:28:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.80	152	108317	20.00	ng	0.00
21) Naphthalene-d8	8.09	136	286963	20.00	ng	0.00
39) Acenaphthene-d10	9.83	164	238024	20.00	ng	0.00
64) Phenanthrene-d10	11.32	188	507514	20.00	ng	0.00
76) Chrysene-d12	13.96	240	428823	20.00	ng	0.00
86) Perylene-d12	15.40	264	494534	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.41	112	704864	111.57	ng	0.00
7) Phenol-d6	6.43	99	777214	98.77	ng	0.00
23) Nitrobenzene-d5	7.37	82	504892	98.35	ng	0.00
42) 2,4,6-Tribromophenol	10.62	330	309980	112.01	ng	0.00
45) 2-Fluorobiphenyl	9.16	172	1155242	74.26	ng	0.00
79) Terphenyl-d14	12.91	244	1442730	74.32	ng	0.00
Target Compounds						
31) Naphthalene	8.12	128	2630790	186.702	ng	96
37) 2-Methylnaphthalene	8.81	142	2120835	230.561	ng	96
38) 1-Methylnaphthalene	8.90	142	1250649	144.489	ng	# 97
46) 1,1'-Biphenyl	9.26	154	97903	5.508	ng	98
50) Dimethylphthalate	9.55	163	278588	16.251	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061920\
Data File : BF120670.D
Acq On : 19 Jun 2020 18:19
Operator : JU/CG
Sample : L3020-17
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-27(16.5-18.5)

Quant Time: Jun 20 02:31:15 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 11 11:28:22 2020
Response via : Initial Calibration

