

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061920\
 Data File : BF120675.D
 Acq On : 19 Jun 2020 20:43
 Operator : JU/CG
 Sample : L3034-14
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-5

Quant Time: Jun 20 02:05:17 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 11 11:28:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.80	152	143626	20.00	ng	0.00
21) Naphthalene-d8	8.08	136	490554	20.00	ng	0.00
39) Acenaphthene-d10	9.83	164	278593	20.00	ng	0.00
64) Phenanthrene-d10	11.32	188	461340	20.00	ng	0.00
76) Chrysene-d12	13.96	240	412809	20.00	ng	0.00
86) Perylene-d12	15.41	264	423083	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.41	112	958202	114.38	ng	0.00
7) Phenol-d6	6.43	99	1017715	97.54	ng	0.00
23) Nitrobenzene-d5	7.36	82	645318	73.53	ng	0.00
42) 2,4,6-Tribromophenol	10.62	330	318945	98.47	ng	0.00
45) 2-Fluorobiphenyl	9.16	172	1201777	66.00	ng	0.00
79) Terphenyl-d14	12.91	244	1584298	84.77	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.55	163	93751	4.673	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF061920\
Data File : BF120675.D
Acq On : 19 Jun 2020 20:43
Operator : JU/CG
Sample : L3034-14
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-5

Quant Time: Jun 20 02:05:17 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 11 11:28:22 2020
Response via : Initial Calibration

