

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF061925\  
 Data File : BF142796.D  
 Acq On : 19 Jun 2025 21:10  
 Operator : RC/JU  
 Sample : SSTDICV040  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 ICVBF061925

Quant Time: Jun 24 15:59:56 2025  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF061925.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Jun 24 05:28:21 2025  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
Internal Standards							
1) 1,4-Dichlorobenzene-d4	6.881	152	74183	20.000	ng	0.00	
21) Naphthalene-d8	8.163	136	279139	20.000	ng	0.00	
39) Acenaphthene-d10	9.922	164	142968	20.000	ng	0.00	
64) Phenanthrene-d10	11.410	188	228099	20.000	ng	0.00	
76) Chrysene-d12	14.057	240	122573	20.000	ng	0.00	
86) Perylene-d12	15.545	264	153363	20.000	ng	0.00	
System Monitoring Compounds							
5) 2-Fluorophenol	5.493	112	361762	81.192	ng	0.00	
7) Phenol-d6	6.510	99	421918	80.262	ng	0.00	
23) Nitrobenzene-d5	7.445	82	373199	73.334	ng	0.00	
42) 2,4,6-Tribromophenol	10.716	330	117795	80.065	ng	0.00	
45) 2-Fluorobiphenyl	9.239	172	785440	72.171	ng	0.00	
79) Terphenyl-d14	12.998	244	628420	70.839	ng	0.00	
Target Compounds							
2) 1,4-Dioxane	2.634	88	70139	39.357	ng	100	Qvalue
3) Pyridine	3.404	79	185173	41.446	ng	99	
4) n-Nitrosodimethylamine	3.357	42	96831	41.912	ng	99	
6) Aniline	6.545	93	296519	42.392	ng	99	
8) 2-Chlorophenol	6.663	128	202041	42.036	ng	100	
9) Benzaldehyde	6.434	77	148174	47.511	ng	99	
10) Phenol	6.522	94	244839	41.901	ng	99	
11) bis(2-Chloroethyl)ether	6.616	93	176058	42.157	ng	99	
12) 1,3-Dichlorobenzene	6.822	146	228820	42.568	ng	99	
13) 1,4-Dichlorobenzene	6.898	146	229234	42.268	ng	99	
14) 1,2-Dichlorobenzene	7.051	146	218720	42.520	ng	99	
15) Benzyl Alcohol	7.022	79	160044	42.114	ng	100	
16) 2,2'-oxybis(1-Chloropr...	7.157	45	284352	42.379	ng	100	
17) 2-Methylphenol	7.134	107	154746	42.485	ng	98	
18) Hexachloroethane	7.392	117	80788	42.589	ng	97	
19) n-Nitroso-di-n-propyla...	7.292	70	127308	41.640	ng	99	
20) 3+4-Methylphenols	7.287	107	194841	42.904	ng	98	
22) Acetophenone	7.292	105	261474	42.478	ng	99	
24) Nitrobenzene	7.463	77	193310	41.968	ng	99	
25) Isophorone	7.704	82	338001	41.403	ng	99	
26) 2-Nitrophenol	7.781	139	108455	43.224	ng	100	
27) 2,4-Dimethylphenol	7.816	122	183273	42.168	ng	99	
28) bis(2-Chloroethoxy)met...	7.910	93	217774	41.960	ng	99	
29) 2,4-Dichlorophenol	8.022	162	166286	42.196	ng	99	
30) 1,2,4-Trichlorobenzene	8.104	180	182768	42.177	ng	100	
31) Naphthalene	8.187	128	576847	42.218	ng	100	
32) Benzoic acid	7.922	122	101203	45.685	ng	95	
33) 4-Chloroaniline	8.234	127	232045	41.766	ng	98	
34) Hexachlorobutadiene	8.304	225	113068	42.657	ng	99	
35) Caprolactam	8.604	113	42561	40.928	ng	98	
36) 4-Chloro-3-methylphenol	8.716	107	164064	41.890	ng	99	
37) 2-Methylnaphthalene	8.881	142	353484	41.730	ng	100	
38) 1-Methylnaphthalene	8.981	142	364378	41.554	ng	100	
40) 1,2,4,5-Tetrachloroben...	9.045	216	182379	43.233	ng	99	
41) Hexachlorocyclopentadiene	9.034	237	108592	45.361	ng	97	
43) 2,4,6-Trichlorophenol	9.157	196	116946	42.543	ng	98	

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF061925\  
 Data File : BF142796.D  
 Acq On : 19 Jun 2025 21:10  
 Operator : RC/JU  
 Sample : SSTDICV040  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 ICVBF061925

Quant Time: Jun 24 15:59:56 2025  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF061925.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Jun 24 05:28:21 2025  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.198	196	124558	43.045	ng	99
46) 1,1'-Biphenyl	9.345	154	476133	42.159	ng	100
47) 2-Chloronaphthalene	9.369	162	354405	41.817	ng	100
48) 2-Nitroaniline	9.463	65	99137	41.724	ng	97
49) Acenaphthylene	9.786	152	580204	41.586	ng	100
50) Dimethylphthalate	9.645	163	381573	41.232	ng	100
51) 2,6-Dinitrotoluene	9.704	165	84111	41.386	ng	99
52) Acenaphthene	9.957	154	356031	41.825	ng	100
53) 3-Nitroaniline	9.875	138	92889	41.512	ng	99
54) 2,4-Dinitrophenol	9.981	184	44513	43.742	ng	99
55) Dibenzofuran	10.128	168	502290	41.150	ng	100
56) 4-Nitrophenol	10.033	139	62127	40.556	ng	97
57) 2,4-Dinitrotoluene	10.110	165	111649	41.358	ng	100
58) Fluorene	10.475	166	391356	41.052	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.245	232	97630	41.857	ng	97
60) Diethylphthalate	10.339	149	371453	40.937	ng	100
61) 4-Chlorophenyl-phenyle...	10.463	204	185884	40.862	ng	99
62) 4-Nitroaniline	10.486	138	83143	40.628	ng	99
63) Azobenzene	10.622	77	332701	41.025	ng	99
65) 4,6-Dinitro-2-methylph...	10.516	198	59797	43.341	ng	94
66) n-Nitrosodiphenylamine	10.580	169	331671	41.955	ng	99
67) 4-Bromophenyl-phenylether	10.951	248	112604	43.376	ng	96
68) Hexachlorobenzene	11.022	284	122976	42.632	ng	98
69) Atrazine	11.104	200	87455	42.840	ng	99
70) Pentachlorophenol	11.216	266	62292	43.337	ng	99
71) Phenanthrene	11.433	178	514049	41.603	ng	99
72) Anthracene	11.486	178	532397	42.013	ng	99
73) Carbazole	11.645	167	446814	40.858	ng	100
74) Di-n-butylphthalate	11.969	149	481078	42.228	ng	100
75) Fluoranthene	12.627	202	473431	40.372	ng	100
77) Benzidine	12.745	184	218884	47.530	ng	99
78) Pyrene	12.857	202	468692	40.794	ng	100
80) Butylbenzylphthalate	13.469	149	154743	46.431	ng	99
81) Benzo(a)anthracene	14.045	228	346202	41.868	ng	100
82) 3,3'-Dichlorobenzidine	14.010	252	124771	46.521	ng	99
83) Chrysene	14.080	228	321976	43.637	ng	100
84) Bis(2-ethylhexyl)phtha...	14.027	149	230046	47.976	ng	99
85) Di-n-octyl phthalate	14.645	149	425193	47.026	ng	99
87) Indeno(1,2,3-cd)pyrene	17.062	276	491326	42.060	ng	100
88) Benzo(b)fluoranthene	15.110	252	362077	40.801	ng	100
89) Benzo(k)fluoranthene	15.139	252	347044	41.800	ng	100
90) Benzo(a)pyrene	15.486	252	361077	42.117	ng	100
91) Dibenzo(a,h)anthracene	17.074	278	399293	42.069	ng	99
92) Benzo(g,h,i)perylene	17.515	276	397566	42.006	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF061925\  
 Data File : BF142796.D  
 Acq On : 19 Jun 2025 21:10  
 Operator : RC/JU  
 Sample : SSTDICV040  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 ICVBF061925

Quant Time: Jun 24 15:59:56 2025  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF061925.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Jun 24 05:28:21 2025  
 Response via : Initial Calibration

