

Data Path : Z:\HPCHEM1\BNA F\DATA\BF062016\
 Data File : BF088211.D
 Acq On : 21 Jun 2016 10:54
 Operator : UM/SJ
 Sample : H3501-02DL 2X
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SS2-0-6DL

Quant Time: Jun 21 13:08:51 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 20 17:39:32 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.66	152	80469	20.00	ng	-0.01
21) Naphthalene-d8	7.95	136	318523	20.00	ng	0.00
38) Acenaphthene-d10	9.70	164	119017	20.00	ng	0.00
63) Phenanthrene-d10	11.18	188	176779	20.00	ng	0.00
75) Chrysene-d12	13.83	240	99425	20.00	ng	0.02
86) Perylene-d12	15.23	264	105626	20.00	ng	0.06
System Monitoring Compounds						
5) 2-Fluorophenol	5.26	112	296623	63.02	ng	0.00
7) Phenol-d6	6.32	99	359612	63.04	ng	-0.01
23) Nitrobenzene-d5	7.23	82	210082	38.51	ng	-0.01
41) 2,4,6-Tribromophenol	10.49	330	54814	43.62	ng	0.00
44) 2-Fluorobiphenyl	9.03	172	368956	46.17	ng	-0.01
78) Terphenyl-d14	12.78	244	194307	44.47	ng	0.01
Target Compounds						
49) Dimethylphthalate	9.43	163	42799	4.96	ng	Qvalue 100
73) Di-n-butylphthalate	11.75	149	31797	2.87	ng	# 83
79) Butylbenzylphthalate	13.26	149	54572	16.73	ng	89
83) Bis(2-ethylhexyl)phthalate	13.82	149	237702	59.86	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF062016\
Data File : BF088211.D
Acq On : 21 Jun 2016 10:54
Operator : UM/SJ
Sample : H3501-02DL 2X
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SS2-0-6DL

Quant Time: Jun 21 13:08:51 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060716.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 20 17:39:32 2016
Response via : Initial Calibration

