


Data Path : Z:\HPCHEM1\BNA F\DATA\BF062016\
 Data File : BF088190.D
 Acq On : 21 Jun 2016 00:46
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF060716.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jun 20 17:39:32 2016



AutoFind: Scans 862, 863, 864; Background Corrected with Scan 859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.8	8413	PASS
68	69	0.00	2	1.4	121	PASS
69	198	0.00	100	34.0	8721	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.8	12788	PASS
197	198	0.00	2	0.5	117	PASS
198	198	100	100	100.0	25677	PASS
199	198	5	9	6.6	1690	PASS
275	198	10	60	26.4	6781	PASS
365	198	1	100	2.7	684	PASS
441	442	0.01	24	15.4	3780	PASS
442	198	50	100	95.7	24570	PASS
443	442	15	24	18.8	4610	PASS