

Data Path : Z:\HPCHEM1\BNA F\DATA\BF062017\
 Data File : BF096055.D
 Acq On : 20 Jun 2017 20:47
 Operator : SJ/MA
 Sample : I3759-01
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SU-02-061617

Manual Integrations
 APPROVED

mohammad
 6/21/2017 2:16:48 PM

Quant Time: Jun 21 01:43:54 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 20 16:05:47 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.29	152	67959	20.00	ng	0.00
21) Naphthalene-d8	9.32	136	265228	20.00	ng	0.00
38) Acenaphthene-d10	12.15	164	102572	20.00	ng	0.00
63) Phenanthrene-d10	14.54	188	159417	20.00	ng	0.00
75) Chrysene-d12	18.29	240	104051	20.00	ng	0.02
86) Perylene-d12	19.96	264	88936	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.18	112	519761	121.87	ng	0.00
7) Phenol-d6	6.87	99	619138	121.26	ng	0.00
23) Nitrobenzene-d5	8.21	82	343454	80.42	ng	-0.01
41) 2,4,6-Tribromophenol	13.45	330	94916	104.59	ng	0.00
44) 2-Fluorobiphenyl	11.10	172	651617	88.40	ng	0.00
78) Terphenyl-d14	16.93	244	434946	103.74	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	9.36	128	44230	3.32	ng	98
37) 2-Methylnaphthalene	10.49	142	30088	3.35	ng	94
48) Acenaphthylene	11.92	152	396627	35.12	ng	98
49) Dimethylphthalate	11.80	163	118992	15.28	ng	99
51) Acenaphthene	12.20	154	73821	11.32	ng	99
54) Dibenzofuran	12.49	168	59500	6.48	ng	97
57) Fluorene	13.04	166	128543	16.09	ng	93
70) Phenanthrene	14.59	178	1181737	128.47	ng	99
71) Anthracene	14.67	178	500453	52.11	ng	98
72) Carbazole	14.97	167	33005	3.53	ng	# 93
74) Fluoranthene	16.40	202	1905685	209.32	ng	99
77) Pyrene	16.70	202	2118813m	272.91	ng	
80) Benzo(a)anthracene	18.28	228	982766	162.45	ng	94
82) Chrysene	18.33	228	916011	151.52	ng	97
85) Indeno(1,2,3-cd)pyrene	21.12	276	356501	68.92	ng	# 91
87) Benzo(b)fluoranthene	19.56	252	835004m	149.94	ng	
88) Benzo(k)fluoranthene	19.59	252	304575m	57.22	ng	
89) Benzo(a)pyrene	19.91	252	658694	128.69	ng	96
90) Dibenzo(a,h)anthracene	21.12	278	146277m	33.69	ng	
91) Benzo(a,h,i)perylene	21.45	276	340818	77.00	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA F\DATA\BF062017\
Data File : BF096055.D
Acq On : 20 Jun 2017 20:47
Operator : SJ/MA
Sample : I3759-01
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SU-02-061617

Manual Integrations
APPROVED

mohammad
6/21/2017 2:16:48 PM

Quant Time: Jun 21 01:43:54 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 20 16:05:47 2017
Response via : Initial Calibration





#1


1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.29 min Scan# 987
Delta R.T. 0.00 min
Lab File: BF096055.D
Acq: 20 Jun 2017 20:47

Instrument :
BNA_F
ClientSampleId :
SU-02-061617




Tot Ion	Ratio	Lower	Upper
152	100		
150	153.5	126.2	189.2
115	59.2	52.2	78.2

Manual Integrations
APPROVED

mohammad
6/21/2017 2:16:48 PM




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 121.87 ng
RT: 5.18 min Scan# 628
Delta R.T. 0.00 min
Lab File: BF096055.D
Acq: 20 Jun 2017 20:47

Tgt Ion	Ratio	Lower	Upper
112	100		
64	54.6	46.0	69.0
63	26.9	23.4	35.0



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF0
Ion 63.00 (62.70 to 63.70): BF0





#7

Phenol-d6

Concen: 121.26 ng

RT: 6.87 min Scan# 915

Delta R.T. -0.01 min

Lab File: BF096055.D

Acq: 20 Jun 2017 20:47

Instrument :

BNA_F





ClientSampleId :

SU-02-061617

Tot Ion:	99	Resp:	619138
Ion	Ratio	Lower	Upper
99	100		
42	16.1	13.5	20.3
71	32.1	24.5	36.7

Manual Integrations
APPROVED

mohammad
6/21/2017 2:16:48 PM



#21

Naphthalene-d8

Concen: 20.00 ng




RT: 9.32 min Scan# 1332


Delta R.T. -0.01 min

Lab File: BF096055.D

Acq: 20 Jun 2017 20:47

Tgt Ion:	136	Resp:	265228
Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.5	12.7
54	6.9	4.9	7.3
68	5.7	4.1	6.1





#23

Nitrobenzene-d5

Concen: 80.42 ng

RT: 8.21 min Scan# 1143

Delta R.T. -0.01 min

Lab File: BF096055.D

Acq: 20 Jun 2017 20:47

Instrument :

BNA_F





ClientSampled :

SU-02-061617

Tot Ion:	82	Resp:	343454
Ion Ratio	Lower	Upper	
82	100		
128	47.8	34.0	51.0
54	48.4	42.2	63.2

Manual Integrations
APPROVED

mohammad
6/21/2017 2:16:48 PM



#31

Naphthalene

Concen: 3.32 ng




RT: 9.36 min Scan# 1338


Delta R.T. -0.01 min

Lab File: BF096055.D

Acq: 20 Jun 2017 20:47

Tgt Ion:	128	Resp:	44230
Ion Ratio	Lower	Upper	
128	100		
129	10.6	9.0	13.6
127	14.2	10.8	16.2









#37
2-Methylnaphthalene
Concen: 3.35 ng
RT: 10.49 min Scan# 1531
Delta R.T. 0.01 min
Lab File: BF096055.D
Acq: 20 Jun 2017 20:47

Instrument :
BNA_F
ClientSampleId :
SU-02-061617

Tot Ion	Ratio	Lower	Upper
142	100		
141	95.9	71.3	106.9
115	32.8	27.1	40.7




Manual Integrations
APPROVED

mohammad
6/21/2017 2:16:48 PM



#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 12.15 min Scan# 1812
Delta R.T. -0.01 min
Lab File: BF096055.D
Acq: 20 Jun 2017 20:47

Tgt Ion	Ratio	Lower	Upper
164	100		
162	108.0	82.6	123.8
160	47.1	36.2	54.2









#41
 2,4,6-Tribromophenol
 Concen: 104.59 ng
 RT: 13.45 min Scan# 2033
 Delta R.T. -0.01 min
 Lab File: BF096055.D
 Acq: 20 Jun 2017 20:47

Instrument :
 BNA_F
 ClientSampleId :
 SU-02-061617

Tot Ion	Ratio	Lower	Upper
330	100		
332	96.9	76.6	115.0
141	40.2	31.4	47.2




Manual Integrations
 APPROVED


mohammad
 6/21/2017 2:16:48 PM



#44
 2-Fluorobiphenyl
 Concen: 88.40 ng
 RT: 11.10 min Scan# 1635
 Delta R.T. -0.01 min
 Lab File: BF096055.D
 Acq: 20 Jun 2017 20:47

Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.2	29.6	44.4
170	22.9	19.8	29.8





#48

Acenaphthylene

Concen: 35.12 ng

RT: 11.92 min Scan# 1773

Delta R.T. -0.01 min

Lab File: BF096055.D

Acq: 20 Jun 2017 20:47

Instrument :

BNA_F





ClientSampleId :

SU-02-061617

Tot Ion	Ratio	Lower	Upper
152	100		
151	20.0	16.8	25.2
153	12.9	10.8	16.2

Manual Integrations
APPROVED

mohammad
6/21/2017 2:16:48 PM



#49

Dimethylphthalate

Concen: 15.28 ng




RT: 11.80 min Scan# 1753


Delta R.T. -0.02 min

Lab File: BF096055.D

Acq: 20 Jun 2017 20:47

Tgt Ion	Ratio	Lower	Upper
163	100		
194	3.4	2.6	4.0
164	9.9	8.4	12.6





#51

Acenaphthene

Concen: 11.32 ng

RT: 12.20 min Scan# 1821

Delta R.T. -0.01 min

Lab File: BF096055.D

Acq: 20 Jun 2017 20:47

Instrument :

BNA_F

ClientSampleId :

SU-02-061617

Tot Ion:154 Resp: 73821

Ion Ratio Lower Upper

154 100


153 115.2 90.5 135.7

152 54.7 43.6 65.4

Manual Integrations
APPROVED

mohammad


6/21/2017 2:16:48 PM





Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



Time-->



#54

Dibenzofuran

Concen: 6.48 ng

RT: 12.49 min Scan# 1870

Delta R.T. -0.01 min

Lab File: BF096055.D

Acq: 20 Jun 2017 20:47


Tgt Ion:168 Resp: 59500

Ion Ratio Lower Upper

168 100

139 36.0 30.6 45.8


169 14.3 10.8 16.2




Abundance Ion 168.00 (167.70 to 168.70): E


Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E



Time-->









#57
Fluorene
 Concen: 16.09 ng
 RT: 13.04 min Scan# 1964
 Delta R.T. -0.01 min
 Lab File: BF096055.D
 Acq: 20 Jun 2017 20:47

Instrument :
 BNA_F
 ClientSampleId :
 SU-02-061617

Tot Ion	Ratio	Lower	Upper
166	100		
165	106.1	78.8	118.2
167	15.1	10.6	16.0




Manual Integrations
 APPROVED


mohammad
 6/21/2017 2:16:48 PM



#63
Phenanthrene-d10
 Concen: 20.00 ng
 RT: 14.54 min Scan# 2219
 Delta R.T. 0.00 min
 Lab File: BF096055.D
 Acq: 20 Jun 2017 20:47

Tgt Ion	Ratio	Lower	Upper
188	100		
94	11.6	9.4	14.2
80	11.9	10.2	15.2









#70
Phenanthrene
Concen: 128.47 ng
RT: 14.59 min Scan# 2228
Delta R.T. 0.01 min
Lab File: BF096055.D
Acq: 20 Jun 2017 20:47

Instrument :
BNA_F
ClientSampled :
SU-02-061617

Tot Ion:178 Resp: 1181737
Ion Ratio Lower Upper
178 100
176 19.6 16.2 24.4
179 15.5 12.6 19.0




Manual Integrations
APPROVED


mohammad
6/21/2017 2:16:48 PM



#71
Anthracene
Concen: 52.11 ng
RT: 14.67 min Scan# 2241
Delta R.T. 0.00 min
Lab File: BF096055.D
Acq: 20 Jun 2017 20:47

Tgt Ion:178 Resp: 500453
Ion Ratio Lower Upper
178 100
176 18.1 15.8 23.8
179 15.7 12.7 19.1





#72

Carbazole

Concen: 3.53 ng

RT: 14.97 min Scan# 2293

Delta R.T. 0.01 min

Lab File: BF096055.D

Acq: 20 Jun 2017 20:47

Instrument :

BNA_F





ClientSampleId :

SU-02-061617

Tot Ion	Ratio	Lower	Upper
167	100		
166	27.3	18.1	27.1#
139	15.6	11.7	17.5

Manual Integrations
APPROVED

mohammad
6/21/2017 2:16:48 PM



#74

Fluoranthene

Concen: 209.32 ng




RT: 16.40 min Scan# 2535


Delta R.T. 0.03 min

Lab File: BF096055.D

Acq: 20 Jun 2017 20:47

Tgt Ion	Ratio	Lower	Upper
202	100		
101	14.0	0.0	33.2
203	17.7	0.0	38.1





#75

Chrysene-d12

Concen: 20.00 ng

RT: 18.29 min Scan# 2857

Delta R.T. 0.02 min

Lab File: BF096055.D

Acq: 20 Jun 2017 20:47

Instrument :

BNA_F





ClientSampled :

SU-02-061617

Tot Ion	Ratio	Lower	Upper
240	100		
120	14.3	5.8	8.8#
236	24.7	20.5	30.7

Manual Integrations
APPROVED

mohammad
6/21/2017 2:16:48 PM



#77

Pyrene

Concen: 272.91 ng m




RT: 16.70 min Scan# 2587


Delta R.T. 0.04 min

Lab File: BF096055.D

Acq: 20 Jun 2017 20:47

Tgt Ion	Ratio	Lower	Upper
202	100		
200	22.0	17.6	26.4
203	21.7	14.4	21.6#









#78
 Terphenyl-d14
 Concen: 103.74 ng
 RT: 16.93 min Scan# 2626
 Delta R.T. 0.01 min
 Lab File: BF096055.D
 Acq: 20 Jun 2017 20:47

Instrument :
 BNA_F
 ClientSampleId :
 SU-02-061617

Tot Ion	Ratio	Lower	Upper
244	100		
212	7.3	6.0	9.0
122	14.0	10.3	15.5




Manual Integrations
 APPROVED


mohammad
 6/21/2017 2:16:48 PM



#80
 Benzo(a)anthracene
 Concen: 162.45 ng
 RT: 18.28 min Scan# 2855
 Delta R.T. 0.02 min
 Lab File: BF096055.D
 Acq: 20 Jun 2017 20:47

Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.4	22.6	33.8
229	23.0	16.3	24.5









#82
 Chrysene
 Concen: 151.52 ng
 RT: 18.33 min Scan# 2863
 Delta R.T. 0.02 min
 Lab File: BF096055.D
 Acq: 20 Jun 2017 20:47

Instrument :
 BNA_F
 ClientSampleId :
 SU-02-061617

Tot Ion:	228	Resp:	916011
Ion Ratio	Lower	Upper	
228	100		
226	30.9	24.9	37.3
229	23.0	15.8	23.6




Manual Integrations
 APPROVED


mohammad
 6/21/2017 2:16:48 PM



#85
 Indeno(1,2,3-cd)pyrene
 Concen: 68.92 ng
 RT: 21.12 min Scan# 3338
 Delta R.T. 0.02 min
 Lab File: BF096055.D
 Acq: 20 Jun 2017 20:47

Tgt Ion:	276	Resp:	356501
Ion Ratio	Lower	Upper	
276	100		
138	26.3	27.8	41.6#
277	24.8	20.2	30.4









#86
Pervlene-d12
Concen: 20.00 ng
RT: 19.96 min Scan# 3140
Delta R.T. 0.02 min
Lab File: BF096055.D
Acq: 20 Jun 2017 20:47

Instrument :
BNA_F
ClientSampled :
SU-02-061617

Tot Ion	Ratio	Lower	Upper
264	100		
260	22.7	19.5	29.3
265	26.3	17.2	25.8




Manual Integrations
APPROVED


mohammad
6/21/2017 2:16:48 PM



#87
Benzo(b)fluoranthene
Concen: 149.94 ng m
RT: 19.56 min Scan# 3073
Delta R.T. 0.02 min
Lab File: BF096055.D
Acq: 20 Jun 2017 20:47

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.2	18.1	27.1
125	15.3	9.8	14.8









#88
Benzo(k)fluoranthene
Concen: 57.22 ng m
RT: 19.59 min Scan# 3077
Delta R.T. 0.02 min
Lab File: BF096055.D
Acq: 20 Jun 2017 20:47

Instrument :
BNA_F
ClientSampleId :
SU-02-061617

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.4	18.4	27.6
125	13.1	13.1	19.7




Manual Integrations
APPROVED


mohammad
6/21/2017 2:16:48 PM



#89
Benzo(a)pyrene
Concen: 128.69 ng
RT: 19.91 min Scan# 3132
Delta R.T. 0.02 min
Lab File: BF096055.D
Acq: 20 Jun 2017 20:47

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.2	17.5	26.3
125	15.7	11.0	16.4





#90

Dibenzo(a,h)anthracene

Concen: 33.69 ng/mL

RT: 21.12 min Scan# 3337

Delta R.T. 0.02 min

Lab File: BF096055.D

Acq: 20 Jun 2017 20:47

Instrument :

BNA_F

ClientSampleId :

SU-02-061617

Tot Ion:278 Resp: 146277

Ion Ratio Lower Upper

278 100


139 27.6 16.6 24.8#

279 30.5 19.3 28.9#

Manual Integrations
APPROVED

mohammad




6/21/2017 2:16:48 PM



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#91

Benzo(g,h,i)perylene

Concen: 77.00 ng/mL

RT: 21.45 min Scan# 3394

Delta R.T. 0.03 min

Lab File: BF096055.D

Acq: 20 Jun 2017 20:47


Tgt Ion:276 Resp: 340818

Ion Ratio Lower Upper

276 100

277 25.0 18.6 28.0

138 31.1 25.4 38.0



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

