

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062018\
 Data File : BF106648.D
 Acq On : 21 Jun 2018 4:41
 Operator : JU/SJ
 Sample : J3580-01
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 S-1

Manual Integrations
 APPROVED

Sohil
 6/21/2018 11:28:17 AM

Quant Time: Jun 21 08:13:19 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 19 14:32:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.96	152	82799	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	311490	20.00	ng	0.00
38) Acenaphthene-d10	10.00	164	147211	20.00	ng	0.00
63) Phenanthrene-d10	11.49	188	241449	20.00	ng	-0.01
75) Chrysene-d12	14.15	240	230254	20.00	ng	0.00
86) Perylene-d12	15.69	264	171784	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	452419	92.52	ng	0.00
7) Phenol-d6	6.60	99	564748	92.49	ng	0.00
23) Nitrobenzene-d5	7.52	82	367787	65.62	ng	0.00
41) 2,4,6-Tribromophenol	10.79	330	148496	87.03	ng	0.00
44) 2-Fluorobiphenyl	9.32	172	647498	72.23	ng	0.00
78) Terphenyl-d14	13.08	244	615703	64.77	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.71	163	91328	8.272	ng	99
74) Fluoranthene	12.71	202	26039	2.029	ng	97
83) Bis(2-ethylhexyl)phthalate	14.10	149	16321	2.045	ng	# 91
87) Benzo(b)fluoranthene	15.23	252	26575m	2.490	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062018\
Data File : BF106648.D
Acq On : 21 Jun 2018 4:41
Operator : JU/SJ
Sample : J3580-01
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
S-1

Manual Integrations
APPROVED

Sohil
6/21/2018 11:28:17 AM

Quant Time: Jun 21 08:13:19 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 19 14:32:09 2018
Response via : Initial Calibration

