

Data Path : Z:\HPCHEM1\BNA F\DATA\BF062116\  
 Data File : BF088245.D  
 Acq On : 22 Jun 2016 7:51  
 Operator : UM/SJ  
 Sample : H3633-04  
 Misc :  
 ALS Vial : 29 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 P0-010

Quant Time: Jun 22 08:40:09 2016  
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060716.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Jun 22 08:34:05 2016  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.65  | 152  | 84560    | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 7.93  | 136  | 333253   | 20.00  | ng    | -0.01    |
| 38) Acenaphthene-d10        | 9.68  | 164  | 152713   | 20.00  | ng    | -0.01    |
| 63) Phenanthrene-d10        | 11.16 | 188  | 282900   | 20.00  | ng    | 0.00     |
| 75) Chrysene-d12            | 13.79 | 240  | 211437   | 20.00  | ng    | 0.00     |
| 86) Perylene-d12            | 15.17 | 264  | 162682   | 20.00  | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.23  | 112  | 292924   | 59.22  | ng    | 0.00     |
| 7) Phenol-d6                | 6.30  | 99   | 234507   | 39.12  | ng    | -0.01    |
| 23) Nitrobenzene-d5         | 7.22  | 82   | 452310   | 79.26  | ng    | 0.00     |
| 41) 2,4,6-Tribromophenol    | 10.48 | 330  | 178458   | 110.67 | ng    | 0.00     |
| 44) 2-Fluorobiphenyl        | 9.02  | 172  | 852543   | 87.89  | ng    | 0.00     |
| 78) Terphenyl-d14           | 12.74 | 244  | 735379   | 79.14  | ng    | -0.01    |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF062116\  
Data File : BF088245.D  
Acq On : 22 Jun 2016 7:51  
Operator : UM/SJ  
Sample : H3633-04  
Misc :  
ALS Vial : 29 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
P0-010

Quant Time: Jun 22 08:40:09 2016  
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060716.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Jun 22 08:34:05 2016  
Response via : Initial Calibration

