

Data Path : Z:\HPCHEM1\BNA F\DATA\BF062117\
 Data File : BF096090.D
 Acq On : 21 Jun 2017 18:15
 Operator : SJ/MA
 Sample : I3736-02
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 G24-1.5-3

Manual Integrations
 APPROVED

mohammad
 6/22/2017 5:34:42 PM

Quant Time: Jun 22 01:22:22 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 20 16:05:47 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.29	152	74016	20.00	ng	0.00
21) Naphthalene-d8	9.32	136	302039	20.00	ng	0.00
38) Acenaphthene-d10	12.14	164	130296	20.00	ng	-0.01
63) Phenanthrene-d10	14.53	188	205530	20.00	ng	0.00
75) Chrysene-d12	18.27	240	158576	20.00	ng	0.00
86) Perylene-d12	19.94	264	131961	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.19	112	368795	79.40	ng	0.01
7) Phenol-d6	6.87	99	473381	85.13	ng	0.00
23) Nitrobenzene-d5	8.21	82	275381	56.62	ng	-0.01
41) 2,4,6-Tribromophenol	13.44	330	81393	70.60	ng	-0.01
44) 2-Fluorobiphenyl	11.10	172	502052	53.62	ng	-0.01
78) Terphenyl-d14	16.92	244	306353	47.94	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	11.80	163	55318	5.59	ng	99
70) Phenanthrene	14.57	178	27071	2.28	ng	95
74) Fluoranthene	16.37	202	39427	3.36	ng	96
77) Pyrene	16.67	202	35416	2.99	ng	# 92
80) Benzo(a)anthracene	18.26	228	19906	2.16	ng	96
82) Chrysene	18.30	228	20897	2.27	ng	# 94
87) Benzo(b)fluoranthene	19.54	252	31213m	3.78	ng	# 80
89) Benzo(a)pyrene	19.90	252	15316	2.02	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF062117\
Data File : BF096090.D
Acq On : 21 Jun 2017 18:15
Operator : SJ/MA
Sample : I3736-02
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
G24-1.5-3

Manual Integrations
APPROVED

mohammad
6/22/2017 5:34:42 PM

Quant Time: Jun 22 01:22:22 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 20 16:05:47 2017
Response via : Initial Calibration

