

Data Path : Z:\HPCHEM1\BNA F\DATA\BF062117\
 Data File : BF096097.D
 Acq On : 21 Jun 2017 22:00
 Operator : SJ/MA
 Sample : I3736-03 5X
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 G22-0-1.5

Manual Integrations
 APPROVED

mohammad
 6/22/2017 6:02:34 PM

Quant Time: Jun 22 02:01:01 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 20 16:05:47 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.29	152	71828	20.00	ng	0.00
21) Naphthalene-d8	9.32	136	283933	20.00	ng	0.00
38) Acenaphthene-d10	12.14	164	121499	20.00	ng	0.00
63) Phenanthrene-d10	14.54	188	190708	20.00	ng	0.00
75) Chrysene-d12	18.27	240	140753	20.00	ng	0.00
86) Perylene-d12	19.94	264	110565	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.23	112	72884	16.17	ng	0.05
7) Phenol-d6	6.92	99	104761	19.41	ng	0.04
23) Nitrobenzene-d5	8.23	82	57726	12.63	ng	0.01
41) 2,4,6-Tribromophenol	13.46	330	11557	10.75	ng	0.01
44) 2-Fluorobiphenyl	11.10	172	136202	15.60	ng	0.00
78) Terphenyl-d14	16.92	244	87898	15.50	ng	0.00
Target Compounds						Qvalue
70) Phenanthrene	14.57	178	163106	14.82	ng	98
71) Anthracene	14.67	178	57492	5.00	ng	99
74) Fluoranthene	16.37	202	310341	28.50	ng	98
77) Pyrene	16.67	202	273108	26.00	ng	96
80) Benzo(a)anthracene	18.26	228	175441	21.44	ng	97
82) Chrysene	18.30	228	163438	19.98	ng	98
83) Bis(2-ethylhexyl)phthalate	18.34	149	16067	2.48	ng	# 91
85) Indeno(1,2,3-cd)pyrene	21.11	276	50274	7.18	ng	93
87) Benzo(b)fluoranthene	19.54	252	160636m	23.20	ng	
88) Benzo(k)fluoranthene	19.57	252	42415m	6.41	ng	
89) Benzo(a)pyrene	19.89	252	95872	15.07	ng	97
90) Dibenzo(a,h)anthracene	21.10	278	16134	2.99	ng	# 86
91) Benzo(a,h,i)perylene	21.44	276	42383	7.70	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF062117\
Data File : BF096097.D
Acq On : 21 Jun 2017 22:00
Operator : SJ/MA
Sample : I3736-03 5X
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
G22-0-1.5

Manual Integrations
APPROVED

mohammad
6/22/2017 6:02:34 PM

Quant Time: Jun 22 02:01:01 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 20 16:05:47 2017
Response via : Initial Calibration

