

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062121\
 Data File : BF124398.D
 Acq On : 21 Jun 2021 17:45
 Operator : JU/CG
 Sample : M2772-03
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-11

Quant Time: Jun 22 01:32:40 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 15 20:09:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.781	152	32428	20.000	ng	0.00
21) Naphthalene-d8	8.069	136	122979	20.000	ng	# 0.00
39) Acenaphthene-d10	9.822	164	71002	20.000	ng	0.00
64) Phenanthrene-d10	11.310	188	128233	20.000	ng	0.00
76) Chrysene-d12	13.957	240	73462	20.000	ng	0.00
86) Perylene-d12	15.415	264	77679	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.410	112	153014	71.031	ng	0.00
7) Phenol-d6	6.428	99	178170	61.532	ng	-0.02
23) Nitrobenzene-d5	7.351	82	142833	46.064	ng	0.00
42) 2,4,6-Tribromophenol	10.616	330	62545	62.176	ng	0.00
45) 2-Fluorobiphenyl	9.145	172	265707	47.091	ng	0.00
79) Terphenyl-d14	12.898	244	270500	50.555	ng	0.00
Target Compounds						
						Qvalue
75) Fluoranthene	12.527	202	20104	2.391	ng	94
78) Pyrene	12.757	202	19088	2.706	ng	93
87) Indeno(1,2,3-cd)pyrene	16.886	276	4462	4.963	ng	# 65
92) Benzo(g,h,i)perylene	17.321	276	3732	4.438	ng	# 74


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062121\
Data File : BF124398.D
Acq On : 21 Jun 2021 17:45
Operator : JU/CG
Sample : M2772-03
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-11

Quant Time: Jun 22 01:32:40 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 15 20:09:26 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.781 min Scan# 798
Delta R.T. -0.005 min
Lab File: BF124398.D
Acq: 21 Jun 2021 17:45




Instrument :
BNA_F
ClientSampled :
SB-11


Tgt Ion:152 Resp: 32428
Ion Ratio Lower Upper
152 100
150 160.7 127.0 190.6
115 66.6 54.3 81.5



#5
2-Fluorophenol
Concen: 71.031 ng
RT: 5.410 min Scan# 565
Delta R.T. -0.006 min
Lab File: BF124398.D
Acq: 21 Jun 2021 17:45

Tgt Ion:112 Resp: 153014
Ion Ratio Lower Upper
112 100
64 61.8 61.4 92.0
63 36.2 33.0 49.4









#7
Phenol-d6
Concen: 61.532 ng
RT: 6.428 min Scan# 738
Delta R.T. -0.018 min
Lab File: BF124398.D
Acq: 21 Jun 2021 17:45




Instrument :
BNA_F
ClientSampled :
SB-11


Tgt Ion: 99 Resp: 178170
Ion Ratio Lower Upper
99 100
42 20.6 19.8 29.8
71 47.6 31.1 46.7#



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.069 min Scan# 1017
Delta R.T. -0.006 min
Lab File: BF124398.D
Acq: 21 Jun 2021 17:45

Tgt Ion: 136 Resp: 122979
Ion Ratio Lower Upper
136 100
137 11.1 8.7 13.1
54 7.8 9.8 14.6#
68 4.9 4.1 6.1









#23
Nitrobenzene-d5
Concen: 46.064 ng
RT: 7.351 min Scan# 895
Delta R.T. -0.006 min
Lab File: BF124398.D
Acq: 21 Jun 2021 17:45




Instrument :
BNA_F
ClientSampled :
SB-11


Tgt Ion: 82 Resp: 142833
Ion Ratio Lower Upper
82 100
128 38.7 29.4 44.2
54 53.2 54.1 81.1#



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.822 min Scan# 1315
Delta R.T. -0.006 min
Lab File: BF124398.D
Acq: 21 Jun 2021 17:45

Tgt Ion:164 Resp: 71002
Ion Ratio Lower Upper
164 100
162 103.0 80.7 121.1
160 44.1 37.3 55.9









#42
2,4,6-Tribromophenol
Concen: 62.176 ng
RT: 10.616 min Scan# 1450
Delta R.T. -0.006 min
Lab File: BF124398.D
Acq: 21 Jun 2021 17:45




Instrument :
BNA_F
ClientSampled :
SB-11


Tgt Ion: 330 Resp: 62545
Ion Ratio Lower Upper
330 100
332 96.3 77.1 115.7
141 37.9 27.9 41.9



#45
2-Fluorobiphenyl
Concen: 47.091 ng
RT: 9.145 min Scan# 1200
Delta R.T. -0.006 min
Lab File: BF124398.D
Acq: 21 Jun 2021 17:45

Tgt Ion: 172 Resp: 265707
Ion Ratio Lower Upper
172 100
171 37.0 28.1 42.1
170 23.6 19.2 28.8









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.310 min Scan# 1568
Delta R.T. -0.006 min
Lab File: BF124398.D
Acq: 21 Jun 2021 17:45




Instrument :
BNA_F
ClientSampled :
SB-11


Tgt Ion:188 Resp: 128233
Ion Ratio Lower Upper
188 100
94 7.9 6.9 10.3
80 9.7 8.2 12.2



#75
Fluoranthene
Concen: 2.391 ng
RT: 12.527 min Scan# 1775
Delta R.T. -0.006 min
Lab File: BF124398.D
Acq: 21 Jun 2021 17:45

Tgt Ion:202 Resp: 20104
Ion Ratio Lower Upper
202 100
101 6.5 0.0 30.9
203 16.2 0.0 37.5









#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.957 min Scan# 2018
Delta R.T. -0.006 min
Lab File: BF124398.D
Acq: 21 Jun 2021 17:45




Instrument :
BNA_F
ClientSampled :
SB-11


Tgt Ion:	240	Resp:	73462
Ion Ratio	Lower	Upper	
240	100		
120	8.8	7.2	10.8
236	23.7	19.0	28.4



#78
Pyrene
Concen: 2.706 ng
RT: 12.757 min Scan# 1814
Delta R.T. -0.006 min
Lab File: BF124398.D
Acq: 21 Jun 2021 17:45

Tgt Ion:	202	Resp:	19088
Ion Ratio	Lower	Upper	
202	100		
200	24.0	16.1	24.1
203	16.4	14.8	22.2









#79
Terphenyl-d14
Concen: 50.555 ng
RT: 12.898 min Scan# 1838
Delta R.T. -0.006 min
Lab File: BF124398.D
Acq: 21 Jun 2021 17:45




Instrument :
BNA_F
ClientSampled :
SB-11


Tgt Ion:244 Resp: 270500
Ion Ratio Lower Upper
244 100
212 7.2 5.7 8.5
122 8.9 9.7 14.5#



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.415 min Scan# 2266
Delta R.T. 0.000 min
Lab File: BF124398.D
Acq: 21 Jun 2021 17:45

Tgt Ion:264 Resp: 77679
Ion Ratio Lower Upper
264 100
260 20.5 18.7 28.1
265 22.9 17.8 26.8









#87
Indeno(1,2,3-cd)pyrene
Concen: 4.963 ng
RT: 16.886 min Scan# 2516
Delta R.T. 0.012 min
Lab File: BF124398.D
Acq: 21 Jun 2021 17:45

Instrument :
BNA_F
ClientSampled :
SB-11

Tgt Ion:276 Resp: 4462
Ion Ratio Lower Upper
276 100
138 0.0 16.2 24.2#
277 11.2 20.1 30.1#



#92
Benzo(g,h,i)perylene
Concen: 4.438 ng
RT: 17.321 min Scan# 2590
Delta R.T. 0.006 min
Lab File: BF124398.D
Acq: 21 Jun 2021 17:45

Tgt Ion:276 Resp: 3732
Ion Ratio Lower Upper
276 100
277 18.2 18.2 27.2#
138 38.5 13.8 20.6#

