


Data Path : Z:\HPCHEM1\BNA_F\Data\BF062215\
 Data File : BF080105.D
 Acq On : 22 Jun 2015 22:09
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF061715.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jun 23 00:53:56 2015



AutoFind: Scans 789, 790, 791; Background Corrected with Scan 787

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	45.0	8196	PASS
68	69	0.00	2	1.8	125	PASS
69	198	0.00	100	38.9	7091	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	47.5	8660	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	18216	PASS
199	198	5	9	6.7	1226	PASS
275	198	10	60	25.6	4666	PASS
365	198	1	100	2.8	515	PASS
441	442	0.01	24	16.2	2702	PASS
442	198	50	100	91.5	16659	PASS
443	442	15	24	20.3	3384	PASS