

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062222\
 Data File : BF128933.D
 Acq On : 22 Jun 2022 15:54
 Operator : CG\JU
 Sample : PB145718TB
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB145718TB

Quant Time: Jun 23 01:29:40 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 22 01:02:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.745	152	114393	20.000	ng	-0.01
21) Naphthalene-d8	8.028	136	412393	20.000	ng	-0.01
39) Acenaphthene-d10	9.786	164	251854	20.000	ng	0.00
64) Phenanthrene-d10	11.274	188	474114	20.000	ng	# 0.00
76) Chrysene-d12	13.915	240	370268	20.000	ng	#-0.01
86) Perylene-d12	15.362	264	288024	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.393	112	890407	123.982	ng	0.00
7) Phenol-d6	6.410	99	1140052	130.046	ng	0.00
23) Nitrobenzene-d5	7.316	82	816927	92.230	ng	-0.01
42) 2,4,6-Tribromophenol	10.580	330	349678	117.508	ng	0.00
45) 2-Fluorobiphenyl	9.110	172	1487819	82.752	ng	0.00
79) Terphenyl-d14	12.868	244	1822706	85.538	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062222\
Data File : BF128933.D
Acq On : 22 Jun 2022 15:54
Operator : CG\JU
Sample : PB145718TB
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB145718TB

Quant Time: Jun 23 01:29:40 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 22 01:02:29 2022
Response via : Initial Calibration

