

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062222\
 Data File : BF128925.D
 Acq On : 22 Jun 2022 11:21
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 23 01:24:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 22 01:02:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.466	0.469	-0.6	95	0.00
3	Pyridine	1.238	1.282	-3.6	99	0.00
4	n-Nitrosodimethylamine	0.859	0.817	4.9	90	-0.02
5 S	2-Fluorophenol	1.256	1.261	-0.4	99	0.00
6	Aniline	1.653	1.746	-5.6	106	0.00
7 S	Phenol-d6	1.533	1.598	-4.2	104	0.00
8	2-Chlorophenol	1.285	1.337	-4.0	104	0.00
9	Benzaldehyde	0.934	1.005	-7.6	133	0.00
10 C	Phenol	1.620	1.694	-4.6	103	0.00
11	bis(2-Chloroethyl)ether	1.187	1.240	-4.5	104	0.00
12	1,3-Dichlorobenzene	1.485	1.551	-4.4	104	0.00
13 C	1,4-Dichlorobenzene	1.500	1.545	-3.0	103	0.00
14	1,2-Dichlorobenzene	1.403	1.468	-4.6	106	0.00
15	Benzyl Alcohol	1.282	1.322	-3.1	104	0.00
16	2,2'-oxybis(1-Chloropropane	1.956	2.027	-3.6	107	-0.01
17	2-Methylphenol	1.015	1.424	-40.3#	143	0.00
18	Hexachloroethane	0.558	0.585	-4.8	103	0.00
19 P	n-Nitroso-di-n-propylamine	0.953	0.999	-4.8	104	0.00
20	3+4-Methylphenols	1.356	1.424	-5.0	104	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
22	Acetophenone	0.523	0.544	-4.0	103	0.00
23 S	Nitrobenzene-d5	0.430	0.454	-5.6	106	0.00
24	Nitrobenzene	0.394	0.425	-7.9	107	0.00
25	Isophorone	0.652	0.689	-5.7	106	0.00
26 C	2-Nitrophenol	0.174	0.195	-12.1	108	0.00
27	2,4-Dimethylphenol	0.265	0.289	-9.1	105	0.00
28	bis(2-Chloroethoxy)methane	0.412	0.435	-5.6	104	0.00
29 C	2,4-Dichlorophenol	0.306	0.324	-5.9	103	0.00
30	1,2,4-Trichlorobenzene	0.345	0.360	-4.3	103	0.00
31	Naphthalene	1.026	1.084	-5.7	109	0.00
32	Benzoic acid	0.188	0.125	33.5#	64	0.00
33	4-Chloroaniline	0.387	0.418	-8.0	112	0.00
34 C	Hexachlorobutadiene	0.245	0.259	-5.7	110	0.00
35	Caprolactam	0.085	0.090	-5.9	110	0.00
36 C	4-Chloro-3-methylphenol	0.334	0.357	-6.9	110	0.00
37	2-Methylnaphthalene	0.653	0.691	-5.8	110	0.00
38	1-Methylnaphthalene	0.631	0.672	-6.5	113	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	106	0.00
40	1,2,4,5-Tetrachlorobenzene	0.621	0.663	-6.8	109	0.00
41 P	Hexachlorocyclopentadiene	0.373	0.347	7.0	92	0.00
42 S	2,4,6-Tribromophenol	0.236	0.217	8.1	91	0.00
43 C	2,4,6-Trichlorophenol	0.420	0.387	7.9	95	0.00
44	2,4,5-Trichlorophenol	0.434	0.395	9.0	93	0.00
45 S	2-Fluorobiphenyl	1.428	1.474	-3.2	107	0.00
46	1,1'-Biphenyl	1.501	1.561	-4.0	108	0.00
47	2-Chloronaphthalene	1.145	1.194	-4.3	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062222\
 Data File : BF128925.D
 Acq On : 22 Jun 2022 11:21
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 23 01:24:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 22 01:02:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.392	0.431	-9.9	113	0.00
49	Acenaphthylene	1.753	1.807	-3.1	107	0.00
50	Dimethylphthalate	1.369	1.432	-4.6	108	0.00
51	2,6-Dinitrotoluene	0.270	0.298	-10.4	110	0.00
52 C	Acenaphthene	1.142	1.091	4.5	97	0.00
53	3-Nitroaniline	0.293	0.324	-10.6	111	0.00
54 P	2,4-Dinitrophenol	0.150	0.143	4.7	94	0.00
55	Dibenzofuran	1.641	1.672	-1.9	106	0.00
56 P	4-Nitrophenol	0.213	0.188	11.7	93	0.00
57	2,4-Dinitrotoluene	0.361	0.382	-5.8	105	0.00
58	Fluorene	1.272	1.323	-4.0	105	0.00
59	2,3,4,6-Tetrachlorophenol	0.359	0.312	13.1	88	0.00
60	Diethylphthalate	1.376	1.415	-2.8	106	0.00
61	4-Chlorophenyl-phenylether	0.660	0.684	-3.6	106	0.00
62	4-Nitroaniline	0.297	0.298	-0.3	101	0.00
63	Azobenzene	1.355	1.405	-3.7	106	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	0.118	0.128	-8.5	107	0.00
66 c	n-Nitrosodiphenylamine	0.615	0.660	-7.3	107	0.00
67	4-Bromophenyl-phenylether	0.223	0.239	-7.2	108	0.00
68	Hexachlorobenzene	0.248	0.259	-4.4	104	0.00
69	Atrazine	0.198	0.203	-2.5	102	0.00
70 C	Pentachlorophenol	0.124	0.100	19.4	79	0.00
71	Phenanthrene	1.003	1.049	-4.6	107	0.00
72	Anthracene	0.993	1.032	-3.9	106	0.00
73	Carbazole	0.922	0.969	-5.1	107	0.00
74	Di-n-butylphthalate	1.123	1.182	-5.3	105	0.00
75 C	Fluoranthene	1.049	1.118	-6.6	106	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
77	Benzydine	0.399	0.411	-3.0	106	0.00
78	Pyrene	1.488	1.590	-6.9	106	0.00
79 S	Terphenyl-d14	1.151	1.209	-5.0	99	0.00
80	Butylbenzylphthalate	0.625	0.699	-11.8	103	0.00
81	Benzo(a)anthracene	1.247	1.351	-8.3	102	0.00
82	3,3'-Dichlorobenzidine	0.267	0.299	-12.0	107	0.00
83	Chrysene	1.189	1.277	-7.4	102	0.00
84	Bis(2-ethylhexyl)phthalate	0.812	0.937	-15.4	107	0.00
85 c	Di-n-octyl phthalate	1.224	1.410	-15.2	106	0.00
86 I	Perylene-d12	1.000	1.000	0.0	97	0.00
87	Indeno(1,2,3-cd)pyrene	1.205	1.201	0.3	106	0.00
88	Benzo(b)fluoranthene	1.278	1.429	-11.8	104	0.00
89	Benzo(k)fluoranthene	1.202	1.217	-1.2	95	0.00
90 C	Benzo(a)pyrene	1.004	1.047	-4.3	99	0.00
91	Dibenzo(a,h)anthracene	1.000	1.024	-2.4	106	0.00
92	Benzo(g,h,i)perylene	0.991	0.991	0.0	107	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062222\
Data File : BF128925.D
Acq On : 22 Jun 2022 11:21
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 23 01:24:38 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 22 01:02:29 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0