

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062222\
 Data File : BF128925.D
 Acq On : 22 Jun 2022 11:21
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 23 01:24:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 22 01:02:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	40.208	-0.5	95	0.00
3	Pyridine	40.000	41.406	-3.5	99	0.00
4	n-Nitrosodimethylamine	40.000	38.044	4.9	90	-0.02
5 S	2-Fluorophenol	80.000	80.340	-0.4	99	0.00
6	Aniline	40.000	42.258	-5.6	106	0.00
7 S	Phenol-d6	80.000	83.421	-4.3	104	0.00
8	2-Chlorophenol	40.000	41.626	-4.1	104	0.00
9	Benzaldehyde	40.000	43.063	-7.7	133	0.00
10 C	Phenol	40.000	41.824	-4.6	103	0.00
11	bis(2-Chloroethyl)ether	40.000	41.790	-4.5	104	0.00
12	1,3-Dichlorobenzene	40.000	41.772	-4.4	104	0.00
13 C	1,4-Dichlorobenzene	40.000	41.211	-3.0	103	0.00
14	1,2-Dichlorobenzene	40.000	41.857	-4.6	106	0.00
15	Benzyl Alcohol	40.000	41.233	-3.1	104	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.458	-3.6	107	-0.01
17	2-Methylphenol	40.000	56.092	-40.2#	143	0.00
18	Hexachloroethane	40.000	41.937	-4.8	103	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.924	-4.8	104	0.00
20	3+4-Methylphenols	40.000	42.015	-5.0	104	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	103	0.00
22	Acetophenone	40.000	41.590	-4.0	103	0.00
23 S	Nitrobenzene-d5	80.000	84.548	-5.7	106	0.00
24	Nitrobenzene	40.000	43.080	-7.7	107	0.00
25	Isophorone	40.000	42.252	-5.6	106	0.00
26 C	2-Nitrophenol	40.000	44.830	-12.1	108	0.00
27	2,4-Dimethylphenol	40.000	43.722	-9.3	105	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.237	-5.6	104	0.00
29 C	2,4-Dichlorophenol	40.000	42.323	-5.8	103	0.00
30	1,2,4-Trichlorobenzene	40.000	41.747	-4.4	103	0.00
31	Naphthalene	40.000	42.241	-5.6	109	0.00
32	Benzoic acid	40.000	26.549	33.6#	64	0.00
33	4-Chloroaniline	40.000	43.167	-7.9	112	0.00
34 C	Hexachlorobutadiene	40.000	42.272	-5.7	110	0.00
35	Caprolactam	40.000	42.498	-6.2	110	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.694	-6.7	110	0.00
37	2-Methylnaphthalene	40.000	42.367	-5.9	110	0.00
38	1-Methylnaphthalene	40.000	42.646	-6.6	113	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	106	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.719	-6.8	109	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.223	6.9	92	0.00
42 S	2,4,6-Tribromophenol	80.000	73.526	8.1	91	0.00
43 C	2,4,6-Trichlorophenol	40.000	36.892	7.8	95	0.00
44	2,4,5-Trichlorophenol	40.000	36.352	9.1	93	0.00
45 S	2-Fluorobiphenyl	80.000	82.566	-3.2	107	0.00
46	1,1'-Biphenyl	40.000	41.596	-4.0	108	0.00
47	2-Chloronaphthalene	40.000	41.724	-4.3	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062222\
 Data File : BF128925.D
 Acq On : 22 Jun 2022 11:21
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 23 01:24:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 22 01:02:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.073	-10.2	113	0.00
49	Acenaphthylene	40.000	41.225	-3.1	107	0.00
50	Dimethylphthalate	40.000	41.833	-4.6	108	0.00
51	2,6-Dinitrotoluene	40.000	44.123	-10.3	110	0.00
52 C	Acenaphthene	40.000	38.243	4.4	97	0.00
53	3-Nitroaniline	40.000	44.194	-10.5	111	0.00
54 P	2,4-Dinitrophenol	40.000	38.123	4.7	94	0.00
55	Dibenzofuran	40.000	40.770	-1.9	106	0.00
56 P	4-Nitrophenol	40.000	35.465	11.3	93	0.00
57	2,4-Dinitrotoluene	40.000	42.362	-5.9	105	0.00
58	Fluorene	40.000	41.600	-4.0	105	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	34.808	13.0	88	0.00
60	Diethylphthalate	40.000	41.147	-2.9	106	0.00
61	4-Chlorophenyl-phenylether	40.000	41.428	-3.6	106	0.00
62	4-Nitroaniline	40.000	40.137	-0.3	101	0.00
63	Azobenzene	40.000	41.506	-3.8	106	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.446	-8.6	107	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.924	-7.3	107	0.00
67	4-Bromophenyl-phenylether	40.000	42.906	-7.3	108	0.00
68	Hexachlorobenzene	40.000	41.832	-4.6	104	0.00
69	Atrazine	40.000	41.100	-2.8	102	0.00
70 C	Pentachlorophenol	40.000	32.202	19.5	79	0.00
71	Phenanthrene	40.000	41.809	-4.5	107	0.00
72	Anthracene	40.000	41.590	-4.0	106	0.00
73	Carbazole	40.000	42.031	-5.1	107	0.00
74	Di-n-butylphthalate	40.000	42.120	-5.3	105	0.00
75 C	Fluoranthene	40.000	42.653	-6.6	106	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	95	0.00
77	Benzydine	40.000	41.204	-3.0	106	0.00
78	Pyrene	40.000	42.734	-6.8	106	0.00
79 S	Terphenyl-d14	80.000	84.041	-5.1	99	0.00
80	Butylbenzylphthalate	40.000	44.771	-11.9	103	0.00
81	Benzo(a)anthracene	40.000	43.343	-8.4	102	0.00
82	3,3'-Dichlorobenzidine	40.000	44.891	-12.2	107	0.00
83	Chrysene	40.000	42.970	-7.4	102	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	46.119	-15.3	107	0.00
85 c	Di-n-octyl phthalate	40.000	46.096	-15.2	106	0.00
86 I	Perylene-d12	20.000	20.000	0.0	97	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.840	0.4	106	0.00
88	Benzo(b)fluoranthene	40.000	44.718	-11.8	104	0.00
89	Benzo(k)fluoranthene	40.000	40.498	-1.2	95	0.00
90 C	Benzo(a)pyrene	40.000	41.680	-4.2	99	0.00
91	Dibenzo(a,h)anthracene	40.000	40.969	-2.4	106	0.00
92	Benzo(g,h,i)perylene	40.000	39.990	0.0	107	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062222\
Data File : BF128925.D
Acq On : 22 Jun 2022 11:21
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 23 01:24:38 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 22 01:02:29 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0