

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062222\
 Data File : BF128944.D
 Acq On : 22 Jun 2022 21:17
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 23 01:33:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 22 01:02:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	-0.01
2	1,4-Dioxane	0.466	0.479	-2.8	106	-0.03
3	Pyridine	1.238	1.285	-3.8	108	-0.02
4	n-Nitrosodimethylamine	0.859	0.833	3.0	101	-0.04
5 S	2-Fluorophenol	1.256	1.266	-0.8	108	0.00
6	Aniline	1.653	1.699	-2.8	113	0.00
7 S	Phenol-d6	1.533	1.582	-3.2	112	0.00
8	2-Chlorophenol	1.285	1.343	-4.5	114	0.00
9	Benzaldehyde	0.934	1.035	-10.8	149	-0.01
10 C	Phenol	1.620	1.677	-3.5	111	0.00
11	bis(2-Chloroethyl)ether	1.187	1.272	-7.2	116	0.00
12	1,3-Dichlorobenzene	1.485	1.548	-4.2	113	-0.01
13 C	1,4-Dichlorobenzene	1.500	1.569	-4.6	115	-0.01
14	1,2-Dichlorobenzene	1.403	1.478	-5.3	117	0.00
15	Benzyl Alcohol	1.282	1.288	-0.5	110	0.00
16	2,2'-oxybis(1-Chloropropane	1.956	2.103	-7.5	121	-0.01
17	2-Methylphenol	1.015	1.463	-44.1#	160#	0.00
18	Hexachloroethane	0.558	0.594	-6.5	114	0.00
19 P	n-Nitroso-di-n-propylamine	0.953	1.036	-8.7	118	0.00
20	3+4-Methylphenols	1.356	1.463	-7.9	117	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	115	0.00
22	Acetophenone	0.523	0.544	-4.0	115	0.00
23 S	Nitrobenzene-d5	0.430	0.453	-5.3	118	0.00
24	Nitrobenzene	0.394	0.418	-6.1	118	0.00
25	Isophorone	0.652	0.694	-6.4	119	0.00
26 C	2-Nitrophenol	0.174	0.199	-14.4	123	0.00
27	2,4-Dimethylphenol	0.265	0.286	-7.9	116	0.00
28	bis(2-Chloroethoxy)methane	0.412	0.439	-6.6	117	0.00
29 C	2,4-Dichlorophenol	0.306	0.327	-6.9	116	0.00
30	1,2,4-Trichlorobenzene	0.345	0.358	-3.8	114	0.00
31	Naphthalene	1.026	1.063	-3.6	119	0.00
32	Benzoic acid	0.188	0.117	37.8#	67	0.00
33	4-Chloroaniline	0.387	0.361	6.7	108	0.00
34 C	Hexachlorobutadiene	0.245	0.250	-2.0	119	0.00
35	Caprolactam	0.085	0.089	-4.7	121	0.00
36 C	4-Chloro-3-methylphenol	0.334	0.350	-4.8	121	0.00
37	2-Methylnaphthalene	0.653	0.683	-4.6	121	0.00
38	1-Methylnaphthalene	0.631	0.662	-4.9	124	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	117	0.00
40	1,2,4,5-Tetrachlorobenzene	0.621	0.657	-5.8	119	0.00
41 P	Hexachlorocyclopentadiene	0.373	0.351	5.9	102	0.00
42 S	2,4,6-Tribromophenol	0.236	0.219	7.2	101	0.00
43 C	2,4,6-Trichlorophenol	0.420	0.396	5.7	107	0.00
44	2,4,5-Trichlorophenol	0.434	0.409	5.8	106	0.00
45 S	2-Fluorobiphenyl	1.428	1.467	-2.7	117	0.00
46	1,1'-Biphenyl	1.501	1.558	-3.8	119	0.00
47	2-Chloronaphthalene	1.145	1.209	-5.6	122	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062222\
 Data File : BF128944.D
 Acq On : 22 Jun 2022 21:17
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 23 01:33:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 22 01:02:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.392	0.434	-10.7	125	0.00
49	Acenaphthylene	1.753	1.780	-1.5	116	0.00
50	Dimethylphthalate	1.369	1.452	-6.1	120	0.00
51	2,6-Dinitrotoluene	0.270	0.299	-10.7	122	0.00
52 C	Acenaphthene	1.142	1.095	4.1	107	0.00
53	3-Nitroaniline	0.293	0.318	-8.5	120	0.00
54 P	2,4-Dinitrophenol	0.150	0.135	10.0	97	0.00
55	Dibenzofuran	1.641	1.660	-1.2	116	-0.01
56 P	4-Nitrophenol	0.213	0.182	14.6	99	0.00
57	2,4-Dinitrotoluene	0.361	0.385	-6.6	116	0.00
58	Fluorene	1.272	1.323	-4.0	116	0.00
59	2,3,4,6-Tetrachlorophenol	0.359	0.324	9.7	100	0.00
60	Diethylphthalate	1.376	1.437	-4.4	118	0.00
61	4-Chlorophenyl-phenylether	0.660	0.682	-3.3	116	0.00
62	4-Nitroaniline	0.297	0.294	1.0	110	0.00
63	Azobenzene	1.355	1.442	-6.4	120	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
65	4,6-Dinitro-2-methylphenol	0.118	0.125	-5.9	112	0.00
66 c	n-Nitrosodiphenylamine	0.615	0.659	-7.2	116	0.00
67	4-Bromophenyl-phenylether	0.223	0.242	-8.5	117	-0.01
68	Hexachlorobenzene	0.248	0.266	-7.3	115	0.00
69	Atrazine	0.198	0.183	7.6	99	0.00
70 C	Pentachlorophenol	0.124	0.104	16.1	89	0.00
71	Phenanthrene	1.003	1.039	-3.6	114	-0.01
72	Anthracene	0.993	1.025	-3.2	114	0.00
73	Carbazole	0.922	0.955	-3.6	114	0.00
74	Di-n-butylphthalate	1.123	1.187	-5.7	114	0.00
75 C	Fluoranthene	1.049	1.064	-1.4	108	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
77	Benzidine	0.399	0.340	14.8	83	0.00
78	Pyrene	1.488	1.680	-12.9	106	0.00
79 S	Terphenyl-d14	1.151	1.262	-9.6	98	0.00
80	Butylbenzylphthalate	0.625	0.713	-14.1	100	0.00
81	Benzo(a)anthracene	1.247	1.331	-6.7	96	0.00
82	3,3'-Dichlorobenzidine	0.267	0.320	-19.9	109	0.00
83	Chrysene	1.189	1.254	-5.5	96	0.00
84	Bis(2-ethylhexyl)phthalate	0.812	0.923	-13.7	101	0.00
85 c	Di-n-octyl phthalate	1.224	1.342	-9.6	96	0.00
86 I	Perylene-d12	1.000	1.000	0.0	98	0.00
87	Indeno(1,2,3-cd)pyrene	1.205	1.333	-10.6	120	0.00
88	Benzo(b)fluoranthene	1.278	1.237	3.2	92	0.00
89	Benzo(k)fluoranthene	1.202	1.305	-8.6	103	0.00
90 C	Benzo(a)pyrene	1.004	1.043	-3.9	100	0.00
91	Dibenzo(a,h)anthracene	1.000	1.114	-11.4	117	0.00
92	Benzo(g,h,i)perylene	0.991	1.107	-11.7	121	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062222\
Data File : BF128944.D
Acq On : 22 Jun 2022 21:17
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 23 01:33:27 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 22 01:02:29 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0