

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF062415\
 Data File : BF080148.D
 Acq On : 25 Jun 2015 1:10
 Operator : TP/IZ
 Sample : G2717-03
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 AC-UNDERCUT(0-2)-6

Quant Time: Jun 25 04:52:47 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF061715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 25 00:52:28 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.32	152	39714	20.00	ng	0.00
21) Naphthalene-d8	8.61	136	151073	20.00	ng	0.00
38) Acenaphthene-d10	10.38	164	76634	20.00	ng	0.00
63) Phenanthrene-d10	11.88	188	154155	20.00	ng	-0.01
75) Chrysene-d12	14.83	240	163315	20.00	ng	-0.10
86) Perylene-d12	16.70	264	163674	20.00	ng	-0.14
System Monitoring Compounds						
5) 2-Fluorophenol	5.92	112	271437	120.99	ng	0.00
7) Phenol-d6	6.92	99	381907	127.92	ng	0.00
23) Nitrobenzene-d5	7.86	82	197299	76.03	ng	-0.01
41) 2,4,6-Tribromophenol	11.17	330	88134	117.61	ng	0.00
44) 2-Fluorobiphenyl	9.68	172	448479	83.46	ng	0.00
78) Terphenyl-d14	13.57	244	498426	71.28	ng	-0.05
Target Compounds						Qvalue
49) Dimethylphthalate	10.07	163	42352	7.35	ng	# 98
70) Phenanthrene	11.90	178	27435	2.94	ng	97
73) Di-n-butylphthalate	12.44	149	69540	7.02	ng	# 98
74) Fluoranthene	13.16	202	43965	4.42	ng	94
77) Pyrene	13.42	202	39328	3.91	ng	96
80) Benzo(a)anthracene	14.81	228	23068	2.32	ng	92
82) Chrysene	14.86	228	24343	2.65	ng	95
85) Indeno(1,2,3-cd)pyrene	18.53	276	45363	3.92	ng	# 91
87) Benzo(b)fluoranthene	16.16	252	35142m	3.55	ng	
89) Benzo(a)pyrene	16.62	252	26441m	2.81	ng	
90) Dibenzo(a,h)anthracene	18.55	278	32703	3.40	ng	# 86
91) Benzo(g,h,i)perylene	19.09	276	39608	4.07	ng	# 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF062415\
Data File : BF080148.D
Acq On : 25 Jun 2015 1:10
Operator : TP/IZ
Sample : G2717-03
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
AC-UNDERCUT(0-2)-6

Quant Time: Jun 25 04:52:47 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF061715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 25 00:52:28 2015
Response via : Initial Calibration

