

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF062416\
 Data File : BF088365.D
 Acq On : 25 Jun 2016 9:17
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 99 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Quant Time: Jun 27 08:43:48 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF060716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 04:27:12 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.72	152	557826	20.00	ng	-0.13
21) Naphthalene-d8	7.74	136	1203	20.00	ng	-0.39
38) Acenaphthene-d10	9.40	164	2052	20.00	ng	-0.48
63) Phenanthrene-d10	10.80	188	1241	20.00	ng	-0.56
75) Chrysene-d12	0.00	240	0	0.00	ng	-13.99
86) Perylene-d12	0.00	264	0	0.00	ng	-15.39

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	0.00	112	0	0.00	ng	
7) Phenol-d6	6.35	99	12867	0.33	ng	-0.13
23) Nitrobenzene-d5	7.15	82	969963	47082.42	ng	-0.26
41) 2,4,6-Tribromophenol	0.00	330	0	0.00	ng	
44) 2-Fluorobiphenyl	8.94	172	1513057	12384.60	ng	-0.27
78) Terphenyl-d14	0.00	244	0	0.00	ng	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
6) Aniline	6.23	93	383063	6.81	ng	98
8) 2-Chlorophenol	6.35	128	288165	7.46	ng	87
9) Benzaldehyde	6.10	77	174317	5.50	ng	88
10) Phenol	6.24	94	368075	6.91	ng	75
11) bis(2-Chloroethyl)ether	6.31	93	290927	8.39	ng	86
12) 1,3-Dichlorobenzene	6.73	146	237306	5.73	ng	# 90
13) 1,4-Dichlorobenzene	6.58	146	316033	7.57	ng	98
14) 1,2-Dichlorobenzene	6.73	146	247041	6.51	ng	# 88
15) Benzyl Alcohol	6.73	79	203547	6.58	ng	97
16) 2,2'-oxybis(1-Chloropropan	6.86	45	262490	5.60	ng	# 1
17) 2-Methylphenol	7.02	107	276474	8.83	ng	# 77
18) Hexachloroethane	7.07	117	110323	7.45	ng	93
19) n-Nitroso-di-n-propylamine	7.00	70	176819	6.99	ng	89
20) 3+4-Methylphenols	7.02	107	272512	7.46	ng	# 80
22) Acetophenone	6.99	105	375902	13982.36	ng	# 97
24) Nitrobenzene	7.16	77	249317	12493.50	ng	87
25) Isophorone	7.40	82	525730	13777.13	ng	99
26) 2-Nitrophenol	7.47	139	158017	14781.94	ng	96
27) 2,4-Dimethylphenol	7.54	122	239952	12191.37	ng	99
28) bis(2-Chloroethoxy)methane	7.47	93	15380	649.83	ng	# 67
29) 2,4-Dichlorophenol	7.74	162	228631	13912.65	ng	92
30) 1,2,4-Trichlorobenzene	7.79	180	234877	13126.03	ng	99
31) Naphthalene	7.87	128	798983	13420.94	ng	99
32) Benzoic acid	7.70	122	71962	6639.84	ng	91
33) 4-Chloroaniline	7.87	127	105368	3963.62	ng	# 34
34) Hexachlorobutadiene	7.99	225	121843	12911.28	ng	98
35) Caprolactam	8.34	113	65359	12007.23	ng	# 56
36) 4-Chloro-3-methylphenol	8.44	107	246990	14053.91	ng	94
37) 2-Methylnaphthalene	8.44	142	203028	5174.19	ng	# 24
39) 1,2,4,5-Tetrachlorobenzene	8.73	216	216698	Below Cal		# 1
40) Hexachlorocyclopentadiene	8.72	237	43664	1319.17	ng	98
42) 2,4,6-Trichlorophenol	8.86	196	148427	3617.03	ng	98
43) 2,4,5-Trichlorophenol	8.86	196	148427	3476.64	ng	# 86
45) 1,1'-Biphenyl	9.03	154	583677	4388.48	ng	98

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF062416\
 Data File : BF088365.D
 Acq On : 25 Jun 2016 9:17
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 99 Sample Multiplier: 1


Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Quant Time: Jun 27 08:43:48 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF060716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 04:27:12 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)

46) 2-Chloronaphthalene	9.05	162	481470	3625.91	ng	100
47) 2-Nitroaniline	9.03	65	5598	142.91	ng #	16
48) Acenaphthylene	9.46	152	775823	3673.19	ng	99
49) Dimethylphthalate	9.35	163	609201	4098.41	ng	99
50) 2,6-Dinitrotoluene	9.05	165	16791	493.25	ng #	1
51) Acenaphthene	9.63	154	465766	3534.89	ng	98
52) 3-Nitroaniline	9.58	138	132801	3380.84	ng	96
54) Dibenzofuran	9.47	168	60416	332.95	ng #	54
55) 4-Nitrophenol	9.58	139	10624	348.47	ng #	1
56) 2,4-Dinitrotoluene	9.40	165	125174	2861.24	ng #	71
57) Fluorene	10.15	166	512303	4674.72	ng	98
58) 2,3,4,6-Tetrachlorophenol	9.90	232	105911	3156.78	ng #	61
59) Diethylphthalate	10.03	149	546410	3631.57	ng	99
60) 4-Chlorophenyl-phenylether	10.14	204	202095	3349.02	ng	91
61) 4-Nitroaniline	9.80	138	12590	337.90	ng #	56
62) Azobenzene	10.30	77	513920	3454.02	ng	97
64) 4,6-Dinitro-2-methylphenol	9.93	198	8249	1018.92	ng #	25
65) n-Nitrosodiphenylamine	9.93	169	12035	292.26	ng #	1
71) Anthracene	11.10	178	754262	11482.50	ng	99
72) Carbazole	10.89	167	25810	419.88	ng #	68
73) Di-n-butylphthalate	11.44	149	4136	53.15	ng #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.72 min Scan# 449
Delta R.T. -0.13 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion:152 Resp: 557826
Ion Ratio Lower Upper
152 100
150 154.7 123.8 185.6
115 70.9 39.6 59.4#



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5

2-Fluorophenol
Concen: 0.00 ng
Expected RT: 5.43 min

Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Tgt Ion: 112
Sig Exp Ratio
112 100
64 52.7
63 27.3





#6

Aniline

Concen: 6.81 ng

RT: 6.23 min Scan# 406

Delta R.T. -0.27 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion: 93 Resp: 383063

Ion Ratio Lower Upper

93 100

66 36.0 29.8 44.8

65 18.1 14.9 22.3



#7

Phenol-d6

Concen: 0.33 ng

RT: 6.35 min Scan# 417

Delta R.T. -0.13 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17




Tgt Ion: 99 Resp: 12867


Ion Ratio Lower Upper

99 100

42 6.6 12.2 18.2#

71 3.9 23.4 35.0#









#8
2-Chlorophenol
Concen: 7.46 ng
RT: 6.35 min Scan# 417
Delta R.T. -0.27 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion: 128 Resp: 288165
Ion Ratio Lower Upper
128 100
130 32.8 12.0 52.0
64 36.9 30.8 70.8




Abundance Ion 128.00 (127.70 to 128.70): E
Ion 130.00 (129.70 to 130.70): E
Ion 64.00 (63.70 to 64.70): BF0






#9
Benzaldehyde
Concen: 5.50 ng
RT: 6.10 min Scan# 395
Delta R.T. -0.29 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Tgt Ion: 77 Resp: 174317
Ion Ratio Lower Upper
77 100
105 97.2 90.6 130.6
106 92.9 85.3 125.3



Abundance Ion 77.00 (76.70 to 77.70): BF0
Ion 105.00 (104.70 to 105.70): E
Ion 106.00 (105.70 to 106.70): E





#10

Phenol

Concen: 6.91 ng

RT: 6.24 min Scan# 407

Delta R.T. -0.25 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion: 94 Resp: 368075

Ion Ratio Lower Upper

94 100

65 34.2 5.9 45.9

66 51.8 14.0 54.0



#11

bis(2-Chloroethyl)ether

Concen: 8.39 ng

RT: 6.31 min Scan# 413

Delta R.T. -0.27 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17




Tgt Ion: 93 Resp: 290927


Ion Ratio Lower Upper

93 100

63 69.6 67.6 107.6

95 32.3 12.5 52.5





#12

1,3-Dichlorobenzene

Concen: 5.73 ng

RT: 6.73 min Scan# 450

Delta R.T. -0.06 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:146 Resp: 237306

Ion Ratio Lower Upper

146 100

148 62.5 50.9 76.3




75 47.2 25.6 38.4#



Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 75.00 (74.70 to 75.70): BF0



#13

1,4-Dichlorobenzene

Concen: 7.57 ng

RT: 6.58 min Scan# 437

Delta R.T. -0.27 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17


Tgt Ion:146 Resp: 316033

Ion Ratio Lower Upper

146 100

148 64.5 52.0 78.0



111 39.2 33.8 50.8




Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E









#14
1,2-Dichlorobenzene
Concen: 6.51 ng
RT: 6.73 min Scan# 450
Delta R.T. -0.29 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 146 Resp: 247041
Ion Ratio Lower Upper
146 100
148 62.5 52.3 78.5
111 52.5 28.7 43.1#



#15
Benzyl Alcohol
Concen: 6.58 ng
RT: 6.73 min Scan# 450
Delta R.T. -0.26 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Tgt Ion: 79 Resp: 203547
Ion Ratio Lower Upper
79 100
108 78.7 65.0 97.6
77 64.7 50.3 75.5









#16
 2,2'-oxybis(1-Chloropropane)
 Concen: 5.60 ng
 RT: 6.86 min Scan# 461
 Delta R.T. -0.27 min
 Lab File: BF088365.D
 Acq: 25 Jun 2016 9:17




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion: 45 Resp: 262490
 Ion Ratio Lower Upper
 45 100
 77 51.5 0.0 32.5#
 79 47.9 0.0 29.5#



#17
 2-Methylphenol
 Concen: 8.83 ng
 RT: 7.02 min Scan# 475
 Delta R.T. -0.09 min
 Lab File: BF088365.D
 Acq: 25 Jun 2016 9:17

Tgt Ion: 107 Resp: 276474
 Ion Ratio Lower Upper
 107 100
 108 86.4 90.9 136.3#
 77 39.3 36.6 55.0
 79 26.2 36.5 54.7#









#18
Hexachloroethane
Concen: 7.45 ng
RT: 7.07 min Scan# 480
Delta R.T. -0.29 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 117 Resp: 110323
Ion Ratio Lower Upper
117 100
119 98.0 77.4 116.2
201 112.6 80.4 120.6



#19
n-Nitroso-di-n-propylamine
Concen: 6.99 ng
RT: 7.00 min Scan# 474
Delta R.T. -0.26 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Tgt Ion: 70 Resp: 176819
Ion Ratio Lower Upper
70 100
42 50.2 49.6 74.4
101 9.2 7.1 10.7
130 20.7 15.4 23.2









#20
3+4-Methylphenols
Concen: 7.46 ng
RT: 7.02 min Scan# 475
Delta R.T. -0.24 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	107	Resp:	272512
Ion Ratio	Lower	Upper	
107	100		
108	86.4	67.8	107.8
77	39.3	59.3	99.3#
79	26.2	7.0	47.0



#21
Naphthalene-d8
Concen: 20.00 ng
RT: 7.74 min Scan# 538
Delta R.T. -0.39 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Tgt Ion:	136	Resp:	1203
Ion Ratio	Lower	Upper	
136	100		
137	85.8	9.1	13.7#
54	49.2	6.6	10.0#
68	0.0	5.1	7.7#









#22
Acetophenone
Concen: 13982.36 ng
RT: 6.99 min Scan# 473
Delta R.T. -0.26 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion: 105 Resp: 375902
Ion Ratio Lower Upper
105 100
71 2.0 5.3 7.9#
51 23.0 18.2 27.4
120 21.4 18.0 27.0




Abundance Ion 105.00 (104.70 to 105.70): E
Ion 71.00 (70.70 to 71.70): BF0
Ion 51.00 (50.70 to 51.70): BF0
Ion 120.00 (119.70 to 120.70): E






#23
Nitrobenzene-d5
Concen: 47082.42 ng
RT: 7.15 min Scan# 487
Delta R.T. -0.26 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Tgt Ion: 82 Resp: 969963
Ion Ratio Lower Upper
82 100
128 48.7 35.9 53.9
54 45.0 40.9 61.3



Abundance Ion 82.00 (81.70 to 82.70): BF0
Ion 128.00 (127.70 to 128.70): E
Ion 54.00 (53.70 to 54.70): BF0









#24
Nitrobenzene
Concen: 12493.50 ng
RT: 7.16 min Scan# 488
Delta R.T. -0.26 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion	Ratio	Lower	Upper
77	100		
123	45.3	45.0	67.4
65	13.7	10.2	15.2



#25
Isophorone
Concen: 13777.13 ng
RT: 7.40 min Scan# 509
Delta R.T. -0.27 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Tgt Ion	Ratio	Lower	Upper
82	100		
95	7.2	5.4	8.2
138	18.5	14.6	21.8









#26
 2-Nitrophenol
 Concen: 14781.94 ng
 RT: 7.47 min Scan# 515
 Delta R.T. -0.27 min
 Lab File: BF088365.D
 Acq: 25 Jun 2016 9:17

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion:139 Resp: 158017
 Ion Ratio Lower Upper
 139 100
 109 29.6 23.0 34.4
 65 53.6 45.8 68.8




Abundance Ion 139.00 (138.70 to 139.70): E
 Ion 109.00 (108.70 to 109.70): E
 Ion 65.00 (64.70 to 65.70): BF0






#27
 2,4-Dimethylphenol
 Concen: 12191.37 ng
 RT: 7.54 min Scan# 521
 Delta R.T. -0.25 min
 Lab File: BF088365.D
 Acq: 25 Jun 2016 9:17

Tgt Ion:122 Resp: 239952
 Ion Ratio Lower Upper
 122 100
 107 102.4 82.4 123.6
 121 58.4 46.3 69.5



Abundance Ion 122.00 (121.70 to 122.70): E
 Ion 107.00 (106.70 to 107.70): E
 Ion 121.00 (120.70 to 121.70): E





#28

bis(2-Chloroethoxy)methane

Concen: 649.83 ng

RT: 7.47 min Scan# 515

Delta R.T. -0.41 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion: 93 Resp: 15380

Ion Ratio Lower Upper

93 100

95 9.9 26.5 39.7#

123 20.8 11.6 17.4#



#29

2,4-Dichlorophenol

Concen: 13912.65 ng

RT: 7.74 min Scan# 538

Delta R.T. -0.25 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17




Tgt Ion: 162 Resp: 228631


Ion Ratio Lower Upper

162 100

164 64.8 43.8 83.8

98 29.8 21.3 61.3









#30
 1,2,4-Trichlorobenzene
 Concen: 13126.03 ng
 RT: 7.79 min Scan# 543
 Delta R.T. -0.27 min
 Lab File: BF088365.D
 Acq: 25 Jun 2016 9:17

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion:180 Resp: 234877
 Ion Ratio Lower Upper
 180 100
 182 95.0 76.0 114.0
 145 35.8 26.5 39.7




Abundance Ion 180.00 (179.70 to 180.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 145.00 (144.70 to 145.70): E






#31
 Naphthalene
 Concen: 13420.94 ng
 RT: 7.87 min Scan# 550
 Delta R.T. -0.27 min
 Lab File: BF088365.D
 Acq: 25 Jun 2016 9:17

Tgt Ion:128 Resp: 798983
 Ion Ratio Lower Upper
 128 100
 129 11.3 9.4 14.2
 127 13.6 10.9 16.3



Abundance Ion 128.00 (127.70 to 128.70): E
 Ion 129.00 (128.70 to 129.70): E
 Ion 127.00 (126.70 to 127.70): E





#32

Benzoic acid

Concen: 6639.84 ng

RT: 7.70 min Scan# 535

Delta R.T. -0.22 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion:122 Resp: 71962

Ion Ratio Lower Upper

122 100

105 115.3 101.6 141.6

77 78.0 70.6 110.6



#33

4-Chloroaniline

Concen: 3963.62 ng

RT: 7.87 min Scan# 550

Delta R.T. -0.33 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17

Tgt Ion:127 Resp: 105368




Ion Ratio Lower Upper


127 100

129 83.1 26.3 39.5#

65 0.0 24.7 37.1#

92 0.0 15.4 23.0#









#34
Hexachlorobutadiene
Concen: 12911.28 ng
RT: 7.99 min Scan# 560
Delta R.T. -0.29 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion:225 Resp: 121843
Ion Ratio Lower Upper
225 100
223 63.0 50.9 76.3
227 62.8 51.9 77.9




Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E






#35
Caprolactam
Concen: 12007.23 ng
RT: 8.34 min Scan# 591
Delta R.T. -0.24 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Tgt Ion:113 Resp: 65359
Ion Ratio Lower Upper
113 100
55 113.9 158.6 198.6#
56 83.5 110.6 150.6#



Abundance Ion 113.00 (112.70 to 113.70): E
Ion 55.00 (54.70 to 55.70): BF0
Ion 56.00 (55.70 to 56.70): BF0









#36
 4-Chloro-3-methylphenol
 Concen: 14053.91 ng
 RT: 8.44 min Scan# 600
 Delta R.T. -0.24 min
 Lab File: BF088365.D
 Acq: 25 Jun 2016 9:17




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:107 Resp: 246990
 Ion Ratio Lower Upper
 107 100
 144 27.4 20.8 31.2
 142 85.7 63.4 95.2



#37
 2-Methylnaphthalene
 Concen: 5174.19 ng
 RT: 8.44 min Scan# 600
 Delta R.T. -0.40 min
 Lab File: BF088365.D
 Acq: 25 Jun 2016 9:17

Tgt Ion:142 Resp: 203028
 Ion Ratio Lower Upper
 142 100
 141 10.6 71.0 106.6#
 115 1.5 22.6 33.8#





#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 9.40 min Scan# 684

Delta R.T. -0.48 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:164 Resp: 2052

Ion Ratio Lower Upper

164 100

162 0.0 85.4 128.2#




160 0.0 39.5 59.3#



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#39

1,2,4,5-Tetrachlorobenzene

Concen: Below Cal

RT: 8.73 min Scan# 625

Delta R.T. -0.27 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17

Tgt Ion:216 Resp: 216698


Ion Ratio Lower Upper

216 100

214 79.7 2.9 4.3#

179 23.1 0.0 0.0#

108 21.3 0.5 0.7#






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E









#40
Hexachlorocyclopentadiene
Concen: 1319.17 ng
RT: 8.72 min Scan# 624
Delta R.T. -0.27 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion: 237 Resp: 43664
Ion Ratio Lower Upper
237 100
235 62.1 43.4 83.4
272 14.7 0.0 32.3




Abundance Ion 236.80 (236.50 to 237.50): E
Ion 234.90 (234.60 to 235.60): E
Ion 271.75 (271.45 to 272.45): E




#41
2,4,6-Tribromophenol
Concen: 0.00 ng
Expected RT: 10.67 min

Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Tgt Ion: 330
Sig Exp Ratio
330 100
332 0.0
141 0.0





#42

2,4,6-Trichlorophenol

Concen: 3617.03 ng

RT: 8.86 min Scan# 636

Delta R.T. -0.26 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:196 Resp: 148427

Ion Ratio Lower Upper

196 100

198 95.3 78.0 117.0




200 30.7 25.4 38.2



Abundance Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 200.00 (199.70 to 200.70): E



#43

2,4,5-Trichlorophenol

Concen: 3476.64 ng

RT: 8.86 min Scan# 636

Delta R.T. -0.31 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17

Tgt Ion:196 Resp: 148427


Ion Ratio Lower Upper

196 100

198 95.3 77.5 116.3

97 55.7 32.0 48.0#

132 46.5 20.9 31.3#






Abundance Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 97.00 (96.70 to 97.70): BF0

Ion 132.00 (131.70 to 132.70): E





#44

2-Fluorobiphenyl

Concen: 12384.60 ng

RT: 8.94 min Scan# 643

Delta R.T. -0.27 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17

Instrument :


BNA_F

ClientSampleId :

SSTDCCC040EC

Tgt Ion:172 Resp: 1513057



Ion	Ratio	Lower	Upper
172	100		
171	36.6	28.6	43.0
170	24.8	19.2	28.8



Abundance Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E



#45

1,1'-Biphenyl

Concen: 4388.48 ng

RT: 9.03 min Scan# 651


Delta R.T. -0.27 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17

Tgt Ion:154 Resp: 583677


Ion	Ratio	Lower	Upper
154	100		
153	41.2	20.6	60.6
76	13.6	0.0	35.1




Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 76.00 (75.70 to 76.70): BF0









#46
 2-Chloronaphthalene
 Concen: 3625.91 ng
 RT: 9.05 min Scan# 653
 Delta R.T. -0.27 min
 Lab File: BF088365.D
 Acq: 25 Jun 2016 9:17




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion: 162 Resp: 481470
 Ion Ratio Lower Upper
 162 100
 127 44.0 35.2 52.8
 164 32.8 26.6 39.8



#47
 2-Nitroaniline
 Concen: 142.91 ng
 RT: 9.03 min Scan# 651
 Delta R.T. -0.40 min
 Lab File: BF088365.D
 Acq: 25 Jun 2016 9:17

Tgt Ion: 65 Resp: 5598
 Ion Ratio Lower Upper
 65 100
 92 0.0 54.6 82.0#
 138 14.3 77.0 115.4#





#48

Acenaphthylene

Concen: 3673.19 ng

RT: 9.46 min Scan# 689

Delta R.T. -0.29 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:152 Resp: 775823

Ion Ratio Lower Upper




152 100

151 20.8 16.2 24.4

153 13.3 11.0 16.6



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E



#49

Dimethylphthalate

Concen: 4098.41 ng

RT: 9.35 min Scan# 679

Delta R.T. -0.26 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17


Tgt Ion:163 Resp: 609201

Ion Ratio Lower Upper



163 100


194 3.4 2.8 4.2

164 10.1 8.3 12.5



Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E





#50

2,6-Dinitrotoluene

Concen: 493.25 ng

RT: 9.05 min Scan# 653

Delta R.T. -0.62 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:165 Resp: 16791

Ion Ratio Lower Upper

165 100

63 249.8 28.1 42.1#




89 9.7 32.2 48.2#



Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF0

Ion 89.00 (88.70 to 89.70): BF0



#51

Acenaphthene

Concen: 3534.89 ng

RT: 9.63 min Scan# 704

Delta R.T. -0.29 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17


Tgt Ion:154 Resp: 465766

Ion Ratio Lower Upper

154 100

153 113.7 89.5 134.3



152 54.7 42.7 64.1




Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E









#52
3-Nitroaniline
Concen: 3380.84 ng
RT: 9.58 min Scan# 699
Delta R.T. -0.26 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion:138 Resp: 132801
Ion Ratio Lower Upper
138 100
108 10.4 8.5 12.7
92 115.2 95.7 143.5




Abundance Ion 138.00 (137.70 to 138.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 92.00 (91.70 to 92.70): BF0






#54
Dibenzofuran
Concen: 332.95 ng
RT: 9.47 min Scan# 690
Delta R.T. -0.62 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Tgt Ion:168 Resp: 60416
Ion Ratio Lower Upper
168 100
139 0.0 26.4 39.6#
169 6.7 10.4 15.6#



Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E









#55
4-Nitrophenol
Concen: 348.47 ng
RT: 9.58 min Scan# 699
Delta R.T. -0.42 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion:139 Resp: 10624
Ion Ratio Lower Upper
139 100
109 14.2 48.9 88.9#
65 1643.2 84.9 124.9#




Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BF0






#56
2,4-Dinitrotoluene
Concen: 2861.24 ng
RT: 9.40 min Scan# 684
Delta R.T. -0.66 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Tgt Ion:165 Resp: 125174
Ion Ratio Lower Upper
165 100
63 62.1 22.0 33.0#
89 58.6 41.1 61.7
182 0.0 2.0 3.0#



Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 89.00 (88.70 to 89.70): BF0
Ion 182.00 (181.70 to 182.70): E





#57

Fluorene

Concen: 4674.72 ng

RT: 10.15 min Scan# 749

Delta R.T. -0.29 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:166 Resp: 512303

Ion Ratio Lower Upper

166 100

165 99.3 77.8 116.8


167 13.4 11.2 16.8





Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



Time-->



#58

2,3,4,6-Tetrachlorophenol

Concen: 3156.78 ng

RT: 9.90 min Scan# 727

Delta R.T. -0.31 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17

Tgt Ion:232 Resp: 105911


Ion Ratio Lower Upper

232 100

131 43.9 0.9 1.3#

130 2.4 1.5 2.3#

166 25.9 0.5 0.7#




Abundance Ion 232.00 (231.70 to 232.70): E


Ion 131.00 (130.70 to 131.70): E


Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E



Time-->





#59

Diethylphthalate

Concen: 3631.57 ng

RT: 10.03 min Scan# 739

Delta R.T. -0.27 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:149 Resp: 546410

Ion Ratio Lower Upper

149 100

177 20.5 16.9 25.3




150 12.7 10.1 15.1



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E



#60

4-Chlorophenyl-phenylether

Concen: 3349.02 ng

RT: 10.14 min Scan# 748

Delta R.T. -0.29 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17


Tgt Ion:204 Resp: 202095

Ion Ratio Lower Upper

204 100

206 32.0 26.3 39.5



141 79.2 55.0 82.6




Abundance Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E









#61
4-Nitroaniline
Concen: 337.90 ng
RT: 9.80 min Scan# 719
Delta R.T. -0.65 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion: 138 Resp: 12590
Ion Ratio Lower Upper
138 100
92 29.7 27.3 67.3
108 20.8 46.7 86.7#




Abundance Ion 138.00 (137.70 to 138.70): E
Ion 92.00 (91.70 to 92.70): BF0
Ion 108.00 (107.70 to 108.70): E






#62
Azobenzene
Concen: 3454.02 ng
RT: 10.30 min Scan# 762
Delta R.T. -0.29 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Tgt Ion: 77 Resp: 513920
Ion Ratio Lower Upper
77 100
182 21.6 4.4 44.4
105 23.4 3.8 43.8
51 27.3 9.3 49.3



Abundance Ion 77.00 (76.70 to 77.70): BF0
Ion 182.00 (181.70 to 182.70): E
Ion 105.00 (104.70 to 105.70): E
Ion 51.00 (50.70 to 51.70): BF0





#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 10.80 min Scan# 806

Delta R.T. -0.56 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:188 Resp: 1241

Ion Ratio Lower Upper

188 100

94 953.3 9.0 13.6#




80 110.6 10.0 15.0#



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF0

Ion 80.00 (79.70 to 80.70): BF0



#64

4,6-Dinitro-2-methylphenol

Concen: 1018.92 ng

RT: 9.93 min Scan# 730

Delta R.T. -0.55 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17


Tgt Ion:198 Resp: 8249

Ion Ratio Lower Upper

198 100

51 4.8 39.6 79.6#



105 0.0 36.6 76.6#




Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BF0

Ion 105.00 (104.70 to 105.70): E









#65
 n-Nitrosodiphenylamine
 Concen: 292.26 ng
 RT: 9.93 min Scan# 730
 Delta R.T. -0.62 min
 Lab File: BF088365.D
 Acq: 25 Jun 2016 9:17

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion:169 Resp: 12035
 Ion Ratio Lower Upper
 169 100
 168 373.5 53.4 80.0#
 167 109.2 29.4 44.0#




Abundance Ion 169.00 (168.70 to 169.70): E
 Ion 168.00 (167.70 to 168.70): E
 Ion 167.00 (166.70 to 167.70): E






#71
 Anthracene
 Concen: 11482.50 ng
 RT: 11.10 min Scan# 832
 Delta R.T. -0.34 min
 Lab File: BF088365.D
 Acq: 25 Jun 2016 9:17

Tgt Ion:178 Resp: 754262
 Ion Ratio Lower Upper
 178 100
 176 20.0 15.6 23.4
 179 15.9 12.8 19.2



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E





#72

Carbazole

Concen: 419.88 ng

RT: 10.89 min Scan# 814

Delta R.T. -0.70 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:167 Resp: 25810

Ion Ratio Lower Upper

167 100

166 7.8 17.8 26.6#

139 0.0 11.2 16.8#




Abundance Ion 167.00 (166.70 to 167.70): E


Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E

Time-->



Time-->



#73

Di-n-butylphthalate

Concen: 53.15 ng

RT: 11.44 min Scan# 862

Delta R.T. -0.49 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17


Tgt Ion:149 Resp: 4136

Ion Ratio Lower Upper

149 100

150 10.8 7.8 11.6

104 0.0 4.0 6.0#




Abundance Ion 149.00 (148.70 to 149.70): E


Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E

Time-->



Time-->



#75

Chrysene-d12

Concen: 0.00 ng

Expected RT: 13.99 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC



Tgt Ion: 240

Sig Exp Ratio

240 100

120 8.5

236 25.2



#78

Terphenyl-d14

Concen: 0.00 ng

Expected RT: 12.95 min

Lab File: BF088365.D

Acq: 25 Jun 2016 9:17


Tgt Ion: 244


Sig Exp Ratio

244 100

212 7.5

122 14.0





#86
Perylene-d12
Concen: 0.00 ng
Expected RT: 15.39 min
Lab File: BF088365.D
Acq: 25 Jun 2016 9:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion: 264
Sig Exp Ratio
264 100
260 24.0
265 21.1

