

Data Path : Z:\HPCHEM1\BNA F\DATA\BF062417\
 Data File : BF096186.D
 Acq On : 24 Jun 2017 12:54
 Operator : SJ/MA
 Sample : I3872-04 5X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Jun 26 01:21:08 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 20 16:05:47 2017
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.26	152	70085	20.00	ng	-0.03
21) Naphthalene-d8	9.30	136	276999	20.00	ng	-0.03
38) Acenaphthene-d10	12.12	164	113066	20.00	ng	-0.04
63) Phenanthrene-d10	14.51	188	167935	20.00	ng	-0.03
75) Chrysene-d12	18.25	240	142843	20.00	ng	-0.02
86) Perylene-d12	19.93	264	117087	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.19	112	73397	16.69	ng	0.00
7) Phenol-d6	6.89	99	88730	16.85	ng	0.02
23) Nitrobenzene-d5	8.20	82	54143	12.14	ng	-0.02
41) 2,4,6-Tribromophenol	13.43	330	15397	15.39	ng	-0.02
44) 2-Fluorobiphenyl	11.07	172	122263	15.05	ng	-0.04
78) Terphenyl-d14	16.90	244	74017	12.86	ng	-0.03
Target Compounds						
2) 1,4-Dioxane	1.54	88	29391	14.05	ng	Qvalue # 23
22) Acetophenone	7.92	105	22538	3.48	ng	# 50
31) Naphthalene	9.33	128	62320	4.48	ng	98
37) 2-Methylnaphthalene	10.47	142	26406	2.81	ng	91
49) Dimethylphthalate	11.78	163	20054	2.34	ng	97
51) Acenaphthene	12.17	154	20546	2.86	ng	94
54) Dibenzofuran	12.47	168	32751	3.24	ng	97
55) 4-Nitrophenol	12.47	139	12232	13.67	ng	# 35
57) Fluorene	13.02	166	39103	4.44	ng	99
70) Phenanthrene	14.54	178	195738	20.20	ng	98
71) Anthracene	14.64	178	58656	5.80	ng	97
72) Carbazole	14.96	167	19929	2.02	ng	99
74) Fluoranthene	16.34	202	152279	15.88	ng	99
77) Pyrene	16.64	202	126701	11.89	ng	# 96
80) Benzo(a)anthracene	18.24	228	55242	6.65	ng	97
82) Chrysene	18.29	228	46097	5.55	ng	96
85) Indeno(1,2,3-cd)pyrene	21.11	276	18993	2.67	ng	# 88
87) Benzo(b)fluoranthene	19.53	252	52295	7.13	ng	# 94
88) Benzo(k)fluoranthene	19.53	252	52295	7.46	ng	97
89) Benzo(a)pyrene	19.88	252	36576	5.43	ng	# 90
91) Benzo(g,h,i)perylene	21.44	276	15482	2.66	ng	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF062417\
Data File : BF096186.D
Acq On : 24 Jun 2017 12:54
Operator : SJ/MA
Sample : I3872-04 5X
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :

Quant Time: Jun 26 01:21:08 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 20 16:05:47 2017
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.26 min Scan# 982
Delta R.T. -0.03 min
Lab File: BF096186.D
Acq: 24 Jun 2017 12:54

Instrument :
BNA_F
ClientSampleId :

Tot Ion:152 Resp: 70085
Ion Ratio Lower Upper
152 100
150 155.7 126.2 189.2
115 59.2 52.2 78.2




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 14.05 ng
RT: 1.54 min Scan# 9
Delta R.T. -0.07 min
Lab File: BF096186.D
Acq: 24 Jun 2017 12:54

Tgt Ion: 88 Resp: 29391
Ion Ratio Lower Upper
88 100
58 0.0 61.4 92.0#
43 4.8 23.4 35.0#



Abundance Ion 88.00 (87.70 to 88.70): BF0
Ion 58.00 (57.70 to 58.70): BF0
Ion 43.00 (42.70 to 43.70): BF0





#5

2-Fluorophenol

Concen: 16.69 ng

RT: 5.19 min Scan# 629


Delta R.T. 0.01 min

Lab File: BF096186.D




Acq: 24 Jun 2017 12:54

Instrument :
BNA_F
ClientSampled :

Tot Ion:	112	Resp:	73397
Ion	Ratio	Lower	Upper
112	100		
64	54.9	46.0	69.0
63	26.0	23.4	35.0



Abundance Ion 112.00 (111.70 to 112.70): BFO
Ion 64.00 (63.70 to 64.70): BFO
Ion 63.00 (62.70 to 63.70): BFO



#7

Phenol-d6

Concen: 16.85 ng


RT: 6.89 min Scan# 919

Delta R.T. 0.02 min



Lab File: BF096186.D

Acq: 24 Jun 2017 12:54

Tgt Ion:	99	Resp:	88730
Ion	Ratio	Lower	Upper
99	100		
42	16.2	13.5	20.3
71	29.0	24.5	36.7



Abundance Ion 99.00 (98.70 to 99.70): BFO
Ion 42.00 (41.70 to 42.70): BFO
Ion 71.00 (70.70 to 71.70): BFO





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 9.30 min Scan# 1328

Delta R.T. -0.03 min

Lab File: BF096186.D

Acq: 24 Jun 2017 12:54

Instrument :

BNA_F

ClientSampleId :

Tot Ion:136 Resp: 276999


Ion Ratio Lower Upper

136 100

137 10.9 8.5 12.7

54 6.8 4.9 7.3

68 5.5 4.1 6.1






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0



#22

Acetophenone

Concen: 3.48 ng

RT: 7.92 min Scan# 1093

Delta R.T. -0.08 min

Lab File: BF096186.D

Acq: 24 Jun 2017 12:54

Tgt Ion:105 Resp: 22538


Ion Ratio Lower Upper

105 100

71 3.5 2.8 4.2

51 4.0 30.7 46.1#

120 0.0 17.4 26.0#






Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF0

Ion 51.00 (50.70 to 51.70): BF0

Ion 120.00 (119.70 to 120.70): E









#23
Nitrobenzene-d5
Concen: 12.14 ng
RT: 8.20 min Scan# 1142
Delta R.T. -0.02 min
Lab File: BF096186.D
Acq: 24 Jun 2017 12:54




Instrument :
BNA_F
ClientSampled :


Tot Ion: 82 Resp: 54143
Ion Ratio Lower Upper
82 100
128 45.7 34.0 51.0
54 45.2 42.2 63.2



#31
Naphthalene
Concen: 4.48 ng
RT: 9.33 min Scan# 1334
Delta R.T. -0.03 min
Lab File: BF096186.D
Acq: 24 Jun 2017 12:54

Tgt Ion: 128 Resp: 62320
Ion Ratio Lower Upper
128 100
129 11.9 9.0 13.6
127 14.8 10.8 16.2





#37

2-Methylnaphthalene

Concen: 2.81 ng

RT: 10.47 min Scan# 1527

Delta R.T. -0.02 min

Lab File: BF096186.D

Acq: 24 Jun 2017 12:54

Instrument :

BNA_F

ClientSampleId :


Tot Ion:142 Resp: 26406

Ion Ratio Lower Upper

142 100

141 96.9 71.3 106.9




115 40.0 27.1 40.7



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 12.12 min Scan# 1807

Delta R.T. -0.04 min

Lab File: BF096186.D

Acq: 24 Jun 2017 12:54


Tgt Ion:164 Resp: 113066

Ion Ratio Lower Upper

164 100

162 101.8 82.6 123.8



160 41.6 36.2 54.2




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#41
 2,4,6-Tribromophenol
 Concen: 15.39 ng
 RT: 13.43 min Scan# 2031
 Delta R.T. -0.02 min
 Lab File: BF096186.D
 Acq: 24 Jun 2017 12:54

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:330 Resp: 15397
 Ion Ratio Lower Upper
 330 100
 332 100.9 76.6 115.0
 141 41.1 31.4 47.2




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 15.05 ng
 RT: 11.07 min Scan# 1630
 Delta R.T. -0.04 min
 Lab File: BF096186.D
 Acq: 24 Jun 2017 12:54

Tgt Ion:172 Resp: 122263
 Ion Ratio Lower Upper
 172 100
 171 36.5 29.6 44.4
 170 23.3 19.8 29.8



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#49

Dimethylphthalate

Concen: 2.34 ng

RT: 11.78 min Scan# 1750

Delta R.T. -0.04 min

Lab File: BF096186.D

Acq: 24 Jun 2017 12:54

Instrument :

BNA_F

ClientSampleId :


Tot Ion:163 Resp: 20054

Ion Ratio Lower Upper

163 100

194 2.8 2.6 4.0




164 11.9 8.4 12.6



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#51

Acenaphthene

Concen: 2.86 ng

RT: 12.17 min Scan# 1816

Delta R.T. -0.04 min

Lab File: BF096186.D

Acq: 24 Jun 2017 12:54


Tgt Ion:154 Resp: 20546

Ion Ratio Lower Upper

154 100

153 119.5 90.5 135.7



152 59.5 43.6 65.4




Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E





#54

Dibenzenofuran

Concen: 3.24 ng

RT: 12.47 min Scan# 1867

Delta R.T. -0.02 min

Lab File: BF096186.D

Acq: 24 Jun 2017 12:54

Instrument :

BNA_F

ClientSampleId :


Tot Ion:168 Resp: 32751

Ion Ratio Lower Upper

168 100

139 37.1 30.6 45.8




169 15.4 10.8 16.2



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E



#55

4-Nitrophenol

Concen: 13.67 ng

RT: 12.47 min Scan# 1867

Delta R.T. -0.16 min

Lab File: BF096186.D

Acq: 24 Jun 2017 12:54


Tgt Ion:139 Resp: 12232

Ion Ratio Lower Upper

139 100

109 9.5 34.1 74.1#



65 3.9 31.4 71.4#




Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF0









#57
 Fluorene
 Concen: 4.44 ng
 RT: 13.02 min Scan# 1960
 Delta R.T. -0.03 min
 Lab File: BF096186.D
 Acq: 24 Jun 2017 12:54




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:166 Resp: 39103
 Ion Ratio Lower Upper
 166 100
 165 98.8 78.8 118.2
 167 14.3 10.6 16.0



#63
 Phenanthrene-d10
 Concen: 20.00 ng
 RT: 14.51 min Scan# 2214
 Delta R.T. -0.03 min
 Lab File: BF096186.D
 Acq: 24 Jun 2017 12:54

Tgt Ion:188 Resp: 167935
 Ion Ratio Lower Upper
 188 100
 94 10.9 9.4 14.2
 80 10.4 10.2 15.2









#70
Phenanthrene
Concen: 20.20 ng
RT: 14.54 min Scan# 2220
Delta R.T. -0.04 min
Lab File: BF096186.D
Acq: 24 Jun 2017 12:54




Instrument :
BNA_F
ClientSampleId :


Tot Ion:178 Resp: 195738
Ion Ratio Lower Upper
178 100
176 18.7 16.2 24.4
179 15.5 12.6 19.0



#71
Anthracene
Concen: 5.80 ng
RT: 14.64 min Scan# 2237
Delta R.T. -0.02 min
Lab File: BF096186.D
Acq: 24 Jun 2017 12:54

Tgt Ion:178 Resp: 58656
Ion Ratio Lower Upper
178 100
176 18.2 15.8 23.8
179 17.2 12.7 19.1






#72
Carbazole
 Concen: 2.02 ng
 RT: 14.96 min Scan# 2291
 Delta R.T. -0.01 min
 Lab File: BF096186.D
 Acq: 24 Jun 2017 12:54




Instrument :
 BNA_F
 ClientSampled :

Tot Ion:167 Resp: 19929

Ion	Ratio	Lower	Upper
167	100		
166	22.3	18.1	27.1
139	13.7	11.7	17.5




Abundance Ion 167.00 (166.70 to 167.70): E
 Ion 166.00 (165.70 to 166.70): E
 Ion 139.00 (138.70 to 139.70): E





#74
Fluoranthene
 Concen: 15.88 ng
 RT: 16.34 min Scan# 2526
 Delta R.T. -0.02 min
 Lab File: BF096186.D
 Acq: 24 Jun 2017 12:54


Tgt Ion:202 Resp: 152279

Ion	Ratio	Lower	Upper
202	100		
101	13.1	0.0	33.2
203	17.3	0.0	38.1



Abundance Ion 202.00 (201.70 to 202.70): E
 Ion 101.00 (100.70 to 101.70): E
 Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 18.25 min Scan# 2850

Delta R.T. -0.02 min

Lab File: BF096186.D

Acq: 24 Jun 2017 12:54

Instrument :

BNA_F

ClientSampleId :


Tot Ion:240 Resp: 142843

Ion Ratio Lower Upper

240 100

120 14.5 5.8 8.8#




236 24.5 20.5 30.7



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#77

Pyrene

Concen: 11.89 ng

RT: 16.64 min Scan# 2577

Delta R.T. -0.02 min

Lab File: BF096186.D

Acq: 24 Jun 2017 12:54


Tgt Ion:202 Resp: 126701

Ion Ratio Lower Upper

202 100

200 21.3 17.6 26.4



203 21.6 14.4 21.6#




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E









#78
 Terphenyl-d14
 Concen: 12.86 ng
 RT: 16.90 min Scan# 2620
 Delta R.T. -0.03 min
 Lab File: BF096186.D
 Acq: 24 Jun 2017 12:54




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:244 Resp: 74017
 Ion Ratio Lower Upper
 244 100
 212 7.0 6.0 9.0
 122 12.7 10.3 15.5



#80
 Benzo(a)anthracene
 Concen: 6.65 ng
 RT: 18.24 min Scan# 2848
 Delta R.T. -0.02 min
 Lab File: BF096186.D
 Acq: 24 Jun 2017 12:54

Tgt Ion:228 Resp: 55242
 Ion Ratio Lower Upper
 228 100
 226 28.6 22.6 33.8
 229 23.4 16.3 24.5





#82

Chrysene

Concen: 5.55 ng

RT: 18.29 min Scan# 2856

Delta R.T. -0.02 min

Lab File: BF096186.D

Acq: 24 Jun 2017 12:54


Instrument :

BNA_F

ClientSampleId :

Tot Ion:228 Resp: 46097

Ion	Ratio	Lower	Upper
228	100		
226	32.2	24.9	37.3
229	22.9	15.8	23.6






Abundance

Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#85

Indeno(1,2,3-cd)pyrene

Concen: 2.67 ng

RT: 21.11 min Scan# 3336


Delta R.T. 0.01 min

Lab File: BF096186.D

Acq: 24 Jun 2017 12:54

Tgt Ion:276 Resp: 18993

Ion	Ratio	Lower	Upper
276	100		
138	28.0	27.8	41.6
277	18.5	20.2	30.4






Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#86
Pervlene-d12
Concen: 20.00 ng
RT: 19.93 min Scan# 3135
Delta R.T. -0.01 min
Lab File: BF096186.D
Acq: 24 Jun 2017 12:54




Instrument :
BNA_F
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
264	100		
260	23.4	19.5	29.3
265	21.3	17.2	25.8



#87
Benzo(b)fluoranthene
Concen: 7.13 ng
RT: 19.53 min Scan# 3068
Delta R.T. -0.01 min
Lab File: BF096186.D
Acq: 24 Jun 2017 12:54

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.3	18.1	27.1
125	18.7	9.8	14.8









#88
Benzo(k)fluoranthene
Concen: 7.46 ng
RT: 19.53 min Scan# 3068
Delta R.T. -0.04 min
Lab File: BF096186.D
Acq: 24 Jun 2017 12:54




Instrument :
BNA_F
ClientSampleId :


Tot Ion:252 Resp: 52295
Ion Ratio Lower Upper
252 100
253 22.3 18.4 27.6
125 18.7 13.1 19.7



#89
Benzo(a)pyrene
Concen: 5.43 ng
RT: 19.88 min Scan# 3127
Delta R.T. -0.01 min
Lab File: BF096186.D
Acq: 24 Jun 2017 12:54

Tgt Ion:252 Resp: 36576
Ion Ratio Lower Upper
252 100
253 24.1 17.5 26.3
125 21.6 11.0 16.4#






#91
 Benzo(a,h,i)perylene
 Concen: 2.66 ng
 RT: 21.44 min Scan# 3392
 Delta R.T. 0.02 min
 Lab File: BF096186.D
 Acq: 24 Jun 2017 12:54

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
276	100		
277	32.0	18.6	28.0
138	34.6	25.4	38.0



Abundance Ion 276.00 (275.70 to 276.70): E
 Ion 277.00 (276.70 to 277.70): E
 Ion 138.00 (137.70 to 138.70): E

